

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 01:00

Date of Inspection: 09/18/2020

Serial Number: 80-001119
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 06/07/2021 <i>08</i>	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1185592 Exp: 08/13/2022
0.000	0.049	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.081

Number of Simulators Used: 5

Remarks:
NA

10/7/2020
Exp. Date
Correction
WSP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

2669/833

Signature and Printed Name WILLIAM S MCKENZIE

09/18/2020
Date