

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORMOND BEACH P.D.  
Time of Inspection: 15:58

Date of Inspection: 11/05/2020

Serial Number: 80-001156  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
/ 0.000	0.055	0.084	0.205	0.081
/ 0.000	0.054	0.084	0.205	0.080
/ 0.000	0.054	0.084	0.204	0.080

Number of Simulators Used: 5

**Remarks:**

00: Ambient Fail MOUTH ALCOHOL.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Robert B Pearson #257*

ROBERT B PEARSON

Signature and Printed Name

11/05/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORMOND BEACH P.D.  
Time of Inspection: 14:31

Date of Inspection: 11/05/2020

Serial Number: 80-001156  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
0.008 / 0.000	0.056 / 0.055	0.086 / 0.085	0.206	0.081
0.007 / 0.000	0.056 / 0.055	0.085 / 0.085	0.206	0.080
0.008 / 0.000	0.056 / 0.055	0.085 / 0.085	0.206	0.080

Number of Simulators Used: 5

**Remarks:**

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance.

00: REPLACED WATER BROUGHT TO TEMPERATURE TEST PASSED SECOND TIME.

05: RAN TEST SECOND TIME WITHIN TOLERANCE

08: RAN TEST SECOND TIME WITHIN TOLERANCE

WILL BE DOING ANOTHER INSPECTION

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #267

ROBERT B PEARSON

Signature and Printed Name

11/05/2020  
Date