

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH SHORES
Time of Inspection: 00:11

Date of Inspection: 09/08/2020

Serial Number: 80-001242
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1305520 Exp: 04/01/2023
0.000	0.049	0.080	0.212 / 0.199	0.080
0.000	0.048	0.079	0.204 / 0.198	0.080
0.000	0.048	0.079	0.201 / 0.198	0.079

Number of Simulators Used: 5

Remarks:


20: Control Outside Tolerance.

↑

Tipped over .20 sim by accident. Let air dry and retried test, (M)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NORMAN R MEDDERS

Signature and Printed Name

09/08/2020
Date