

Florida Department of Law Enforcement Alcohol Testing Program

12808P

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 14:15

Date of Inspection: 02/12/2020

Serial Number: 80-001280
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.049	0.080	0.191	0.079
0.000	0.050	0.078	0.197	0.078
0.000	0.049	0.076	0.199	0.078

Number of Simulators Used: 5

Remarks:

INT DET. SIM NOT CONNECTED. CONNECTED AND RERAN.

REMOVED FROM EVID. USE AT PTDJF 02-12-2020

AGENCY INSPECTOR JACKSONVILLE SO SYSTEMS OFFICE

[Signature] 69016

ANALYSIS DISPERSED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

ROBERT D LIGHTSEY

Signature and Printed Name

02/12/2020
Date