

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CO. SO
Time of Inspection: 12:08

Date of Inspection: 08/07/2020

Serial Number: 80-001329
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1167545 Exp: 07/02/2022
0.000	0.048	0.078	0.196	0.000 / 0.084
0.000	0.048	0.077	0.196	0.000 / 0.083
0.000	0.048	0.079	0.197	0.083 / 0.083

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance. *(Instrument) 10-22-20*
Amended Dry Gas Connector not Attached 10-22-20

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MARK A KANTORSKI
Signature and Printed Name

08/07/2020
Date



Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Indian River County Sheriff's Office

Instrument Serial Number: 80-001329

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Mark Kantorski
Date of Inspection: 08/07/2020
Time of Inspection: 12:08:10
Agency Inspection Discrepancy: Incomplete, Untimely/Not Received, Erroneous Information, Procedural, Other (Required Information Missing)
Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
Lot Number, Expiration Date for g/ 210L, Alcohol Reference Solution, Dry Gas Standard is, Incorrect, Expired.
FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The REASON for repeating the following test(s); OR the Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
Alcohol Free Subject Test, Mouth Alcohol Test, Alcohol Free Test, Interferent Detect Test, 0.05 g/210L Test, 0.08 g/210L Test, 0.20 g/210L Test, 0.08 g/210L Dry Gas Standard Test
FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
Other:

OTHER ELECTRONIC DATA REVIEW
Login Records, Cylinder Change Records, Control Test Records, Diagnostic Check Records
Comments: The 0.08 g/210 L Dry Gas Standard Test was repeated. The reason for repeating the test, as well as corrective action taken prior to repeating the test, must be included. Please see below for the required corrective action and return to your Department Inspector by 10/18/2020.

CORRECTIVE ACTION
Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 10/18/2020 (Date).
Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date).
Upload the Agency Inspection(s).
Remove the instrument from evidentiary use until otherwise directed by the Department.
No action required
Other:

Taylor Muzdro

Signature of Alcohol Testing Program Staff Member

9/18/2020
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Indian River County Sheriff's Office

Instrument Serial Number: 80-001329

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Mark Kantorski
Date of Inspection: 08/07/2020
Time of Inspection: 12:08:10
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [] Erroneous Information [] Procedural [x] Other (Required Information Missing)
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[] Lot Number [] Expiration Date for _g/ 210L [] Alcohol Reference Solution [] Dry Gas Standard is [] Incorrect [] Expired.
[x] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [x] REASON for repeating the following test(s); OR the [x] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [x] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments:
The 0.08 g/210 L Dry Gas Standard Test was repeated. The reason for repeating the test, as well as corrective action taken prior to repeating the test, must be included. Please see below for the required corrective action and return to your Department Inspector by 10/18/2020.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 10/18/2020 (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Taylor Gutshew
Signature of Alcohol Testing Program Staff Member

9/18/2020
Date