

Amended

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 07:16

Date of Inspection: 12/24/2020

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:19390 Exp: 12/30/2021	0.08g/210L Test (g/210L) Lot#:20330 Exp: 07/10/2022	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.048	0.077	0.196	0.081
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080

Number of Simulators Used: 5

Remarks:

Inspection was non-compliant due to the use of non-approved solutions.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LUTHER E EAST

Signature and Printed Name

12/24/2020  
Date

Amended

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 01:24

Date of Inspection: 11/12/2020

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:19390 Exp: 12/30/2021	0.08g/210L Test (g/210L) Lot#:20330 Exp: 07/10/2022	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.046	0.077	0.196	0.080
0.000	0.046	0.077	0.196	0.080
0.000	0.047	0.077	0.195	0.080


Number of Simulators Used: 5

Remarks:

Inspection was non-compliant due to the use of non-approved solutions.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

LUIS H JUSINO

11/12/2020  
Date

*Amended*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD  
Time of Inspection: 23:03

Date of Inspection: 10/19/2020

Serial Number: 80-002097  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:20330 Exp: 07/10/2022	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00919080A3 Exp: 03/05/2021
0.000	0.045	0.077	0.195	0.080
0.000	0.046	0.077	0.197	0.080
0.000	0.046	0.076	0.197	0.080

Number of Simulators Used: 5

Remarks:

*Inspection was non-compliant due to the use of non-approved solutions*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*LS*  
*7 May 21*

*[Signature]*  
Signature and Printed Name

10/19/2020  
Date

LOUIS H JUSINO