

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 10:01

Date of Inspection: 09/21/2020

Serial Number: 80-005810

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.047	0.077	0.195	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ  
Signature and Printed Name

09/21/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 10:42

Date of Inspection: 08/24/2020

Serial Number: 80-005810  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.047	0.080	0.195	0.079
0.000	0.047	0.079	0.196	0.080
0.000	0.047	0.079	0.196	0.079

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

08/24/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 09:19

Date of Inspection: 07/30/2020

Serial Number: 80-005810

Software: 810G.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.080
0.000	0.047	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

07/30/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 11:32

Date of Inspection: 06/22/2020

Serial Number: 80-005810

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17919080A1 Exp: 08/05/2021
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.076	0.198	0.080
0.000	0.048	0.077	0.199	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ  
Signature and Printed Name

06/22/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 08:40

Date of Inspection: 05/26/2020

Serial Number: 80-005810

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.03 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.047	0.075	0.190 / 0.162	
0.000	0.046	0.075	0.189 / 0.171	
0.000	0.045	0.075	0.190 / 0.169	

Number of Simulators Used: 5

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: INSTRUMENT 80-005810 IS READING TOO LOW

05/26/2020 - SHIPPED TO FDLE FOR ANNUAL INSPECTION *G. Franz*

Copy

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L Franz*

GORDON L FRANZ

Signature and Printed Name

05/26/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 07:40

Date of Inspection: 04/23/2020

Serial Number: 80-005810

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.049	0.079	0.193	0.080
0.000	0.049	0.078	0.194	0.080
0.000	0.049	0.078	0.195	0.081

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

04/23/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

80-005810  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND FD

Time of Inspection: 09:21

Date of Inspection: 03/03/2020

Serial Number: 80-005810

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.047	0.076	0.190	0.081
0.000	0.046	0.077	0.191	0.081
0.000	0.046	0.077	0.192	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L Franz \_\_\_\_\_  
Signature and Printed Name GORDON L FRANZ

03/03/2020  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

70-005810  
43K

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Serial Number: 80-005810

Time of Inspection: 09:41

Date of Inspection: 02/11/2020

Software: 3100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.047	0.076	0.195	0.080
0.000	0.046	0.075	0.195	0.080
0.000	0.047	0.076	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L Franz*

GORDON L FRANZ

Signature and Printed Name

02/11/2020  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

80-005810  
BK

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 08:36

Date of Inspection: 01/27/2020

Serial Number: 80-005810  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#: 201808E Exp: 08/22/2020	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17919080A1 Exp: 08/05/2021
0.000	0.048	0.079	0.198	0.080
0.000	0.047	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

Copy

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

01/27/2020  
Date