



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Miami PD

Instrument Serial Number: 80-006456

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Jose Pastor
Date of Inspection: 10/27/2020 and 11/25/2020
Time of Inspection: 11:38:48 and 10:54:25
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [x] Erroneous Information [] Procedural [] Other
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[] Lot Number [x] Expiration Date for 0.08 g/ 210L [x] Alcohol Reference Solution [] Dry Gas Standard is [x] Incorrect [] Expired.
[] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [] REASON for repeating the following test(s); OR the [] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments:
The expiration date (day) for the 0.08 g/210 L Alcohol Reference Solution is incorrect on the October and November 2020 Inspections. Please see the corrective action below and return to the Department Inspector by 02/04/2021.

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 02/04/2021 (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Taylor Gussow

Signature of Alcohol Testing Program Staff Member

1/2/2021

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 10:54

Date of Inspection: 11/25/2020

Serial Number: 80-006456
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/20/2021	0.20g/210L Test (g/210L) Lot#: 201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 989917 Exp: 07/25/2021
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.081
0.000	0.048	0.079	0.201	0.081

Number of Simulators Used: 5

Remarks:

Amended 1/28/21 SP 5533

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jose F Pastor #5533 JOSE F PASTOR
Signature and Printed Name

11/25/2020
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 11:38

Date of Inspection: 10/27/2020

Serial Number: 80-006456
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201811C Exp: 11/13/2020	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/20/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:989917 Exp: 07/25/2021
0.000	0.049	0.079	0.200	0.082
0.000	0.049	0.079	0.200	0.082
0.000	0.050	0.079	0.200	0.081


Number of Simulators Used: 5

Remarks:

Amended 1/28/21 JF5527

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

JOSE F PASTOR

10/27/2020
 Date