



INSTRUMENT PROCESSING SHEET

Agency FDLE

S/N 80-000228

Florida Department of Law Enforcement

Date In 11/13/2019

DI Completion Date 1/22/20

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u>	<b>Quality Checks</b> Performed By <u>SP</u>	<b>Flow Calibration</b> Performed By _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>190</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input type="checkbox"/> Stability Checks	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)

Notes: Case broken at left keyboard connector. Unable to store keyboard correctly.

**Final Release Date**  
  
**FDLE**  
  
**JAN 30 2020**  
  
**Alcohol Testing Program**

Simulator	Serial #	Lot #/Exp
0.050	SD1012	201905A 5-14-21
0.080	DR1279	201905B 5-14-21
0.200	SD1013	201904D 4-30-21
0.080 DGS	N/A	AG916501 6-14-21

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 21.2  
 External Digital Therm. ID#: 300505  
 34°C +/- .2 Serial #: MP5088  
 34°C +/- .2 Serial #: MP5089  
 34°C +/- .2 Serial #: MP5090

**Calibration Adjustment** Performed By SP

Barometric Pressure Gauge 1028 ID # 28427

Simulator	Serial Number	Lot Number	Expiration
0.000	MP5091	N/A	N/A
0.040	MP5082	19080	3-4-21
0.100	MP5083	19160	7-9-21
0.200	MP5084	19040	1-29-21
0.300	MP5085	19010	1-3-21
0.080 DGS	N/A	08819080A1	6-5-21

**Department Inspection** Performed By SP

Barometric Pressure ID# 28421  
 Gauge 1024 Instrument 1022  
 Mouth Alcohol Solution Lot # 2018-B  
 Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	MP5086
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	MP5088	201905A	5-14-21
0.080	MP5089	201905B	5-14-21
0.200	MP5090	201904D	4-30-21
0.080 DGS	N/A	AG916501	6-14-21

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_

Notes/Suggested Service: \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

DPGM 1/30/20 Brett/Lickland 1/30/2020  
 Tech Review / Date Admin Review / Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000228, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000228</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/22/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:16</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2020

Date

*Shayla Platt*

SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 July 2018

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*ADAM*  
*BK 1/30/2020*



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 11:16

Date of Inspection: 01/22/2020

Serial Number: 80-000228

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.050	0.083	0.207	0.080
0.000	0.051	0.082	0.207	0.081
0.000	0.051	0.083	0.208	0.081
0.000	0.051	0.082	0.208	0.081
0.000	0.051	0.082	0.208	0.081
0.000	0.051	0.083	0.207	0.081
0.000	0.051	0.083	0.208	0.080
0.000	0.051	0.082	0.207	0.080
0.000	0.052	0.082	0.207	0.081
0.000	0.051	0.082	0.206	0.081

Standard Deviations	0.0004	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

PDM  
BK  
1/30/2020

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

01/22/2020  
Date

Stability Checks # 80-000228

FIDE  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/08/2020  
Software: 8100.27

SN 80-000228

g/210L

Test	Time
Air Blank	0.000
Control Test	0.081
Air Blank	0.000
Control Test	0.080
Air Blank	0.000
Control Test	0.081
Air Blank	0.000
Control Test Stats	0.0807
Average	0.0006
Std Dev	0.0006
Rel Std Dev(%)	0.7157

DGS

SP  
Operator's Signature

FIDE  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/08/2020  
Software: 8100.27

SN 80-000228

g/210L

Test	Time
Air Blank	0.000
Control Test	0.198
Air Blank	0.000
Control Test	0.196
Air Blank	0.000
Control Test	0.195
Air Blank	0.000
Control Test Stats	0.1963
Average	0.0015
Std Dev	0.0015
Rel Std Dev(%)	0.7780

SP  
Operator's Signature

FIDE  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/08/2020  
Software: 8100.27

SN 80-000228

g/210L

Test	Time
Air Blank	0.000
Control Test	0.077
Air Blank	0.000
Control Test	0.077
Air Blank	0.000
Control Test	0.078
Air Blank	0.000
Control Test Stats	0.0773
Average	0.0006
Std Dev	0.0006
Rel Std Dev(%)	0.7466

SP  
Operator's Signature

FIDE  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
01/08/2020  
Software: 8100.27

SN 80-000228

g/210L

Test	Time
Air Blank	0.000
Control Test	0.049
Air Blank	0.000
Control Test	0.049
Air Blank	0.000
Control Test	0.048
Air Blank	0.000
Control Test Stats	0.0487
Average	0.0006
Std Dev	0.0006
Rel Std Dev(%)	1.1863

SP  
Operator's Signature

updated time to reflect correct time SP

SPM  
TSK 1/30/2020



FOLE

Influx  
Model 6000  
01/17/2020  
08:47:04

Auto Calibration  
Max Power Res Value = 33  
Auto Range Res Value = 21

Sol Value = 0.000 g/210L \*\*\*  
Fit Value = 0.000 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum to = 12434, Sum to = 14226

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8630 (-0.0250)  
Sample #2 = 0.1140 (-0.0180)  
Sample #3 = 0.0950 (-0.0300)  
Sample #4 = 0.1160 (-0.0010)  
Avg % Abs = 0.1083 (-0.0160)  
STD DEV = 0.0116 (0.0161)  
REL STD DEV = 10.699 (103.58)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1350 (-0.0230)  
Sample #2 = 0.1410 (-0.0060)  
Sample #3 = 0.1110 (-0.0060)  
Sample #4 = 0.1230 (0.0070)  
Avg % Abs = 0.1250 (-0.0017)  
STD DEV = 0.0151 (0.0075)  
REL STD DEV = 12.080 (450.333)

Channel 1 Data:  
Sol Value = 0.040 g/210L \*\*\*  
Fit Value = 0.1905 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum to = 12434, Sum to = 14227

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8140 (0.0060)  
Sample #2 = 0.8310 (0.0290)  
Sample #3 = 0.8290 (-0.0010)  
Sample #4 = 0.8160 (0.0440)  
Avg % Abs = 0.8253 (0.0240)  
STD DEV = 0.0081 (0.0229)  
REL STD DEV = 0.987 (95.470)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5650 (0.0000)  
Sample #2 = 1.5270 (0.0580)  
Sample #3 = 1.5320 (0.0350)  
Sample #4 = 1.5270 (0.0630)  
Avg % Abs = 1.5287 (0.0520)  
STD DEV = 0.0029 (0.0149)  
REL STD DEV = 0.189 (28.718)

Sol Value = 0.100 g/210L \*\*\*  
Fit Value = 0.4762 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum to = 12430, Sum to = 14220

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.8940 (-0.0120)  
Sample #2 = 1.8700 (0.0490)  
Sample #3 = 1.8780 (0.0200)  
Sample #4 = 1.8520 (0.1720)  
Avg % Abs = 1.8657 (0.0470)  
STD DEV = 0.0133 (0.0261)  
REL STD DEV = 0.713 (55.442)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6320 (-0.0250)  
Sample #2 = 3.5470 (0.0650)  
Sample #3 = 3.5420 (0.0830)  
Sample #4 = 3.5300 (0.1050)  
Avg % Abs = 3.5357 (0.0910)  
STD DEV = 0.0087 (0.0122)  
REL STD DEV = 0.247 (13.369)

Sol Value = 0.200 g/210L \*\*\*  
Fit Value = 0.9524 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum to = 12431, Sum to = 14219

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.7410 (-0.0200)  
Sample #2 = 3.5950 (0.1050)  
Sample #3 = 3.5740 (0.1550)  
Sample #4 = 3.5650 (0.1350)  
Avg % Abs = 3.5760 (0.1317)  
STD DEV = 0.0154 (0.0252)  
REL STD DEV = 0.429 (66.114)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.0660 (-0.0040)  
Sample #2 = 6.7960 (0.1990)  
Sample #3 = 6.7690 (0.2350)  
Sample #4 = 6.7760 (0.2280)  
Avg % Abs = 6.7810 (0.2287)  
STD DEV = 0.0151 (0.0191)  
REL STD DEV = 0.223 (8.650)

Sol Value = 0.300 g/210L \*\*\*  
Fit Value = 1.4286 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
Sum to = 12425, Sum to = 14213

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.3590 (-0.0010)  
Sample #2 = 5.2510 (0.1270)  
Sample #3 = 5.2180 (0.1580)  
Sample #4 = 5.1660 (0.1860)  
Avg % Abs = 5.2190 (0.1577)  
STD DEV = 0.0315 (0.0305)  
REL STD DEV = 0.604 (19.345)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.0760 (0.0100)  
Sample #2 = 9.7940 (0.3020)  
Sample #3 = 9.7620 (0.3220)  
Sample #4 = 9.7630 (0.3300)  
Avg % Abs = 9.7730 (0.3180)  
STD DEV = 0.0182 (0.0144)  
REL STD DEV = 0.186 (4.535)

Auto Cal Data:  
Sol Val = 0.000 mg/l or 0.000 g/210L  
% Abs = 0.108  
Std Dev = 0.01 Rel Std Dev = 10.70  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.825  
Std Dev = 0.01 Rel Std Dev = 0.99  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.867  
Std Dev = 0.01 Rel Std Dev = 0.71  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.578  
Std Dev = 0.02 Rel Std Dev = 0.43  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.219  
Std Dev = 0.03 Rel Std Dev = 0.60  
Zero Order Coef = -287.85  
First Order Coef = 2644.11  
Second Order Coef = 28.28  
Standard Deviation = 10.688113

Channel 2 Data:  
Sol Val = 0.000 mg/l or 0.000 g/210L  
% Abs = 0.125  
Std Dev = 0.02 Rel Std Dev = 12.08  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.529  
Std Dev = 0.00 Rel Std Dev = 0.15  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.540  
Std Dev = 0.01 Rel Std Dev = 0.25  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.781  
Std Dev = 0.02 Rel Std Dev = 0.22  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.773  
Std Dev = 0.02 Rel Std Dev = 0.19  
Zero Order Coef = -160.12  
First Order Coef = 1329.63  
Second Order Coef = 15.09  
Standard Deviation = 19.602335

Channel 1 Data:  
Sol Val = 0.040 mg/l or 0.000 g/210L  
% Abs = 0.000  
Std Dev = 0.02 Rel Std Dev = 0.43  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.529  
Std Dev = 0.00 Rel Std Dev = 0.15  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.540  
Std Dev = 0.01 Rel Std Dev = 0.25  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.781  
Std Dev = 0.02 Rel Std Dev = 0.22  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.773  
Std Dev = 0.02 Rel Std Dev = 0.19  
Zero Order Coef = -160.12  
First Order Coef = 1329.63  
Second Order Coef = 15.09  
Standard Deviation = 19.602335

Solution Stats Quadratic Fit Chan 1:  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0000  
0.040 -0.040 -0.0002  
0.100 0.100 0.0003  
0.200 0.200 -0.0002  
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2:  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.000 -0.0001  
0.040 0.040 -0.0001  
0.100 0.099 0.0006  
0.200 0.201 -0.0005  
0.300 0.300 0.0002

Sol Value = 0.060 g/210L \*\*\*  
Fit Value = 0.3810 mg/l \*\*\*\*  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1

Channel 1 Data:  
Sample #1 = 3300.00  
Sample #2 = 3271.00  
Sample #3 = 3277.00  
Sample #4 = 3149.00  
Average Result = 3232.3333  
STD DEV = 72.2311  
REL STD DEV = 2.235  
\*\*\*\*\* CHANNEL 2

Channel 2 Data:  
Sample #1 = 3430.00  
Sample #2 = 3437.00  
Sample #3 = 3452.00  
Sample #4 = 3430.00  
Average Result = 3439.6667  
STD DEV = 11.2398  
REL STD DEV = 0.327  
\*\*\*\*\* CHANNEL 1

Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1428  
3 um H2O Adjust (mg/l\*10,000) = 577  
9 um H2O Adjust (mg/l\*10,000) = 378  
\*\*\*\*\* AUTO CAL PASS

CAL ADJUSTMENT  
# 80-000228 SP

SK 1/30/2020

Post Cal-Adjust  
Stability Checks #80-000228

FOLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000228  
01/17/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.050	10:12
Air Blank	0.000	10:12
Control Test	0.050	10:13
Air Blank	0.000	10:14
Control Test	0.049	10:14
Air Blank	0.000	10:15
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

SP

Operator's Signature

FOLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000228  
01/17/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.082	10:22
Air Blank	0.000	10:23
Control Test	0.081	10:23
Air Blank	0.000	10:24
Control Test	0.080	10:25
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

SP

Operator's Signature

FOLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000228  
01/17/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.205	10:34
Air Blank	0.000	10:35
Control Test	0.204	10:35
Air Blank	0.000	10:36
Control Test	0.204	10:37
Air Blank	0.000	10:38
Control Test Stats		
Average	0.2043	
Std Dev	0.0006	
Rel Std Dev(%)	0.2826	

SP

Operator's Signature

FOLE  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000228  
01/17/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.080	10:03
Air Blank	0.000	10:04
Control Test	0.081	10:04
Air Blank	0.000	10:04
Control Test	0.081	10:05
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

SP

Operator's Signature

QPM BK 1/30/2020

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

14:01

Time of Inspection: 02:01 SP

Date of Inspection: 01/08/2020

Serial Number: 80-000228  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:

Time-Date changed. BYPASSED AI TO OPERATE INSTRUMENT

Room  
BK  
1/30/2020

N/A COMPLIANCE NOT  
determined.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

01/08/2020  
Date