




INSTRUMENT PROCESSING SHEET

Agency FHP Panama CityS/N 80-000784

Florida Department of Law Enforcement

Date In 04/07/2020DI Completion Date 5/11/20 Ship P/U H/D CMI EE

Intake Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>RS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>225</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.12</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	<u>MP5088</u>	<u>201905A</u> <u>05-14-2021</u>																																																
0.080	<u>MP5089</u>	<u>201905B</u> <u>05-14-2021</u>																																																
0.200	<u>MP5090</u>	<u>201904D</u> <u>04-30-2021</u>																																																
0.080 DGS	<u>N/A</u>	<u>AG931603</u> <u>11-12-2021</u>																																																
Final Release Date _____ _____	Calibration Adjustment Performed By <u>SP</u> Barometric Pressure Gauge <u>1019</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>05819080A</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>	0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	<u>N/A</u>	<u>05819080A</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>
Simulator	Serial Number	Lot Number	Expiration																																															
0.000	<u>MP5091</u>	<u>N/A</u>	<u>N/A</u>																																															
0.040	<u>MP5082</u>	<u>19080</u>	<u>3-4-21</u>																																															
0.100	<u>MP5083</u>	<u>19160</u>	<u>7-9-21</u>																																															
0.200	<u>MP5084</u>	<u>19040</u>	<u>1-29-21</u>																																															
0.300	<u>MP5085</u>	<u>19010</u>	<u>1-3-21</u>																																															
0.080 DGS	<u>N/A</u>	<u>05819080A</u>	<u>6-5-21</u>																																															
Simulator	Serial Number	Lot Number	Expiration																																															
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																															
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																															
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																															
0.080 DGS	<u>N/A</u>	<u>AG931603</u>	<u>11-12-21</u>																																															
Notes/Suggested Service: _____ _____ _____ _____ _____	Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1019</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>																																				
Simulator	Serial Number																																																	
0.000	<u>MP5086</u>																																																	
Interferent	<u>MP5087</u>																																																	
0.050	<u>MP5088</u>																																																	
0.080	<u>MP5089</u>																																																	
0.200	<u>MP5090</u>																																																	



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000784, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000784</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/11/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>10:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/11/2020

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:48

Date of Inspection: 05/11/2020

Serial Number: 80-000784
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080

Standard Deviations	0.0005	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/11/2020
Date

Stability Checks

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000784
04/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:07
Control Test	0.048	06:07
Air Blank	0.000	06:08
Control Test	0.047	06:09
Air Blank	0.000	06:09
Control Test	0.047	06:10
Air Blank	0.000	06:10
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	


FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000784
04/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:12
Control Test	0.078	06:13
Air Blank	0.000	06:13
Control Test	0.078	06:14
Air Blank	0.000	06:15
Control Test	0.078	06:15
Air Blank	0.000	06:16
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000784
04/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:19
Control Test	0.200	06:19
Air Blank	0.000	06:20
Control Test	0.199	06:21
Air Blank	0.000	06:21
Control Test	0.199	06:22
Air Blank	0.000	06:22
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	


wet



Operator's Signature



Operator's Signature



Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000784
04/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:30
Control Test	0.082	06:31
Air Blank	0.000	06:31
Control Test	0.081	06:32
Air Blank	0.000	06:32
Control Test	0.082	06:32
Air Blank	0.000	06:33
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Dry



Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 05/11/2020 07:24:41
 SN 80-000784
 Auto Calibration
 Max Power Res Value = 91
 Auto Range Res Value = 65

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.6050 (-0.0190)
 Sample #2 = 1.5970 (0.0050)
 Sample #3 = 1.6110 (0.0030)
 Sample #4 = 1.6010 (0.0060)
 Avg % Abs = 1.6030 (0.0047)
 STD DEV = 0.0072 (0.0015)
 REL STD DEV = 0.450 (32.733)

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12469, 9um Io = 12676
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1940 (-0.0260)
 Sample #2 = 0.1640 (0.0120)
 Sample #3 = 0.1790 (0.0050)
 Sample #4 = 0.1710 (0.0450)
 Avg % Abs = 0.1713 (0.0207)
 STD DEV = 0.0075 (0.0214)
 REL STD DEV = 4.381 (1103.364)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1950 (-0.0200)
 Sample #2 = 0.1870 (-0.0060)
 Sample #3 = 0.1690 (0.0120)
 Sample #4 = 0.1910 (0.0140)
 Avg % Abs = 0.1823 (0.0067)
 STD DEV = 0.0117 (0.0110)
 REL STD DEV = 6.427 (165.227)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12479, 9um Io = 12670
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8790 (-0.0290)
 Sample #2 = 0.8660 (0.0190)
 Sample #3 = 0.8650 (0.0400)
 Sample #4 = 0.8750 (0.0140)
 Avg % Abs = 0.8753 (0.0243)
 STD DEV = 0.0105 (0.0138)
 REL STD DEV = 1.200 (56.696)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.171
 Std Dev = 0.01 Rel Std Dev = 4.38
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.875
 Std Dev = 0.01 Rel Std Dev = 1.20
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.914
 Std Dev = 0.02 Rel Std Dev = 0.91
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.593
 Std Dev = 0.01 Rel Std Dev = 0.14
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.201
 Std Dev = 0.03 Rel Std Dev = 0.49
 Zero Order Coef = -451.44
 First Order Coef = 2655.63
 Second Order Coef = 33.87
 Standard Deviation = 5.619753

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.182
 Std Dev = 0.01 Rel Std Dev = 6.43
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.603
 Std Dev = 0.01 Rel Std Dev = 0.45
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.650
 Std Dev = 0.02 Rel Std Dev = 0.51
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.952
 Std Dev = 0.01 Rel Std Dev = 0.08
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.040
 Std Dev = 0.01 Rel Std Dev = 0.12
 Zero Order Coef = -237.15
 First Order Coef = 1318.01
 Second Order Coef = 12.74
 Standard Deviation = 13.748235

<<<<< CHANNEL 2 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.182
 Std Dev = 0.01 Rel Std Dev = 6.43
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.603
 Std Dev = 0.01 Rel Std Dev = 0.45
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.650
 Std Dev = 0.02 Rel Std Dev = 0.51
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.952
 Std Dev = 0.01 Rel Std Dev = 0.08
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.040
 Std Dev = 0.01 Rel Std Dev = 0.12
 Zero Order Coef = -237.15
 First Order Coef = 1318.01
 Second Order Coef = 12.74
 Standard Deviation = 13.748235

<<<<< CHANNEL 2 >>>>>
 Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12465, 9um Io = 12668
 <<<<< CHANNEL 1 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6210 (-0.0030)
 Sample #2 = 3.5980 (0.0220)
 Sample #3 = 3.5930 (0.0090)
 Sample #4 = 3.5980 (0.0440)
 Avg % Abs = 3.5930 (0.0250)
 STD DEV = 0.0050 (0.0177)
 REL STD DEV = 0.139 (70.767)

 : Solution Stats Quadratic Fit Chan 1 :
 : Act Fit Residual :
 : g/210L g/210L g/210L :
 : 0.000 0.000 -0.0001 :
 : 0.040 0.040 0.0001 :
 : 0.100 0.100 0.0001 :
 : 0.200 0.200 -0.0002 :
 : 0.300 0.300 0.0001 :

 : Solution Stats Quadratic Fit Chan 2 :
 : Act Fit Residual :
 : g/210L g/210L g/210L :
 : 0.000 0.000 -0.0001 :
 : 0.040 0.040 -0.0001 :
 : 0.100 0.100 0.0004 :
 : 0.200 0.200 -0.0004 :
 : 0.300 0.300 0.0001 :

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 2843.00
 Sample #2 = 2754.00
 Sample #3 = 2702.00
 Sample #4 = 2699.00
 Average Result = 2718.3333
 STD DEV = 30.9246
 REL STD DEV = 1.138
 ***** CHANNEL 2 *****
 Sample #1 = 3262.00
 Sample #2 = 3270.00
 Sample #3 = 3242.00
 Sample #4 = 3260.00
 Average Result = 3257.3333
 STD DEV = 14.1892
 REL STD DEV = 0.436

***** CHANNEL 1 *****
 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1018
 3 um H2O Adjust (mg/l*10,000) = 1091
 9 um H2O Adjust (mg/l*10,000) = 552
 ***** AUTO CAL PASS *****

CAL ADJUSTMENT 8P
 80-000784

Post Cal Adjust Stability Checks #80-000784

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
05/11/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.051	08:14
Air Blank	0.000	08:15
Control Test	0.049	08:16
Air Blank	0.000	08:16
Control Test	0.050	08:17
Air Blank	0.000	08:17
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
05/11/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:18
Control Test	0.079	08:19
Air Blank	0.000	08:19
Control Test	0.080	08:20
Air Blank	0.000	08:21
Control Test	0.080	08:21
Air Blank	0.000	08:22
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
05/11/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:26
Control Test	0.202	08:27
Air Blank	0.000	08:28
Control Test	0.201	08:28
Air Blank	0.000	08:29
Control Test	0.200	08:29
Air Blank	0.000	08:30
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

SP

Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000
05/11/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.080	08:23
Air Blank	0.000	08:24
Control Test	0.081	08:24
Air Blank	0.000	08:25
Control Test	0.080	08:25
Air Blank	0.000	08:25
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DGS

SP

Operator's Signature