



INSTRUMENT PROCESSING SHEET

Agency Perry PD

S/N 80-000811

Florida Department of Law Enforcement

Date In 11-17-2020

DI Completion Date 11-18-2020

Ship P/U H/D CMI E

Intake Performed By IS Quality Checks Performed By IS Flow Calibration Performed By

Notes: Instrument did not identify acetone as interferen during agency inspection.

Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.20 11:14:08 -05'00'

Table with 3 columns: Simulator, Serial #, Lot #/Exp. Rows include 0.050 MP5088, 0.080 MP5086, 0.200 MP5090, 0.080 DGS N/A.

Maintenance Performed By Temperature Checks Performed By IS Lab Temp 20.73 External Digital Therm. ID#: 300503

Calibration Adjustment Performed By IS Barometric Pressure Gauge 1031 ID # 28421. Includes table with Simulator, Serial Number, Lot Number, Expiration.

Department Inspection Performed By IS Barometric Pressure ID# 30793 Gauge 1029 Instrument 1029. Includes table with Simulator, Serial Number.

Notes/Suggested Service: Stability checks nominal, performed optical bench cal adjust due to instrument not identifying acetone as interferent during agency inspection. Tech Review: Put check marks for Cal. Adjust and Post stab 11-20-2020

Attachments Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, Other. Includes checkboxes for instrument compliance and return to evidentiary use.

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY POLICE DEPT  
Time of Inspection: 13:47

Date of Inspection: 11/18/2020

Serial Number: 80-000811  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
Standard Deviations	0.0003	0.0004	0.0003	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

*MX*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name


11/18/2020  
Date

*SR*  
2020.11.  
20  
11:10:43  
-05'00'

# Stability Checks

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:04
Control Test	0.048	09:05
Air Blank	0.000	09:06
Control Test	0.049	09:06
Air Blank	0.000	09:07
Control Test	0.049	09:08
Air Blank	0.000	09:08
Control Test Stats		
Average	0.0487	
Std Dev	0.0016	
Rel Std Dev(%)	1.1863	




Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:10
Control Test	0.079	09:11
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:12
Control Test	0.079	09:13
Air Blank	0.000	09:14
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.201	09:15
Air Blank	0.000	09:16
Control Test	0.201	09:17
Air Blank	0.000	09:17
Control Test	0.201	09:18
Air Blank	0.000	09:18
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature


PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.080	09:21
Air Blank	0.000	09:21
Control Test	0.080	09:22
Air Blank	0.000	09:22
Control Test	0.081	09:22
Air Blank	0.000	09:23
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry



Operator's Signature

2020.11.  
 20  
 11:09:45  
 -05'00'  




# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000811, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000811</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PERRY POLICE DEPT</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/18/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:47</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/18/2020 Date Israel Soto

**ISRAEL SOTO,**  
**Department Inspector**

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2020.11.  
20  
11:08:59  
-05'00'



# Optical Bench Cal Adj

*Small Adj*

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
11/18/2020 09:38:11

Auto Calibration  
Max Power Res Value = 53  
Auto Range Res Value = 30

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12867, Sum Io = 13353

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0770 (-0.0063)  
Sample #2 = 0.0630 (0.0220)  
Sample #3 = 0.0830 (0.0490)  
Sample #4 = 0.0760 (0.0500)  
Avg % Abs = 0.0740 (0.0403)  
STD DEV = 0.0101 (0.0159)  
REL STD DEV = 13.715 (39.384)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1040 (-0.0070)  
Sample #2 = 0.0980 (-0.0070)  
Sample #3 = 0.0990 (0.0000)  
Sample #4 = 0.1120 (0.0000)  
Avg % Abs = 0.1030 (-0.0023)  
STD DEV = 0.0078 (0.0040)  
REL STD DEV = 7.583 (173.205)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12853, Sum Io = 13347

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.7970 (-0.0120)  
Sample #2 = 0.7500 (0.0120)  
Sample #3 = 0.7670 (0.0180)  
Sample #4 = 0.7780 (0.0070)  
Avg % Abs = 0.7650 (0.0123)  
STD DEV = 0.0141 (0.0055)  
REL STD DEV = 1.844 (44.556)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5760 (-0.0180)  
Sample #2 = 1.5450 (0.0020)  
Sample #3 = 1.5540 (0.0040)  
Sample #4 = 1.5350 (0.0000)  
Avg % Abs = 1.5447 (0.0020)  
STD DEV = 0.0095 (0.0020)  
REL STD DEV = 0.615 (100.000)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12848, Sum Io = 13346

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7830 (-0.0170)  
Sample #2 = 1.7680 (0.0140)  
Sample #3 = 1.7760 (0.0140)  
Sample #4 = 1.7440 (0.0460)  
Avg % Abs = 1.7627 (0.0247)  
STD DEV = 0.0167 (0.0185)  
REL STD DEV = 0.945 (74.899)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5690 (-0.0160)  
Sample #2 = 3.5340 (0.0070)  
Sample #3 = 3.6460 (0.0030)  
Sample #4 = 3.6560 (0.0080)  
Avg % Abs = 3.6453 (0.0060)  
STD DEV = 0.0110 (0.0026)  
REL STD DEV = 0.302 (44.096)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12840, Sum Io = 13343

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4450 (-0.0190)  
Sample #2 = 3.4590 (-0.0170)  
Sample #3 = 3.4570 (-0.0150)  
Sample #4 = 3.4240 (0.0180)  
Avg % Abs = 3.4467 (-0.0047)  
STD DEV = 0.0197 (0.0197)  
REL STD DEV = 0.570 (421.186)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.0190 (0.0030)  
Sample #2 = 7.0120 (0.0040)  
Sample #3 = 7.0140 (0.0000)  
Sample #4 = 7.0030 (0.0210)  
Avg % Abs = 7.0097 (0.0083)  
STD DEV = 0.0059 (0.0112)  
REL STD DEV = 0.084 (133.806)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12842, Sum Io = 13342

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.0580 (0.0000)  
Sample #2 = 5.0410 (0.0200)  
Sample #3 = 5.0160 (0.0290)  
Sample #4 = 5.0140 (0.0570)  
Avg % Abs = 5.0237 (0.0353)  
STD DEV = 0.0150 (0.0193)  
REL STD DEV = 0.299 (54.611)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.1640 (-0.0090)  
Sample #2 = 10.1140 (0.0410)  
Sample #3 = 10.1280 (0.0320)  
Sample #4 = 10.1220 (0.0480)  
Avg % Abs = 10.1213 (0.0403)  
STD DEV = 0.0070 (0.0080)  
REL STD DEV = 0.069 (19.886)

\*\*\*\* AUTO CAL DATA \*\*\*\*  
Channel 1 Data:  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.074  
Std Dev = 0.01 Rel Std Dev = 13.71  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.765  
Std Dev = 0.01 Rel Std Dev = 1.84  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.763  
Std Dev = 0.02 Rel Std Dev = 0.94  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.447  
Std Dev = 0.02 Rel Std Dev = 0.57  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.024  
Std Dev = 0.02 Rel Std Dev = 0.30  
Zero Order Coef = -192.29  
First Order Coef = 2730.68  
Second Order Coef = 29.48  
Standard Deviation = 35.261311

Channel 2 Data:  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.103  
Std Dev = 0.01 Rel Std Dev = 7.58  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.545  
Std Dev = 0.01 Rel Std Dev = 0.62  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.645  
Std Dev = 0.01 Rel Std Dev = 0.30  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.010  
Std Dev = 0.01 Rel Std Dev = 0.08  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.121  
Std Dev = 0.01 Rel Std Dev = 0.07  
Zero Order Coef = -119.70  
First Order Coef = 1283.54  
Second Order Coef = 13.70  
Standard Deviation = 19.400053

*MX*  
*SP*  
2020.1  
1.20  
11:08:31  
-05'00'

# Post Stability Checks

Optical Bench Cal Adj

*SS*

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	-0.0002
0.100	0.099	0.0010
0.200	0.201	-0.0010
0.300	0.300	0.0003

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.100	0.0004
0.200	0.201	-0.0006
0.300	0.300	0.0002

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 3325.00  
 Sample #2 = 3352.00  
 Sample #3 = 3366.00  
 Sample #4 = 3204.00  
 Average Result = 3307.3333  
 STD DEV = 89.7626  
 REL STD DEV = 2.714

\*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3430.00  
 Sample #2 = 3435.00  
 Sample #3 = 3452.00  
 Sample #4 = 3418.00  
 Average Result = 3435.0000  
 STD DEV = 17.0000  
 REL STD DEV = 0.495

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1031  
 3 um H2O Adjust (mg/l\*10,000) = 502  
 9 um H2O Adjust (mg/l\*10,000) = 374  
 \*\*\*\* AUTO CAL PASS

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:47
Control Test	0.048	11:48
Air Blank	0.000	11:48
Control Test	0.049	11:49
Air Blank	0.000	11:49
Control Test	0.049	11:50
Air Blank	0.000	11:51
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

*SS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:56
Control Test	0.200	11:57
Air Blank	0.000	11:58
Control Test	0.199	11:58
Air Blank	0.000	11:59
Control Test	0.199	11:59
Air Blank	0.000	12:00
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

*SS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53
Control Test	0.078	11:53
Air Blank	0.000	11:54
Control Test	0.079	11:55
Air Blank	0.000	11:55
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet

*SS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 11/18/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:01
Control Test	0.078	12:01
Air Blank	0.000	12:02
Control Test	0.079	12:02
Air Blank	0.000	12:02
Control Test	0.080	12:03
Air Blank	0.000	12:03
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Dry

*SS*

Operator's Signature

2020.11.  
 20  
 11:07:47  
 -05'00"

*SS*

*MX*



INSTRUMENT PROCESSING SHEET

Agency Perry Police Department S/N 80-000811

Florida Department of Law Enforcement Date In 05/28/2020 DI Completion Date 6/6/2020 Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.06.11 13:01:15 -04'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>SD3962</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>																																									
0.080	<u>SD3962</u>	<u>201905B</u> <u>05-14-2021</u>																																									
0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>																																									
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																									
		<b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.9</u> External Digital Therm. ID#: <u>300502</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5090</u>																																									
<b>Calibration Adjustment</b> <u>1012/</u> Performed By <u>SP</u> Barometric Pressure Gauge <u>1014</u> ID # <u>30793</u>		<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>20060</u></td> <td><u>2/10/22</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>20190</u></td> <td><u>4/6/22</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>20160</u></td> <td><u>3/18/22</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>20030</u></td> <td><u>1/21/22</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>03519080A4</u></td> <td><u>4-5-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>20060</u>	<u>2/10/22</u>	0.100	<u>MP5083</u>	<u>20190</u>	<u>4/6/22</u>	0.200	<u>MP5084</u>	<u>20160</u>	<u>3/18/22</u>	0.300	<u>MP5085</u>	<u>20030</u>	<u>1/21/22</u>	0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	<u>MP5091</u>	N/A	N/A																																								
0.040	<u>MP5082</u>	<u>20060</u>	<u>2/10/22</u>																																								
0.100	<u>MP5083</u>	<u>20190</u>	<u>4/6/22</u>																																								
0.200	<u>MP5084</u>	<u>20160</u>	<u>3/18/22</u>																																								
0.300	<u>MP5085</u>	<u>20030</u>	<u>1/21/22</u>																																								
0.080 DGS	N/A	<u>03519080A4</u>	<u>4-5-21</u>																																								
Simulator	Serial Number																																										
0.000	<u>MP5086</u>																																										
Interferent	<u>MP5087</u>																																										
0.050	<u>MP5088</u>																																										
0.080	<u>MP5089</u>																																										
0.200	<u>MP5090</u>																																										
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP5088</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <u>X2</u> <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050	<u>MP5088</u>	<u>201905A</u>	<u>5-14-21</u>																																								
0.080	<u>MP5089</u>	<u>201905B</u>	<u>5-14-21</u>																																								
0.200	<u>MP5090</u>	<u>201904D</u>	<u>4-30-21</u>																																								
0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>																																								
Notes/Suggested Service: <u>Cal Adjust solutions used for first cal adjustment:</u> <u>0.040 Lot# 19080 exp. 3-4-21</u> <u>0.100 Lot# 19160 exp. 7-9-21</u> <u>0.200 Lot# 19040 exp 1-29-21</u> <u>0.300 Lot# 19010 exp 1-3-21</u> <u>BP-SP</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
2020.06.10 13:41:23 -04'00' <u>Michael D. Haughey</u>		2020.06.11 12:58:17 -04'00' <u>Brett Kuchland</u>																																									
Tech Review / Date		Admin Review / Date																																									



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000811, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000811</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PERRY POLICE DEPT</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/06/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>14:06</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

MH

*Shayla Platt*

06/06/2020

Date

SHAYLA D PLATT,

**Department Inspector**

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

BK

2020.06.1  
1 12:58:41  
-04'00"



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY POLICE DEPT  
Time of Inspection: 14:06

Date of Inspection: 06/06/2020

Serial Number: 80-000811  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.047	0.077	0.198	0.079
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.079

Standard Deviations	0.0004	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*MH*

2020.06.  
11  
BK 12:59:01  
-04'00'

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

06/06/2020  
Date



# Stability Checks

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 05/27/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:25
Control Test	0.046	14:26
Air Blank	0.000	14:26
Control Test	0.046	14:27
Air Blank	0.000	14:28
Control Test	0.046	14:28
Air Blank	0.000	14:29
Control Test Stats		
Average	0.0460	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 05/27/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:31
Control Test	0.079	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:33
Air Blank	0.000	14:33
Control Test	0.078	14:34
Air Blank	0.000	14:35
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet



Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 05/27/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.199	14:38
Air Blank	0.000	14:38
Control Test	0.199	14:39
Air Blank	0.000	14:39
Control Test	0.199	14:40
Air Blank	0.000	14:41
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 05/27/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:42
Control Test	0.078	14:43
Air Blank	0.000	14:43
Control Test	0.078	14:43
Air Blank	0.000	14:44
Control Test	0.078	14:44
Air Blank	0.000	14:45
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

dry



Operator's Signature

MH

BK  
 2020.06.  
 11  
 12:59:23  
 -04'00'

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/05/2020 17:26:25

Auto Calibration  
Max Power Res Value = 53  
Auto Range Res Value = 27

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12354, Sum Io = 13080

Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 0.0710 (-0.0390)  
Sample #2 = 0.0540 (-0.0060)  
Sample #3 = 0.0650 (0.0150)  
Sample #4 = 0.0730 (0.0040)  
Avg % Abs = 0.0640 (0.0043)  
STD DEV = 0.0095 (0.0105)  
REL STD DEV = 14.905 (242.399)

Channel 2 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 0.1040 (-0.0130)  
Sample #2 = 0.1110 (-0.0040)  
Sample #3 = 0.1170 (0.0050)  
Sample #4 = 0.0830 (0.0030)  
Avg % Abs = 0.1037 (0.0013)  
STD DEV = 0.0181 (0.0047)  
REL STD DEV = 17.506 (354.436)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12345, Sum Io = 13075

Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 0.7370 (-0.0170)  
Sample #2 = 0.7800 (-0.0250)  
Sample #3 = 0.7620 (0.0070)  
Sample #4 = 0.7620 (0.0230)  
Avg % Abs = 0.7680 (0.0017)  
STD DEV = 0.0104 (0.0244)  
REL STD DEV = 1.353 (1466.424)

Channel 2 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 1.5340 (-0.0130)  
Sample #2 = 1.5260 (-0.0020)  
Sample #3 = 1.5410 (-0.0020)  
Sample #4 = 1.5670 (0.0000)  
Avg % Abs = 1.5447 (-0.0013)  
STD DEV = 0.0207 (0.0012)  
REL STD DEV = 1.343 (86.603)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12336, Sum Io = 13068

Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 5.0230 (-0.0170)  
Sample #2 = 5.0210 (-0.0120)  
Sample #3 = 5.0110 (0.0230)  
Sample #4 = 5.0130 (0.0290)  
Avg % Abs = 5.0150 (0.0133)  
STD DEV = 0.0053 (0.0221)  
REL STD DEV = 0.106 (166.076)

Channel 2 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 10.1390 (-0.0190)  
Sample #2 = 10.1030 (0.0050)  
Sample #3 = 10.1530 (0.0010)  
Sample #4 = 10.1160 (0.0270)  
Avg % Abs = 10.1240 (0.0110)  
STD DEV = 0.0259 (0.0140)  
REL STD DEV = 0.256 (127.273)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12330, Sum Io = 13065

Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 3.4500 (-0.0150)  
Sample #2 = 3.4510 (0.0030)  
Sample #3 = 3.4310 (0.0070)  
Sample #4 = 3.4380 (0.0100)  
Avg % Abs = 3.4400 (0.0067)  
STD DEV = 0.0101 (0.0035)  
REL STD DEV = 0.295 (52.578)

Channel 2 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 7.0100 (-0.0130)  
Sample #2 = 7.0040 (0.0140)  
Sample #3 = 7.0170 (-0.0060)  
Sample #4 = 7.0060 (0.0120)  
Avg % Abs = 7.0090 (0.0067)  
STD DEV = 0.0070 (0.0110)  
REL STD DEV = 0.100 (165.227)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12325, Sum Io = 13060

Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 1.7630 (-0.0270)  
Sample #2 = 1.7560 (0.0040)  
Sample #3 = 1.7640 (0.0120)  
Sample #4 = 1.7810 (0.0250)  
Avg % Abs = 1.7670 (0.0137)  
STD DEV = 0.0128 (0.0106)  
REL STD DEV = 0.723 (77.552)

Channel 2 header

Sample % Abs (% Abs Ref)  
Sample #1 = 3.6170 (-0.0240)  
Sample #2 = 3.6050 (-0.0070)  
Sample #3 = 3.6210 (0.0000)  
Sample #4 = 3.6430 (-0.0060)  
Avg % Abs = 3.6230 (-0.0043)  
STD DEV = 0.0191 (0.0038)  
REL STD DEV = 0.527 (87.368)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*

Channel 1 data table with 3 columns: Sol Val, % Abs, (% Abs Ref).  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.064  
Std Dev = 0.01 Rel Std Dev = 14.91  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.768  
Std Dev = 0.01 Rel Std Dev = 1.35  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 5.015  
Std Dev = 0.01 Rel Std Dev = 0.11  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.440  
Std Dev = 0.01 Rel Std Dev = 0.30  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 1.767  
Std Dev = 0.01 Rel Std Dev = 0.72  
Zero Order Coef = -1494.55  
First Order Coef = 9826.78  
Second Order Coef = -1736.51  
Standard Deviation = 2769.147705

Channel 2 data table with 3 columns: Sol Val, % Abs, (% Abs Ref).  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.104  
Std Dev = 0.02 Rel Std Dev = 17.51  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.545  
Std Dev = 0.02 Rel Std Dev = 1.34  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 10.124  
Std Dev = 0.03 Rel Std Dev = 0.26  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 7.009  
Std Dev = 0.01 Rel Std Dev = 0.10  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 3.623  
Std Dev = 0.02 Rel Std Dev = 0.53  
Zero Order Coef = -1436.20  
First Order Coef = 4848.70  
Second Order Coef = -424.63  
Standard Deviation = 2724.838135

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.018	0.0183
0.040	0.106	-0.0656
0.100	0.086	0.0136
0.200	0.247	-0.0470
0.300	0.219	0.0806

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.020	0.0197
0.040	0.106	-0.0658
0.100	0.087	0.0133
0.200	0.245	-0.0454
0.300	0.222	0.0783

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Channel 1 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 7961.00  
Sample #2 = 8255.00  
Sample #3 = 8217.00  
Sample #4 = 8125.00  
Average Result = 8199.0000  
STD DEV = 66.8431  
REL STD DEV = 0.815  
Channel 2 data table with 3 columns: Sample, % Abs, (% Abs Ref).  
Sample #1 = 8508.00  
Sample #2 = 8590.00  
Sample #3 = 8547.00  
Sample #4 = 8527.00  
Average Result = 8554.6670  
STD DEV = 32.1921  
REL STD DEV = 0.376  
Dry Gas H2O Adjust Results  
Barometric Pressure = 1012  
3 um H2O Adjust (mg/l\*10,000) = -4389  
9 um H2O Adjust (mg/l\*10,000) = -4744  
\*\*\*\* AUTO CAL PASS

Cal Adjustment #1  
Sims not prepared properly.  
Will repeat cal adjustment.  
SP

MH  
BK  
2020.06.11  
12:59:51  
-04'00'

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/06/2020 10:02:08

Auto Calibration  
Max Power Res Value = 53  
Auto Range Res Value = 30

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12822, Sum hi = 13320

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0800 (0.0130)  
Sample #2 = 0.0520 (0.0430)  
Sample #3 = 0.0810 (0.0590)  
Sample #4 = 0.0990 (0.0490)  
Avg % Abs = 0.0773 (0.0503)  
STD DEV = 0.0237 (0.0081)  
REL STD DEV = 30.664 (16.059)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1080 (-0.0030)  
Sample #2 = 0.0990 (-0.0070)  
Sample #3 = 0.0980 (0.0070)  
Sample #4 = 0.0950 (0.0020)  
Avg % Abs = 0.0973 (0.0007)  
STD DEV = 0.0021 (0.0071)  
REL STD DEV = 2.139 (1064.190)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12808, Sum hi = 13315

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.7420 (-0.0160)  
Sample #2 = 0.7590 (-0.0260)  
Sample #3 = 0.7570 (-0.0160)  
Sample #4 = 0.7770 (-0.0120)  
Avg % Abs = 0.7643 (-0.0180)  
STD DEV = 0.0110 (0.0072)  
REL STD DEV = 1.441 (40.062)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5360 (-0.0040)  
Sample #2 = 1.5320 (-0.0150)  
Sample #3 = 1.5350 (-0.0110)  
Sample #4 = 1.5350 (-0.0160)  
Avg % Abs = 1.5340 (-0.0140)  
STD DEV = 0.0017 (0.0026)  
REL STD DEV = 0.113 (18.898)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12803, Sum hi = 13313

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.7550 (-0.0140)  
Sample #2 = 1.7840 (-0.0110)  
Sample #3 = 1.7670 (-0.0100)  
Sample #4 = 1.7940 (0.0020)  
Avg % Abs = 1.7817 (-0.0063)  
STD DEV = 0.0137 (0.0072)  
REL STD DEV = 0.766 (114.224)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12802, Sum hi = 13313

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4610 (-0.0300)  
Sample #2 = 3.4070 (0.0140)  
Sample #3 = 3.4360 (0.0160)  
Sample #4 = 3.4010 (0.0200)  
Avg % Abs = 3.4147 (0.0167)  
STD DEV = 0.0187 (0.0031)  
REL STD DEV = 0.548 (18.330)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12802, Sum hi = 13312

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 4.9990 (-0.0080)  
Sample #2 = 4.9750 (0.0020)  
Sample #3 = 4.9820 (0.0170)  
Sample #4 = 4.9520 (0.0170)  
Avg % Abs = 4.9697 (0.0120)  
STD DEV = 0.0157 (0.0087)  
REL STD DEV = 0.316 (72.169)

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.0770 (-0.0190)  
Sample #2 = 10.0250 (0.0110)  
Sample #3 = 10.0270 (0.0240)  
Sample #4 = 10.0130 (0.0220)  
Avg % Abs = 10.0217 (0.0190)  
STD DEV = 0.0076 (0.0070)  
REL STD DEV = 0.076 (36.842)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.077  
Std Dev = 0.02 Rel Std Dev = 30.66  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.764  
Std Dev = 0.01 Rel Std Dev = 1.44  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.782  
Std Dev = 0.01 Rel Std Dev = 0.77  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.415  
Std Dev = 0.02 Rel Std Dev = 0.55  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 4.970  
Std Dev = 0.02 Rel Std Dev = 0.32  
Zero Order Coef = -201.80  
First Order Coef = 2711.31  
Second Order Coef = 40.87  
Standard Deviation = 8.398795

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.097  
Std Dev = 0.00 Rel Std Dev = 2.14  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.534  
Std Dev = 0.00 Rel Std Dev = 0.11  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.653  
Std Dev = 0.00 Rel Std Dev = 0.04  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.975  
Std Dev = 0.00 Rel Std Dev = 0.07  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.022  
Std Dev = 0.01 Rel Std Dev = 0.08  
Zero Order Coef = -103.71  
First Order Coef = 1268.14  
Second Order Coef = 16.62  
Standard Deviation = 21.983315

MA  
BK  
2020.06.11  
13:00:13  
-04'00"

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0002
0.100	0.100	0.0001
0.200	0.200	-0.0002
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.039	0.0005
0.100	0.100	0.0002
0.200	0.201	-0.0006
0.300	0.300	0.0002

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3168.00  
Sample #2 = 3165.00  
Sample #3 = 3165.00  
Sample #4 = 3157.00  
Average Result = 3162.3333  
STD DEV = 4.6188  
REL STD DEV = 0.146  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3355.00  
Sample #2 = 3380.00  
Sample #3 = 3371.00  
Sample #4 = 3382.00  
Average Result = 3377.6667  
STD DEV = 5.8595  
REL STD DEV = 0.173  
\*\*\*\*\* DRY GAS H2O ADJUST RESULTS \*\*\*\*\*  
Barometric Pressure = 1014  
3 um H2O Adjust (mg/l\*10,000) = 647  
9 um H2O Adjust (mg/l\*10,000) = 432  
\*\*\*\* AUTO CAL PASS

Cal Adjustment #2 SP

Post Cal Adjust  
Stability Checks

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/06/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.048	11:11
Air Blank	0.000	11:11
Control Test	0.048	11:12
Air Blank	0.000	11:13
Control Test	0.048	11:13
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/06/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:15
Control Test	0.078	11:15
Air Blank	0.000	11:16
Control Test	0.077	11:17
Air Blank	0.000	11:17
Control Test	0.078	11:18
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

Operator's Signature

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/06/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:23
Control Test	0.199	11:24
Air Blank	0.000	11:24
Control Test	0.197	11:25
Air Blank	0.000	11:25
Control Test	0.198	11:26
Air Blank	0.000	11:27
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

SP

Operator's Signature

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
06/06/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:20
Control Test	0.080	11:20
Air Blank	0.000	11:20
Control Test	0.080	11:21
Air Blank	0.000	11:21
Control Test	0.080	11:22
Air Blank	0.000	11:22
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

SP

Operator's Signature

MA

BK  
2020.06.  
11  
13:00:36  
-04'00"





INSTRUMENT PROCESSING SHEET

Agency Perry Police Department s/N 80-00811

Florida Department of Law Enforcement

Date In 02/10/2020 DI Completion Date 2/12/20 Ship P/U H/D CMI EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		<b>Quality Checks</b> Performed By <u>JS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>189</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
<b>Final Release Date</b> <p style="text-align: center;">FDLE FEB 17 2020 Alcohol Testing Program</p>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD1012</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>DR1279</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>SD1011</td> <td>201904D 07-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AGA16501 06-14-2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD1012	201905A 05-14-2021	0.080	DR1279	201905B 05-14-2021	0.200	SD1011	201904D 07-30-2021	0.080 DGS	N/A	AGA16501 06-14-2021	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	SD1012	201905A 05-14-2021																																																													
0.080	DR1279	201905B 05-14-2021																																																													
0.200	SD1011	201904D 07-30-2021																																																													
0.080 DGS	N/A	AGA16501 06-14-2021																																																													
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1017</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5086																																																														
Interferent	MP5087																																																														
0.050	MP5088																																																														
0.080	MP5089																																																														
0.200	MP5090																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
		Tech Review / Date <u>PPM 2/13/20</u> Admin Review / Date <u>Brett Kirkland 2/14/2020</u>																																																													





Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000811, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000811</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PERRY POLICE DEPT</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/12/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>14:41</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/12/2020 Date  
Shayla Platt  
 SHAYLA D PLATT,  
 Department Inspector

FDLE/ATP Form 69 January 2020  
 Issuing Authority: Alcohol Testing Program  
 Service • Integrity • Respect • Quality  
 Page 1 of 1

*Handwritten:* [Signature] TSK 2/14/2020

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY POLICE DEPT  
Time of Inspection: 14:41

Date of Inspection: 02/12/2020

Serial Number: 80-000811  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.080	0.200	0.078
0.000	0.048	0.080	0.200	0.078
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.081	0.199	0.078
0.000	0.049	0.081	0.199	0.079
0.000	0.049	0.081	0.199	0.078
0.000	0.049	0.081	0.199	0.078

Standard Deviations	0.0003	0.0008	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

ROOM  
TBK  
2/14/2020

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

02/12/2020  
Date

# Stability Checks

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 02/10/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:17
Control Test	0.049	14:18
Air Blank	0.000	14:19
Control Test	0.048	14:19
Air Blank	0.000	14:20
Control Test	0.048	14:21
Air Blank	0.000	14:21
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

*DS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 02/10/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:23
Control Test	0.078	14:24
Air Blank	0.000	14:24
Control Test	0.078	14:25
Air Blank	0.000	14:26
Control Test	0.078	14:26
Air Blank	0.000	14:27
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

*DS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 02/10/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:28
Control Test	0.199	14:29
Air Blank	0.000	14:29
Control Test	0.200	14:30
Air Blank	0.000	14:30
Control Test	0.200	14:31
Air Blank	0.000	14:32
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

*DS*

Operator's Signature

PERRY POLICE DEPT  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000811  
 02/10/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:34
Control Test	0.079	14:34
Air Blank	0.000	14:35
Control Test	0.078	14:35
Air Blank	0.000	14:36
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry

*DS*

Operator's Signature

*DSM*  
*BSK*  
*2/14/2020*



**Return Material Authorization**

**Ship to:**  CMI, Inc.  
 Enforcement Electronics

Shipment to repair facility authorized by: Sgt. Lin Gray on 11/21/2019

**Items Returned:** Instrument  Supplies  Other  Describe: \_\_\_\_\_  
Instrument Model: 8000 Serial Number: 80-000811

**Bill To Address:**  
Sgt. Lin Gray  
Perry Police Department  
211 S. Washington Street  
Perry, FL 32347

**Ship to Address:**  
FDLE Off-Site Mail Facility  
c/o Florida Department of Law Enforcemen  
Alcohol Testing Program  
813B Lake Bradford Road  
Tallahassee FL 32304

**Reason for Return:**

Display flashes, then blanks, then instrument returns to red light on, power  
supply problem. This instrument was sent to you just a little while ago for a broken phone  
line socket. When it was received, it was inoperative. Owner almost certainly thinks this will  
be warranty work

**Please choose one of the following options:**

- 1. I \_\_\_\_\_, authorize all repairs.
- 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sgt. Lin Gray  
Phone #: 850-838-6596 Email: lin.gray@perrypolice.net

ATP Contact Name: Patrick Murphy ATP Email: patrickmurphy@fdle.state.fl.us

*Handwritten:*  
BGM  
13K  
7/14/2020



# INSTRUMENT PROCESSING SHEET

Agency Perry PDs/N 80-000811Florida Department of  
Law EnforcementDate In 10/28/2019 DI Completion Date \_\_\_\_\_ Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By _____ <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By _____ <input type="checkbox"/> Lab Temp °C _____ External Digital Therm. ID#: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____ <input type="checkbox"/> 34°C +- .2 Serial #: _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
<b>Final Release Date</b> _____ _____	<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Notes/Suggested Service: _____ _____ _____ _____ _____	<b>Department Inspection</b> Performed By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.050</td><td></td></tr> <tr><td>0.080</td><td></td></tr> <tr><td>0.200</td><td></td></tr> </tbody> </table> <b>Attachments</b> <input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																																					
Simulator	Serial Number																																																	
0.000																																																		
Interferent																																																		
0.050																																																		
0.080																																																		
0.200																																																		

QDM  
 TSK  
 2/19/2021



**Return Material Authorization**

**Ship to:**  CMI, Inc.  
 Enforcement Electronics

Shipment to repair facility authorized by: Sgt. Lin Gray on 11/26/2018

Items Returned: Instrument  Supplies  Other  Describe: \_\_\_\_\_

Instrument Model: 8000 Serial Number: 80-000811

Bill To Address:  
Sgt. Lin Gray  
Perry Police Department  
211 S. Washington Street  
Perry, FL 32347

Ship to Address:  
FDLE Off-Site Mail Facility  
c/o Florida Department of Law Enforcement  
Alcohol Testing Program  
813B Lake Bradford Road  
Tallahassee FL 32304

**Reason for Return:**

Broken phone connector on back panel. Please replace.

**Please choose one of the following options:**

- 1. I \_\_\_\_\_, authorize all repairs.
- 2. I \_\_\_\_\_, authorize repairs up to \$ \_\_\_\_\_.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sgt. Lin Gray

Phone #: 850-838-6596 Email: lin.gray@perry.police.net

ATP Contact Name: Patrick Murphy ATP Email: patrickmurphy@fdle.state.fl.us

*909m*  
*BK*  
*2/14/2020*



### INSTRUMENT PROCESSING SHEET

Agency Perry PD S/N 80-000811

Florida Department of Law Enforcement Date In 11/9/18 DI Completion Date \_\_\_\_\_  Ship  P/U  H/D  CMI  EE

Intake	Performed By <u>SQC</u>	Quality Checks	Performed By _____	Flow Calibration	Performed By _____																																																		
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____		<input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr style="background-color: #cccccc;"> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td></tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)		<table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr style="background-color: #cccccc;"> <th colspan="2">Maintenance</th> <th>Performed By _____</th> </tr> </thead> <tbody> <tr><td><input type="checkbox"/> Battery Replacement</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> Other _____</td><td></td><td></td></tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr style="background-color: #cccccc;"> <th colspan="2">Temperature Checks</th> <th>Performed By _____</th> </tr> </thead> <tbody> <tr><td><input type="checkbox"/> Lab Temp °C _____</td><td></td><td></td></tr> <tr><td>External Digital Therm. ID#: _____</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> 34°C +/- .2 Serial #: _____</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> 34°C +/- .2 Serial #: _____</td><td></td><td></td></tr> <tr><td><input type="checkbox"/> 34°C +/- .2 Serial #: _____</td><td></td><td></td></tr> </tbody> </table>		Maintenance		Performed By _____	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other _____			Temperature Checks		Performed By _____	<input type="checkbox"/> Lab Temp °C _____			External Digital Therm. ID#: _____			<input type="checkbox"/> 34°C +/- .2 Serial #: _____			<input type="checkbox"/> 34°C +/- .2 Serial #: _____			<input type="checkbox"/> 34°C +/- .2 Serial #: _____		
Simulator	Serial #	Lot #/Exp																																																					
0.050																																																							
0.080																																																							
0.200																																																							
0.080 DGS	N/A																																																						
Maintenance		Performed By _____																																																					
<input type="checkbox"/> Battery Replacement																																																							
<input type="checkbox"/> Dry Gas Regulator Replacement																																																							
<input type="checkbox"/> Breath Tube Replacement																																																							
<input type="checkbox"/> Other _____																																																							
Temperature Checks		Performed By _____																																																					
<input type="checkbox"/> Lab Temp °C _____																																																							
External Digital Therm. ID#: _____																																																							
<input type="checkbox"/> 34°C +/- .2 Serial #: _____																																																							
<input type="checkbox"/> 34°C +/- .2 Serial #: _____																																																							
<input type="checkbox"/> 34°C +/- .2 Serial #: _____																																																							
Final Release Date																																																							

Calibration Adjustment	Performed By _____	Department Inspection	Performed By _____																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #cccccc;"> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #cccccc;"> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #cccccc;"> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td></tr> <tr><td>Interferent</td><td></td></tr> <tr><td>0.050</td><td></td></tr> <tr><td>0.080</td><td></td></tr> <tr><td>0.200</td><td></td></tr> </tbody> </table>		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000																																																															
Interferent																																																															
0.050																																																															
0.080																																																															
0.200																																																															
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #cccccc;"> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Form 41</td> <td><input type="checkbox"/> Post-Stability Checks</td> </tr> <tr> <td><input type="checkbox"/> Stability Checks</td> <td><input type="checkbox"/> Flow Calibration</td> </tr> <tr> <td><input type="checkbox"/> Calibration Certificate</td> <td><input type="checkbox"/> Form 40</td> </tr> <tr> <td><input type="checkbox"/> Calibration Adjustment</td> <td><input type="checkbox"/> Other _____</td> </tr> </tbody> </table>		Attachments		<input type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																																				
Attachments																																																															
<input type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks																																																														
<input type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																																														
<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																																														
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																																														

Notes/Suggested Service: _____ _____ _____ _____ _____	<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  _____ Tech Review / Date                      Admin Review / Date
--	--

089M  
 13K  
 2/14/2019

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY POLICE DEPT  
Time of Inspection: 13:39

Date of Inspection: 11/26/2018

Serial Number: 80-000811  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
SKIPPED AI TO OPERATE INSTRUMENT

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PJM  
TBK  
2/14/2021

PATRICK J MURPHY

\_\_\_\_\_  
Signature and Printed Name

11/26/2018  
Date



CMI Inc  
 316 E 9th Street  
 Owensboro KY 42303  
 USA  
 Phone: 866-835-0690  
 Fax: 270-685-6268

Date: 12/12/2018  
 Page: 1 of 2

**Service Estimate: 402980**

PO #:

<b>Ship To:</b> ATP FDLE OFF-SITE MAIL FACILITY 813-B LAKE BRADFORD ROAD TALLAHASSEE FL 32304 USA	<b>Customer Number:</b> 323480	<b>Technician Name:</b> Michael Veeck
--	--------------------------------	--

BT, SHELF

Call Line	Part Number/Description	Revision	Quantity	Service Call Type
1	002480FL UNIT ASSY,I8000	ND	1.00 EA	Out of Warranty

Job: SRV4029800001

Serial Number(s): **80-000811**

**Job Material**

Seq. No.	Part	Description	Quantity	Est. Unit Price	Est. Ext. Price
30	402565	O-RING,NEOPRENE, .301IDx.06THK	1.00	EA 0.25	0.25
80	650517	CERTIFICATE OF CALIBRATION CMI	2.00	EA 0.00	0.00
90	441169	COVER DUST,5/8" x 1/2" McMASTE	1.00	EA 0.00	0.00
100	470145	CAP,PLASTIC, .25IDx.50LG,RED ST	3.00	EA 0.00	0.00
110	470154	CAP,PLSTC, .406IDx.438-.562,RED	1.00	EA 0.00	0.00
120	021307FL	HOSE ASSY,BREATH,I8000	1.00	EA 250.00	250.00
130	471201	O-RING,BUNA-N,-008,3/16inx5/16	1.00	EA 1.95	1.95
140	021354FL	CABLE ASSY,PHONE	1.00	EA 26.50	26.50
150	021357FL	SOLENOID ASSY,I8000	1.00	EA 158.80	158.80
160	690040	BATTERY,3VOLT,LITHIUM,CR2032	1.00	EA 4.99	4.99

**Labor, Freight, & Misc. Charges**

Misc Code	Description	Amount
LABR	Service Repair Labor	212.50
LCAL	Service - Calibration Adjust	45.00
LFT	Service - Cal / Final Test	85.00
SVFT	Freight - Service	35.00

**Total: 819.99**

**NOTE: AN EVALUATION FEE WILL APPLY TO ESTIMATES THAT ARE NOT REPAIRED.**  
 PLEASE SIGN AND FAX TO: 270-685-6268

APPROVED BY: \_\_\_\_\_ DATE: \_\_\_\_\_

**\*TAX NOT INCLUDED**

*QDM*

CMI Inc  
316 E 9th Street  
Owensboro KY 42303  
USA  
Phone: 866-835-0690  
Fax: 270-685-6268

Date: 12/12/2018

Page: 2 of 2

PO #:

**Ship To:**

ATP  
FDLE OFF-SITE MAIL FACILITY  
813-B LAKE BRADFORD ROAD  
TALLAHASSEE FL 32304  
USA

**Service Estimate: 402980**  
**Serial Number(s): 80-000811**

*Adam*