



INSTRUMENT PROCESSING SHEET

Florida Department of
Law Enforcement

Agency Miami-Dade Police Department

S/N 80-000882

Date In 08/12/2020 DI Completion Date 08/12/2020

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DERR</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DERR</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>167</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 1/30/2022</td> </tr> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 1/30/2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DERR</u> ☒ Lab Temp °C <u>22.75C</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/- .2 Serial #: <u>SD3967</u> ☒ 34°C +/- .2 Serial #: <u>SD3968</u> ☒ 34°C +/- .2 Serial #: <u>SD3969</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	SD3967	201905A 05/14/2021																																																
0.080	SD3968	201905B 05/14/2021																																																
0.200	SD3969	201904D 04/30/2021																																																
0.080 DGS	N/A	AG003005 1/30/2022																																																
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.08.19 08:10:07 -04'00'	Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1006</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Michael D. Hagley</u> 2020.08.17 14:55:47 -04'00' <u>Brett Kirkland</u> 2020.08.19 08:07:21 -04'00' Tech Review / Date Admin Review / Date		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969																																				
Simulator	Serial Number																																																	
0.000	SD3965																																																	
Interferent	SD3966																																																	
0.050	SD3967																																																	
0.080	SD3968																																																	
0.200	SD3969																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 13:41

Date of Inspection: 08/12/2020

Serial Number: 80-000882
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.048	0.080	0.199	0.080
0.000	0.048	0.080	0.199	0.079
0.000	0.052	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.080

Standard Deviations	0.0011	0.0004	0.0004	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

MX

BK 2020.08.19
08:07:57 -04'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

08/12/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000882	Miami-Dade Police Department	08/12/2020	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒
MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27	MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27	MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27	MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27 MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2020 Software: 8100.27
Test g/210L Time Air Blank 0.000 10:09 Control Test 0.050 10:09 Air Blank 0.000 10:10 Control Test 0.049 10:11 Air Blank 0.000 10:11 Control Test 0.050 10:12 Air Blank 0.000 10:13 Control Test Stats Average 0.0457 Std Dev 0.0006 Rel Std Dev(%) 1.1625 <i>[Signature]</i> Operator's Signature	Test g/210L Time Air Blank 0.000 10:14 Control Test 0.081 10:15 Air Blank 0.000 10:16 Control Test 0.079 10:16 Air Blank 0.000 10:17 Control Test 0.080 10:18 Air Blank 0.000 10:18 Control Test Stats Average 0.0850 Std Dev 0.0010 Rel Std Dev(%) 1.2500 <i>[Signature]</i> Operator's Signature	Test g/210L Time Air Blank 0.000 10:20 Control Test 0.199 10:21 Air Blank 0.000 10:22 Control Test 0.199 10:22 Air Blank 0.000 10:23 Control Test 0.198 10:24 Air Blank 0.000 10:24 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906 <i>[Signature]</i> Operator's Signature	Test g/210L Time Air Blank 0.000 10:27 Control Test 0.079 10:27 Air Blank 0.000 10:28 Control Test 0.079 10:28 Air Blank 0.000 10:28 Control Test 0.080 10:29 Air Blank 0.000 10:29 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277 <i>[Signature]</i> Operator's Signature

MX
BK
2020.08.19
08:08:24
-0400



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000882, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000882</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/12/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>13:41</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

08/12/2020
Date

David Reyes-Rivera
DAVID E REYES-RIVERA,
Department Inspector

Service • Integrity • Respect • Quality