FDLE
Florida Department of Law Enforcement

# INSTRUMENT PROCESSING SHEET

Agency Pinellas County S/N 80-000889

Date In 01/08/2020 DI Completion Date 1/22/20 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

**Intake** Performed By DP

- ☒ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ☒ Case
- ☒ Keyboard
- ☒ Feet
- ☒ Ports
- ☒ Handle
- ☒ Dry Gas Shelf
- ☒ Breath Tube
- ☒ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord
- ☐ Printer Cable
- ☐ Static Bag
- ☐ 12V DC Cable

Notes: _____

**Final Release Date**

FDLE

JAN 27 2020

Alcohol Testing Program

**Quality Checks** Performed By DS

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 198
- ☒ Flow Verification (L/s)

Flow Column # ATP-102

- 32 mm 0.148 (.139 - .169)
- 36 mm 0.167 (.156 - .190)
- 53 mm 0.234 (.228 - .278)
- 103 mm 0.500 (.447 - .547)

☒ Barometric Pressure Check

Gauge ID # 28427

☒ Stability Checks

| Simulator | Serial # | Lot #/Exp |
|---|---|---|
| 0.050 | MP5088 | 201905A |
| | | 05-14-2021 |
| 0.080 | MP5089 | 201905B |
| | | 05-14-2021 |
| 0.200 | MP5090 | 201904D |
| | | 04-30-2021 |
| 0.080 DGS | N/A | AG916501 |
| | | 06-14-2021 |

**Flow Calibration** Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm

☐ R-Value _____

☐ Post Calibration Verification (L/s)

Flow Column # _____

- 32 mm _____ (.139 - .169)
- 36 mm _____ (.156 - .190)
- 53 mm _____ (.228 - .278)
- 103 mm _____ (.447 - .547)

**Maintenance** Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

**Temperature Checks** Performed By SP

- ☒ Lab Temp °C 21.2

External Digital Therm. ID#: 300505

- ☒ 34°C +-.2 Serial #: MP5088
- ☒ 34°C +-.2 Serial #: MP5089
- ☒ 34°C +-.2 Serial #: MP5090

**Calibration Adjustment** Performed By SP

Barometric Pressure Gauge 1027 ID # 26932

| Simulator | Serial Number | Lot Number | Expiration |
|---|---|---|---|
| 0.000 | MP5091 | N/A | N/A |
| 0.040 | MP5082 | 19080 | 3-4-21 |
| 0.100 | MP5083 | 19140 | 7-9-21 |
| 0.200 | MP5084 | 19040 | 1-29-21 |
| 0.300 | MP5085 | 19010 | 1-3-21 |
| 0.080 DGS | N/A | 08819080A1 | 6-5-21 |

☒ Post Calibration Adjustment Stability Checks

| Simulator | Serial Number | Lot Number | Expiration |
|---|---|---|---|
| 0.050 | MP5088 | 201905A | 5-14-21 |
| 0.080 | MP5089 | 201905B | 5-14-21 |
| 0.200 | MP5090 | 201904D | 4-30-21 |
| 0.080 DGS | N/A | AG916501 | 6-14-21 |

**Department Inspection** Performed By SP

Barometric Pressure ID# 28421

Gauge ~~SP 1027~~ 1024 Instrument 1022

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2019-A

| Simulator | Serial Number |
|---|---|
| 0.000 | MP5086 |
| Interferent | MP5087 |
| 0.050 | MP5088 |
| 0.080 | MP5089 |
| 0.200 | MP5090 |

**Attachments**

- ☒ Form 41
- ☒ Stability Checks
- ☒ Calibration Certificate
- ☒ Calibration Adjustment
- ☒ Post-Stability Checks
- ☐ Flow Calibration
- ☒ Form 40
- ☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

RJM 1/23/20 Brett Kirkland 1/27/2020

Tech Review / Date Admin Review / Date

# Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000889 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

| | |
|---|---|
| Serial Number: | 80-000889 |
| Owning Agency: | PINELLAS COUNTY SO |
| Calibration Date: | 01/22/2020 |
| Calibration Time: | 11:01 |

| | UNCERTAINTY* ± |
|---|---|
| 0.050 g/ 210 L | 0.004 |
| 0.080 g/ 210 L | 0.005 |
| 0.200 g/ 210 L | 0.007 |
| 0.080 g/ 210 L Dry Gas Control | 0.005 |

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/22/2020
Date

Shayla Platt
SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 July 2018
Issuing Authority: Alcohol Testing Program

*Service • Integrity • Respect • Quality*



PDM
BK 1/27/2020

# Florida Department of Law Enforcement
# Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 11:01
Date of Inspection: 01/22/2020
Serial Number: 80-000889
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|---|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021 | 0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021 | 0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021 |
|---|---|---|---|---|
| 0.000 | 0.051 | 0.082 | 0.206 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.206 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.206 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.206 | 0.080 |
| 0.000 | 0.051 | 0.082 | 0.206 | 0.080 |
| 0.000 | 0.051 | 0.082 | 0.205 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.205 | 0.080 |
| 0.000 | 0.051 | 0.082 | 0.206 | 0.079 |
| 0.000 | 0.051 | 0.082 | 0.205 | 0.079 |
| 0.000 | 0.051 | 0.082 | 0.205 | 0.079 |
| Standard Deviations | 0.0000 | 0.0005 | 0.0005 | 0.0005 |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

PDM
BK
1/27/2020

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
SHAYLA D PLATT
Signature and Printed Name

01/22/2020
Date

*FDLE/ATP Form 41 –Revised August 2005*

# Stability Checks

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/10/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:51 |
| Control Test | 0.047 | 08:51 |
| Air Blank | 0.000 | 08:52 |
| Control Test | 0.049 | 08:53 |
| Air Blank | 0.000 | 08:53 |
| Control Test | 0.050 | 08:54 |
| Air Blank | 0.000 | 08:55 |

Control Test Stats

| | |
|---|---|
| Average | 0.0487 |
| Std Dev | 0.0015 |
| Rel Std Dev(%) | 3.1388 |

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/10/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 08:59 |
| Control Test | 0.082 | 08:59 |
| Air Blank | 0.000 | 09:00 |
| Control Test | 0.080 | 09:01 |
| Air Blank | 0.000 | 09:01 |
| Control Test | 0.080 | 09:02 |
| Air Blank | 0.000 | 09:03 |

Control Test Stats

| | |
|---|---|
| Average | 0.080 |
| Std Dev | 0.001 |
| Rel Std Dev(%) | 1.431 |

wet

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/10/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 09:08 |
| Control Test | 0.205 | 09:09 |
| Air Blank | 0.000 | 09:10 |
| Control Test | 0.203 | 09:10 |
| Air Blank | 0.000 | 09:11 |
| Control Test | 0.201 | 09:12 |
| Air Blank | 0.000 | 09:12 |

Control Test Stats

| | |
|---|---|
| Average | 0.203 |
| Std Dev | 0.002 |
| Rel Std Dev(%) | 0.985 |

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/10/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 09:13 |
| Control Test | 0.077 | 09:14 |
| Air Blank | 0.000 | 09:14 |
| Control Test | 0.078 | 09:14 |
| Air Blank | 0.000 | 09:15 |
| Control Test | 0.079 | 09:15 |
| Air Blank | 0.000 | 09:16 |

Control Test Stats

| | |
|---|---|
| Average | 0.078 |
| Std Dev | 0.001 |
| Rel Std Dev(%) | 1.282 |

Dry

Operator's Signature

PDM
BK 1/27/2020

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/17/2020 08:44:26

Auto Calibration
Max Power Res Value = 60
Auto Range Res Value = 52

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12805, 9um Io = 13823

<<<<< CHANNEL 1 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 0.1380 | (-0.0140) |
| Sample #2 = | 0.1690 | (-0.0010) |
| Sample #3 = | 0.1780 | (-0.0220) |
| Sample #4 = | 0.1520 | (0.0270) |

Avg % Abs = 0.1663 (0.0013)
STD DEV = 0.0132 (0.0246)
REL STD DEV = 7.938 (1843.739)

---

<<<<< CHANNEL 2 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 0.1180 | (-0.0170) |
| Sample #2 = | 0.1090 | (-0.0010) |
| Sample #3 = | 0.1410 | (-0.0190) |
| Sample #4 = | 0.1160 | (0.0140) |

Avg % Abs = 0.1220 (-0.0020)
STD DEV = 0.0168 (0.0155)
REL STD DEV = 13.789 (826.136)

---

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12804, 9um Io = 13826

<<<<< CHANNEL 1 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 0.9320 | (-0.0340) |
| Sample #2 = | 0.8530 | (0.0390) |
| Sample #3 = | 0.8820 | (0.0470) |
| Sample #4 = | 0.8820 | (0.0470) |

Avg % Abs = 0.8723 (0.0443)
STD DEV = 0.0167 (0.0046)
REL STD DEV = 1.919 (10.418)

---

<<<<< CHANNEL 2 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 1.5760 | (-0.0140) |
| Sample #2 = | 1.5130 | (0.0600) |
| Sample #3 = | 1.5060 | (0.0630) |
| Sample #4 = | 1.5140 | (0.0690) |

Avg % Abs = 1.5110 (0.0640)
STD DEV = 0.0044 (0.0046)
REL STD DEV = 0.288 (7.160)

---

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12800, 9um Io = 13821

<<<<< CHANNEL 1 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 1.9840 | (-0.0110) |
| Sample #2 = | 1.9510 | (0.0270) |
| Sample #3 = | 1.9280 | (0.0400) |
| Sample #4 = | 1.9130 | (0.0470) |

Avg % Abs = 1.9307 (0.0380)
STD DEV = 0.0191 (0.0101)
REL STD DEV = 0.991 (26.708)

---

<<<<< CHANNEL 2 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 3.6100 | (-0.0030) |
| Sample #2 = | 3.5390 | (0.0730) |
| Sample #3 = | 3.5240 | (0.0830) |
| Sample #4 = | 3.5160 | (0.0790) |

Avg % Abs = 3.5263 (0.0783)
STD DEV = 0.0117 (0.0050)
REL STD DEV = 0.331 (6.425)

---

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12799, 9um Io = 13819

<<<<< CHANNEL 1 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 3.7850 | (-0.0230) |
| Sample #2 = | 3.6590 | (0.0910) |
| Sample #3 = | 3.6530 | (0.0970) |
| Sample #4 = | 3.6170 | (0.1330) |

Avg % Abs = 3.6430 (0.1070)
STD DEV = 0.0227 (0.0227)
REL STD DEV = 0.624 (21.230)

---

<<<<< CHANNEL 2 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 6.9520 | (-0.0060) |
| Sample #2 = | 6.7890 | (0.1720) |
| Sample #3 = | 6.7600 | (0.1860) |
| Sample #4 = | 6.7440 | (0.2120) |

Avg % Abs = 6.7643 (0.1900)
STD DEV = 0.0228 (0.0203)
REL STD DEV = 0.337 (10.683)

---

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%%
Samples Taken = 4, Discarded = 1
3um Io = 12792, 9um Io = 13810

<<<<< CHANNEL 1 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 5.3830 | (-0.0130) |
| Sample #2 = | 5.3040 | (0.0890) |
| Sample #3 = | 5.2800 | (0.0980) |
| Sample #4 = | 5.2570 | (0.1440) |

Avg % Abs = 5.2803 (0.1103)
STD DEV = 0.0235 (0.0295)
REL STD DEV = 0.445 (26.738)

---

<<<<< CHANNEL 2 >>>>>

| Sample | % Abs | (% Abs Ref) |
|---|---|---|
| Sample #1 = | 9.9780 | (-0.0070) |
| Sample #2 = | 9.7660 | (0.2120) |
| Sample #3 = | 9.7550 | (0.2160) |
| Sample #4 = | 9.7030 | (0.2750) |

Avg % Abs = 9.7413 (0.2343)
STD DEV = 0.0337 (0.0353)
REL STD DEV = 0.345 (15.053)

---

***** AUTO CAL DATA *****

<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.166
Std Dev = 0.01 Rel Std Dev = 7.94
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.02 Rel Std Dev = 1.92
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.931
Std Dev = 0.02 Rel Std Dev = 0.99
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.643
Std Dev = 0.02 Rel Std Dev = 0.62
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.280
Std Dev = 0.02 Rel Std Dev = 0.45
Zero Order Coef = -424.05
First Order Coef = 2624.29
Second Order Coef = 30.36
Standard Deviation = 13.773288

---

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.122
Std Dev = 0.02 Rel Std Dev = 13.79
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.511
Std Dev = 0.00 Rel Std Dev = 0.29
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.526
Std Dev = 0.01 Rel Std Dev = 0.33
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.764
Std Dev = 0.02 Rel Std Dev = 0.34
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.741
Std Dev = 0.03 Rel Std Dev = 0.35
Zero Order Coef = -149.04
First Order Coef = 1330.96
Second Order Coef = 15.35
Standard Deviation = 23.333534

---

Solution Stats Quadratic Fit Chan 1

| Act g/210L | Fit g/210L | Residual g/210L |
|---|---|---|
| 0.000 | 0.000 | -0.0003 |
| 0.040 | 0.040 | 0.0003 |
| 0.100 | 0.100 | 0.0001 |
| 0.200 | 0.200 | -0.0003 |
| 0.300 | 0.300 | 0.0001 |

---

Solution Stats Quadratic Fit Chan 2

| Act g/210L | Fit g/210L | Residual g/210L |
|---|---|---|
| 0.000 | 0.000 | -0.0003 |
| 0.040 | 0.040 | 0.0002 |
| 0.100 | 0.099 | 0.0006 |
| 0.200 | 0.201 | -0.0007 |
| 0.300 | 0.300 | 0.0003 |

---

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 2876.00
Sample #2 = 2832.00
Sample #3 = 2866.00
Sample #4 = 2906.00
Average Result = 2868.0000
STD DEV = 37.0405
REL STD DEV = 1.292
**********
***** CHANNEL 2
Sample #1 = 3363.00
Sample #2 = 3388.00
Sample #3 = 3396.00
Sample #4 = 3439.00
Average Result = 3407.6667
STD DEV = 27.4287
REL STD DEV = 0.805
**********
Dry Gas H20 Adjust Results **********
Barometric Pressure = 1027
3 um H20 Adjust (mg/l*10,000) = 941
9 um H20 Adjust (mg/l*10,000) = 402
**** AUTO CAL PASS

CAL ADJUSTMENT
#80-000889 SP

BK 1/27/2020

# Post Cal-Adjust Stability Checks #80-000889

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/17/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:24 |
| Control Test | 0.049 | 10:25 |
| Air Blank | 0.000 | 10:25 |
| Control Test | 0.049 | 10:26 |
| Air Blank | 0.000 | 10:27 |
| Control Test | 0.049 | 10:27 |
| Air Blank | 0.000 | 10:28 |
| Control Test Stats | | |
| Average | 0.0490 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

SP
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/17/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:14 |
| Control Test | 0.081 | 10:15 |
| Air Blank | 0.000 | 10:15 |
| Control Test | 0.080 | 10:16 |
| Air Blank | 0.000 | 10:17 |
| Control Test | 0.080 | 10:17 |
| Air Blank | 0.000 | 10:18 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

SP
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/17/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:39 |
| Control Test | 0.205 | 10:40 |
| Air Blank | 0.000 | 10:41 |
| Control Test | 0.202 | 10:41 |
| Air Blank | 0.000 | 10:42 |
| Control Test | 0.202 | 10:43 |
| Air Blank | 0.000 | 10:43 |
| Control Test Stats | | |
| Average | 0.2030 | |
| Std Dev | 0.0017 | |
| Rel Std Dev(%) | 0.8532 | |

SP
Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000889
01/17/2020
Software: 8100.27

| Test | g/210L | Time |
|---|---|---|
| Air Blank | 0.000 | 10:04 |
| Control Test | 0.080 | 10:04 |
| Air Blank | 0.000 | 10:04 |
| Control Test | 0.080 | 10:05 |
| Air Blank | 0.000 | 10:05 |
| Control Test | 0.080 | 10:06 |
| Air Blank | 0.000 | 10:06 |
| Control Test Stats | | |
| Average | 0.0800 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

DGS

SP
Operator's Signature

PDM BK 1/27/2020

# Florida Department of Law Enforcement
# Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 08:10 Date of Inspection: 01/10/2020 Serial Number: 80-000889
Software: 8100.27

| Check or Test | YES | NO |
|---|---|---|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | | No |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: Exp: | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|---|---|---|---|---|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: ____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

PGM

BK
1/27/2020

N/A compliancy not determined

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto
ISRAEL SOTO
Signature and Printed Name

01/10/2020
Date

*FDLE/ATP Form 40 – March 2004*