



INSTRUMENT PROCESSING SHEET

Agency Pasco CSO

S/N 80-001081

Florida Department of
Law Enforcement

Date In 11/12/2020 DI Completion Date 11/19/2020

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>TDG</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Replaced printer paper</u> 	Quality Checks Performed By <u>TDG</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>240</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11/12/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG931603 11/12/2021	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>TDG</u> ☒ Lab Temp °C <u>22.82</u> External Digital Therm. ID#: <u>300504</u> ☒ 34°C +/- .2 Serial #: <u>MP4863</u> ☒ 34°C +/- .2 Serial #: <u>MP4864</u> ☒ 34°C +/- .2 Serial #: <u>MP5097</u>																																													
Simulator	Serial #	Lot #/Exp																																																												
0.050	MP4863	201905A 05/14/2021																																																												
0.080	MP4864	201905B 05/14/2021																																																												
0.200	MP5097	201904D 04/30/2021																																																												
0.080 DGS	N/A	AG931603 11/12/2021																																																												
Final Release Date FDLE Alcohol Testing Program	Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.23 12:41:59 -05'00'																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ 	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1026</u> Instrument <u>1025</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <u>Israel Soto</u> 11-20-2020 Tech Review / Date <u>CS</u> 2020.11.23 Admin Review / Date		Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																											
0.000		N/A	N/A																																																											
0.040																																																														
0.100																																																														
0.200																																																														
0.300																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number	Lot Number	Expiration																																																											
0.050																																																														
0.080																																																														
0.200																																																														
0.080 DGS	N/A																																																													
Simulator	Serial Number																																																													
0.000	SD1014																																																													
Interferent	SD1015																																																													
0.050	MP4863																																																													
0.080	MP4864																																																													
0.200	MP5097																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 12:24

Date of Inspection: 11/19/2020

Serial Number: 80-001081
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.201	0.078

Standard Deviations	0.0000	0.0000	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

AS

2020.11
23
12:40:4
4-05'00'

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY


Signature and Printed Name

11/19/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001081	Pasco CSO	11/19/2020	TDG MG

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																				
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																				
✓		✓		✓		✓																																																																																																																																																				
PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>09:56</td></tr><tr><td>Control Test</td><td>0.050</td><td>09:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr><tr><td>Control Test</td><td>0.049</td><td>09:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr><tr><td>Control Test</td><td>0.000</td><td>09:58</td></tr><tr><td>Air Blank</td><td>0.050</td><td>09:59</td></tr><tr><td>Control Test</td><td>0.000</td><td>09:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0497</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr></tbody></table> Operator's Signature 		Test	g/210L	Time	Air Blank	0.000	09:56	Control Test	0.050	09:56	Air Blank	0.000	09:57	Control Test	0.049	09:57	Air Blank	0.000	09:57	Control Test	0.000	09:58	Air Blank	0.050	09:59	Control Test	0.000	09:59	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table> Operator's Signature 		Test	g/210L	Time	Air Blank	0.000	10:03	Control Test	0.080	10:04	Air Blank	0.000	10:05	Control Test	0.080	10:05	Air Blank	0.000	10:06	Control Test	0.079	10:07	Air Blank	0.000	10:07	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:15</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2010</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature 		Test	g/210L	Time	Air Blank	0.000	10:12	Control Test	0.201	10:13	Air Blank	0.000	10:14	Control Test	0.201	10:14	Air Blank	0.000	10:15	Control Test	0.201	10:16	Air Blank	0.000	10:16	Control Test Stats			Average	0.2010		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/19/2020 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr><tr><td>Control Test</td><td>0.082</td><td>10:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr><tr><td>Control Test</td><td>0.082</td><td>10:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0817</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr></tbody></table> Operator's Signature 		Test	g/210L	Time	Air Blank	0.000	10:18	Control Test	0.081	10:18	Air Blank	0.000	10:19	Control Test	0.082	10:19	Air Blank	0.000	10:20	Control Test	0.082	10:20	Air Blank	0.000	10:20	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Test	g/210L	Time																																																																																																																																																								
Air Blank	0.000	09:56																																																																																																																																																								
Control Test	0.050	09:56																																																																																																																																																								
Air Blank	0.000	09:57																																																																																																																																																								
Control Test	0.049	09:57																																																																																																																																																								
Air Blank	0.000	09:57																																																																																																																																																								
Control Test	0.000	09:58																																																																																																																																																								
Air Blank	0.050	09:59																																																																																																																																																								
Control Test	0.000	09:59																																																																																																																																																								
Control Test Stats																																																																																																																																																										
Average	0.0497																																																																																																																																																									
Std Dev	0.0006																																																																																																																																																									
Rel Std Dev(%)	1.1625																																																																																																																																																									
Test	g/210L	Time																																																																																																																																																								
Air Blank	0.000	10:03																																																																																																																																																								
Control Test	0.080	10:04																																																																																																																																																								
Air Blank	0.000	10:05																																																																																																																																																								
Control Test	0.080	10:05																																																																																																																																																								
Air Blank	0.000	10:06																																																																																																																																																								
Control Test	0.079	10:07																																																																																																																																																								
Air Blank	0.000	10:07																																																																																																																																																								
Control Test Stats																																																																																																																																																										
Average	0.0797																																																																																																																																																									
Std Dev	0.0006																																																																																																																																																									
Rel Std Dev(%)	0.7247																																																																																																																																																									
Test	g/210L	Time																																																																																																																																																								
Air Blank	0.000	10:12																																																																																																																																																								
Control Test	0.201	10:13																																																																																																																																																								
Air Blank	0.000	10:14																																																																																																																																																								
Control Test	0.201	10:14																																																																																																																																																								
Air Blank	0.000	10:15																																																																																																																																																								
Control Test	0.201	10:16																																																																																																																																																								
Air Blank	0.000	10:16																																																																																																																																																								
Control Test Stats																																																																																																																																																										
Average	0.2010																																																																																																																																																									
Std Dev	0.0000																																																																																																																																																									
Rel Std Dev(%)	0.0000																																																																																																																																																									
Test	g/210L	Time																																																																																																																																																								
Air Blank	0.000	10:18																																																																																																																																																								
Control Test	0.081	10:18																																																																																																																																																								
Air Blank	0.000	10:19																																																																																																																																																								
Control Test	0.082	10:19																																																																																																																																																								
Air Blank	0.000	10:20																																																																																																																																																								
Control Test	0.082	10:20																																																																																																																																																								
Air Blank	0.000	10:20																																																																																																																																																								
Control Test Stats																																																																																																																																																										
Average	0.0817																																																																																																																																																									
Std Dev	0.0006																																																																																																																																																									
Rel Std Dev(%)	0.7070																																																																																																																																																									

Comments:



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001081, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001081</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/19/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>12:24</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

11/19/2020

Date


MICHAEL D HAUGHEY,
Department Inspector

Service • Integrity • Respect • Quality

Page 1 of 1