



# INSTRUMENT PROCESSING SHEET

Agency FHP OrlandoS/N 80-001123

Florida Department of Law Enforcement

Date In 04/02/2020 DI Completion Date 4/9/20 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DA</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>211</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26Q32</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
--	--	--

**Final Release Date**

FDLE Alcohol Testing Program

Digitally signed by FDLE Alcohol Testing Program  
 Date: 2020.04.16 16:20:19 -04'00'

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A 05-14-2021
0.080	MP5089	201905B 05-14-2021
0.200	MP5090	201904D 04-30-2021
0.080 DGS	N/A	AG931603 11-12-2021

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Temperature Checks** Performed By SP

Lab Temp °C 21.14  
 External Digital Therm. ID#: 300502  
 34°C +- .2 Serial #: MP5088  
 34°C +- .2 Serial #: MP5089  
 34°C +- .2 Serial #: MP5090

**Calibration Adjustment** Performed By SP

Barometric Pressure Gauge 1013 ID # 28421

Simulator	Serial Number	Lot Number	Expiration
0.000	MP5086	N/A	N/A
0.040	MP5082	19080	3-4-21
0.100	MP5083	19160	7-9-21
0.200	MP5084	19040	1-29-21
0.300	MP5085	19010	1-3-21
0.080 DGS	N/A	03519080A4	4-5-21

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	MP5088	201905A	5-14-21
0.080	MP5089	201905B	5-14-21
0.200	MP5090	201904D	4-30-21
0.080 DGS	N/A	AG931603	11-12-21

**Department Inspection** Performed By SP

Barometric Pressure ID# 30793  
 Gauge 1008 Instrument 1009  
 Mouth Alcohol Solution Lot # 2019-B  
 Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	MP5091
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

**Attachments**

Form 41  Post-Stability Checks  
 Stability Checks  Flow Calibration  
 Calibration Certificate  Form 40  
 Calibration Adjustment  Other \_\_\_\_\_


Notes/Suggested Service: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

2020.04.16 12:34:33 -04'00' Michael D. Hargray 2020.04.16 16:16:09 Brett Kirkland  
 Tech Review / Date Admin Review / Date



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2729 Fort Knox Blvd.  
 Bldg. 2, Suite 1300  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001123, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001123</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FHP TROOP D ORLANDO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/09/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>10:55</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

*Shayla Platt*

04/09/2020

Date

SHAYLA D PLATT,  
 Department Inspector

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

*MH*

*BK*

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP D ORLANDO  
Time of Inspection: 10:55

Date of Inspection: 04/09/2020

Serial Number: 80-001123  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.198	0.079

Standard Deviations	0.0000	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

MH  
BK

2020.04.16  
16:17:18 -04'00'

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

04/09/2020  
Date

# Stability Checks

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001123  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.049	08:28
Air Blank	0.000	08:28
Control Test	0.049	08:29
Air Blank	0.000	08:29
Control Test	0.048	08:30
Air Blank	0.000	08:30
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001123  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:33
Control Test	0.080	08:34
Air Blank	0.000	08:34
Control Test	0.080	08:35
Air Blank	0.000	08:35
Control Test	0.079	08:36
Air Blank	0.000	08:37
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001123  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.200	08:39
Air Blank	0.000	08:40
Control Test	0.200	08:40
Air Blank	0.000	08:41
Control Test	0.201	08:42
Air Blank	0.000	08:42
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

wet

  
 -----  
 Operator's Signature

  
 -----  
 Operator's Signature

  
 -----  
 Operator's Signature

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001123  
 04/07/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.080	08:44
Air Blank	0.000	08:45
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test	0.080	08:46
Air Blank	0.000	08:46
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MH

BK 2020.04.16  
 16:18:09 -04'00'

Dry

  
 -----  
 Operator's Signature

FHP TROOP D ORLANDO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
04/08/2020

SN 80-001123  
08:17:56

Auto Calibration  
Max Power Res Value = 29  
Auto Range Res Value = 20

Soi Value = 0.000 g/210L \*\*\*  
Fit Value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12783, Sum Io = 13839  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.0180 (-0.0060)  
Sample #2 = 0.0090 (0.0060)  
Sample #3 = 0.0050 (0.0020)  
Sample #4 = 0.0000 (0.0040)  
Avg % Abs = 0.0047 (0.0040)  
STD DEV = 0.0045 (0.0020)  
REL STD DEV = 96.627 (50.000)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = -0.0020 (-0.0020)  
Sample #2 = 0.0160 (-0.0050)  
Sample #3 = 0.0000 (-0.0070)  
Sample #4 = 0.0070 (0.0000)  
Avg % Abs = 0.0077 (-0.0040)  
STD DEV = 0.0080 (0.0036)  
REL STD DEV = 104.619 (50.139)

Soi Value = 0.040 g/210L \*\*\*  
Fit Value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12781, Sum Io = 13836  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8280 (-0.0280)  
Sample #2 = 0.8080 (-0.0100)  
Sample #3 = 0.8150 (0.0000)  
Sample #4 = 0.8270 (-0.0270)  
Avg % Abs = 0.8167 (-0.0123)  
STD DEV = 0.0096 (0.0137)  
REL STD DEV = 1.177 (110.679)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5420 (-0.0030)  
Sample #2 = 1.5590 (-0.0130)  
Sample #3 = 1.5470 (-0.0200)  
Sample #4 = 1.5450 (-0.0040)  
Avg % Abs = 1.5503 (-0.0123)  
STD DEV = 0.0076 (0.0080)  
REL STD DEV = 0.488 (65.034)

Soi Value = 0.100 g/210L \*\*\*  
Fit Value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12784, Sum Io = 13835  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.8400 (-0.0100)  
Sample #2 = 1.8540 (-0.0080)  
Sample #3 = 1.8510 (-0.0210)  
Sample #4 = 1.8300 (-0.0060)  
Avg % Abs = 1.8450 (-0.0117)  
STD DEV = 0.0131 (0.0081)  
REL STD DEV = 0.709 (69.810)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6200 (-0.0020)  
Sample #2 = 3.6280 (-0.0200)  
Sample #3 = 3.6320 (-0.0140)  
Sample #4 = 3.6260 (-0.0150)  
Avg % Abs = 3.6287 (-0.0163)  
STD DEV = 0.0031 (0.0032)  
REL STD DEV = 0.084 (19.681)

Soi Value = 0.200 g/210L \*\*\*  
Fit Value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12787, Sum Io = 13836  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5770 (-0.0190)  
Sample #2 = 3.5600 (0.0080)  
Sample #3 = 3.5490 (0.0210)  
Sample #4 = 3.5650 (0.0080)  
Avg % Abs = 3.5680 (0.0123)  
STD DEV = 0.0092 (0.0075)  
REL STD DEV = 0.230 (60.856)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 7.0070 (-0.0020)  
Sample #2 = 6.9950 (0.0080)  
Sample #3 = 6.9840 (0.0190)  
Sample #4 = 6.9910 (0.0150)  
Avg % Abs = 6.9897 (0.0140)  
STD DEV = 0.0055 (0.0056)  
REL STD DEV = 0.079 (39.770)

Soi Value = 0.300 g/210L \*\*\*  
Fit Value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12787, Sum Io = 13833  
<<<<< CHANNEL 1 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.1450 (0.0000)  
Sample #2 = 5.1710 (0.0000)  
Sample #3 = 5.1550 (0.0060)  
Sample #4 = 5.1690 (-0.0160)  
Avg % Abs = 5.1650 (-0.0033)  
STD DEV = 0.0087 (0.0114)  
REL STD DEV = 0.169 (341.174)

<<<<< CHANNEL 2 >>>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.0830 (-0.0060)  
Sample #2 = 10.1170 (-0.0110)  
Sample #3 = 10.0880 (0.0050)  
Sample #4 = 10.0870 (-0.0060)  
Avg % Abs = 10.0973 (-0.0040)  
STD DEV = 0.0170 (0.0082)  
REL STD DEV = 0.169 (204.634)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<<< CHANNEL 1 >>>>>  
Soi Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.005  
Std Dev = 0.00 Rel Std Dev = 96.63  
Soi Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.817  
Std Dev = 0.01 Rel Std Dev = 1.18  
Soi Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.845  
Std Dev = 0.01 Rel Std Dev = 0.71  
Soi Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.558  
Std Dev = 0.01 Rel Std Dev = 0.23  
Soi Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.165  
Std Dev = 0.01 Rel Std Dev = 0.17  
Zero Order Coef = -64.41  
First Order Coef = 2489.08  
Second Order Coef = 56.42  
Standard Deviation = 61.698788

<<<<< CHANNEL 2 >>>>>  
Soi Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.008  
Std Dev = 0.01 Rel Std Dev = 104.62  
Soi Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.550  
Std Dev = 0.01 Rel Std Dev = 0.49  
Soi Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.629  
Std Dev = 0.00 Rel Std Dev = 0.08  
Soi Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.990  
Std Dev = 0.01 Rel Std Dev = 0.08  
Soi Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 10.097  
Std Dev = 0.02 Rel Std Dev = 0.17  
Zero Order Coef = -31.45  
First Order Coef = 1253.90  
Second Order Coef = 16.25  
Standard Deviation = 29.754848

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3143.00  
Sample #2 = 3163.00  
Sample #3 = 3051.00  
Sample #4 = 3138.00  
Average Result = 3117.3333  
STD DEV = 58.7906  
REL STD DEV = 1.886  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3619.00  
Sample #2 = 3449.00  
Sample #3 = 3388.00  
Sample #4 = 3398.00  
Average Result = 3411.6667  
STD DEV = 32.7160  
REL STD DEV = 0.959  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1013  
3 um H2O Adjust (mg/l \* 10,000) = 692  
9 um H2O Adjust (mg/l \* 10,000) = 398  
\*\*\*\* AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.001 0.0011  
0.040 0.042 -0.0021  
0.100 0.099 0.0009  
0.200 0.200 0.0004  
0.300 0.300 -0.0002

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0005  
0.040 0.041 -0.0010  
0.100 0.099 0.0006  
0.200 0.200 -0.0001  
0.300 0.300 -0.0000

Soi Value = 0.080 g/210L \*\*\*  
Fit Value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Soi Value = 0.080 g/210L \*\*\*  
Fit Value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Soi Value = 0.080 g/210L \*\*\*  
Fit Value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Soi Value = 0.080 g/210L \*\*\*  
Fit Value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

CAL ADJUSTMENT  
# 80-001123 SP

MH  
BK

# Post Cal Adjust Stability Checks #80-001123

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer SN 80-001123  
 Model 8000  
 04/09/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:07
Control Test	0.049	08:08
Air Blank	0.000	08:08
Control Test	0.049	08:09
Air Blank	0.000	08:09
Control Test	0.050	08:10
Air Blank	0.000	08:11
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

*SP*  
 -----  
 Operator's Signature

*MA*

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer SN 80-001123  
 Model 8000  
 04/09/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.080	08:21
Air Blank	0.000	08:22
Control Test	0.079	08:22
Air Blank	0.000	08:23
Control Test	0.080	08:24
Air Blank	0.000	08:24
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

*SP*  
 -----  
 Operator's Signature

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer SN 80-001123  
 Model 8000  
 04/09/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:15
Control Test	0.199	08:15
Air Blank	0.000	08:16
Control Test	0.199	08:17
Air Blank	0.000	08:17
Control Test	0.199	08:18
Air Blank	0.000	08:18
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*SP*  
 -----  
 Operator's Signature

FHP TROOP D ORLANDO  
 Intoxilyzer - Alcohol Analyzer SN 80-001123  
 Model 8000  
 04/09/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:25
Control Test	0.081	08:26
Air Blank	0.000	08:26
Control Test	0.080	08:26
Air Blank	0.000	08:27
Control Test	0.080	08:27
Air Blank	0.000	08:28
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

*SP*

*SP*  
 -----  
 Operator's Signature