



# INSTRUMENT PROCESSING SHEET

Agency Volusia CountyS/N 80-001131Florida Department of  
Law EnforcementDate In 2/27/2020DI Completion Date 4/1/2020 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DP</u>		<b>Quality Checks</b> Performed By <u>DP</u>		<b>Flow Calibration</b> Performed By _____																																																													
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>113</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
<b>Final Release Date</b>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3962</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>SD3962</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____  <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.1</u> External Digital Therm. ID#: <u>300503</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5090</u>																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>																																																															
0.080	<u>SD3962</u>	<u>201905B</u> <u>05-14-2021</u>																																																															
0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>																																																															
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																																															
<b>Calibration Adjustment</b> Performed By <u>SP</u>		<b>Department Inspection</b> Performed By <u>SP</u>																																																															
Barometric Pressure Gauge <u>1019</u> ID # <u>30793</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>G8144</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>G2403</u></td> <td><u>19080</u></td> <td><u>3-4-21</u></td> </tr> <tr> <td>0.100</td> <td><u>G2879</u></td> <td><u>19160</u></td> <td><u>7-9-21</u></td> </tr> <tr> <td>0.200</td> <td><u>G3709</u></td> <td><u>19040</u></td> <td><u>1-29-21</u></td> </tr> <tr> <td>0.300</td> <td><u>G8149</u></td> <td><u>19010</u></td> <td><u>1-3-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>08819050A7</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>201905A</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>201905B</u></td> <td><u>5-14-21</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>201904D</u></td> <td><u>4-30-21</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u></td> <td><u>11-12-21</u></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	<u>G8144</u>	N/A	N/A	0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>	0.100	<u>G2879</u>	<u>19160</u>	<u>7-9-21</u>	0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>	0.300	<u>G8149</u>	<u>19010</u>	<u>1-3-21</u>	0.080 DGS	N/A	<u>08819050A7</u>	<u>6-5-21</u>	Simulator	Serial Number	Lot Number	Expiration	0.050	<u>SD1021</u>	<u>201905A</u>	<u>5-14-21</u>	0.080	<u>DR1275</u>	<u>201905B</u>	<u>5-14-21</u>	0.200	<u>SD1011</u>	<u>201904D</u>	<u>4-30-21</u>	0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>	Barometric Pressure ID# <u>30793</u> Gauge <u>1011</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>		
Simulator	Serial Number	Lot Number	Expiration																																																														
0.000	<u>G8144</u>	N/A	N/A																																																														
0.040	<u>G2403</u>	<u>19080</u>	<u>3-4-21</u>																																																														
0.100	<u>G2879</u>	<u>19160</u>	<u>7-9-21</u>																																																														
0.200	<u>G3709</u>	<u>19040</u>	<u>1-29-21</u>																																																														
0.300	<u>G8149</u>	<u>19010</u>	<u>1-3-21</u>																																																														
0.080 DGS	N/A	<u>08819050A7</u>	<u>6-5-21</u>																																																														
Simulator	Serial Number	Lot Number	Expiration																																																														
0.050	<u>SD1021</u>	<u>201905A</u>	<u>5-14-21</u>																																																														
0.080	<u>DR1275</u>	<u>201905B</u>	<u>5-14-21</u>																																																														
0.200	<u>SD1011</u>	<u>201904D</u>	<u>4-30-21</u>																																																														
0.080 DGS	N/A	<u>AG931603</u>	<u>11-12-21</u>																																																														
Simulator	Serial Number																																																																
0.000	<u>MP5086</u>																																																																
Interferent	<u>MP5087</u>																																																																
0.050	<u>MP5088</u>																																																																
0.080	<u>MP5089</u>																																																																
0.200	<u>MP5090</u>																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																															
		Tech Review / Date _____		Admin Review / Date _____																																																													



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001131, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001131</u>	UNCERTAINTY * ±	
Owning Agency:	<u>VOLUSIA COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/01/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>12:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/01/2020 Date  
  
SHAYLA PLATT,  
Department Inspector

FDLE/ATP Form 69 January 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.  
Time of Inspection: 12:12

Date of Inspection: 04/01/2020

Serial Number: 80-001131  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081

Standard Deviations	0.0000	0.0005	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/01/2020  
Date

# Stability Checks

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 02/28/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:31
Control Test	0.046	13:32
Air Blank	0.000	13:32
Control Test	0.047	13:33
Air Blank	0.000	13:33
Control Test	0.048	13:34
Air Blank	0.000	13:35
Control Test Stats		
Average	0.0470	
Std Dev	0.0010	
Rel Std Dev(%)	2.1277	

*SS*

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 02/28/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:37
Control Test	0.077	13:37
Air Blank	0.000	13:38
Control Test	0.078	13:39
Air Blank	0.000	13:39
Control Test	0.078	13:40
Air Blank	0.000	13:40
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

*SS*

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 02/28/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:41
Control Test	0.197	13:42
Air Blank	0.000	13:43
Control Test	0.198	13:43
Air Blank	0.000	13:44
Control Test	0.199	13:45
Air Blank	0.000	13:45
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

*SS*

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 02/28/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:46
Control Test	0.078	13:46
Air Blank	0.000	13:47
Control Test	0.079	13:47
Air Blank	0.000	13:48
Control Test	0.080	13:48
Air Blank	0.000	13:48
Control Test Stats		
Average	0.0790	
Std Dev	0.0019	
Rel Std Dev(%)	1.2658	

DRY

*SS*

Operator's Signature

LOUISIANA COUNTY S.O.

Intoxilyzer - Alcohol Analyzer

Model 8000

03/13/2020

09:24:

Auto Calibration

Max Power Res Value = 21

Auto Range Res Value = 10

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.5570	(0.0120)
Sample #2	1.5510	(0.0170)
Sample #3	1.5710	(0.0070)
Sample #4	1.5520	(0.0240)

Aug % Abs = 1.5613 (0.0160)  
 STD DEV = 0.0095 (0.0085)  
 REL STD DEV = 0.609 (53.400)

Sol Value = 0.100 g/210L \*\*\*  
 Fit value = 0.4762 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12880, Sum Io = 13699

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	1.9190	(-0.0050)
Sample #2	1.9120	(-0.0240)
Sample #3	1.8880	(0.0140)
Sample #4	1.9240	(-0.0070)

Aug % Abs = 1.9047 (-0.0057)  
 STD DEV = 0.0181 (0.0190)  
 REL STD DEV = 0.953 (335.913)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.7980	(-0.0140)
Sample #2	3.7500	(-0.0200)
Sample #3	3.7710	(-0.0090)
Sample #4	3.7650	(-0.0060)

Aug % Abs = 3.7620 (-0.0117)  
 STD DEV = 0.0108 (0.0074)  
 REL STD DEV = 0.288 (63.181)

Sol Value = 0.200 g/210L \*\*\*  
 Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12878, Sum Io = 13691

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.6370	(-0.0190)
Sample #2	3.6270	(0.0140)
Sample #3	3.6260	(0.0000)
Sample #4	3.6340	(0.0070)

Aug % Abs = 3.6290 (0.0070)  
 STD DEV = 0.0044 (0.0070)  
 REL STD DEV = 0.120 (100.000)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	7.1370	(-0.0160)
Sample #2	7.1040	(0.0240)
Sample #3	7.1430	(0.0000)
Sample #4	7.1460	(0.0040)

Aug % Abs = 7.1310 (0.0093)  
 STD DEV = 0.0234 (0.0129)  
 REL STD DEV = 0.329 (137.766)

Sol Value = 0.300 g/210L \*\*\*  
 Fit value = 1.4286 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12876, Sum Io = 13689

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	5.2930	(-0.0180)
Sample #2	5.2810	(0.0200)
Sample #3	5.2900	(0.0060)
Sample #4	5.3120	(-0.0030)

Aug % Abs = 5.2943 (0.0077)  
 STD DEV = 0.0159 (0.0116)  
 REL STD DEV = 0.301 (151.177)

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	10.2620	(-0.0240)
Sample #2	10.2820	(0.0030)
Sample #3	10.3160	(0.0000)
Sample #4	10.3170	(0.0050)

Aug % Abs = 10.3050 (0.0027)  
 STD DEV = 0.0199 (0.0025)  
 REL STD DEV = 0.193 (94.373)

Sol Value = 0.200 g/210L \*\*\*  
 Fit value = 0.9524 mg/l %%%  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 12878, Sum Io = 13691

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1	3.6370	(-0.0190)
Sample #2	3.6270	(0.0140)
Sample #3	3.6260	(0.0000)
Sample #4	3.6340	(0.0070)

Aug % Abs = 3.6290 (0.0070)  
 STD DEV = 0.0044 (0.0070)  
 REL STD DEV = 0.120 (100.000)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.068  
 Std Dev = 0.01 Rel Std Dev = 10.59  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.805  
 Std Dev = 0.01 Rel Std Dev = 1.23  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.905  
 Std Dev = 0.02 Rel Std Dev = 0.95  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.629  
 Std Dev = 0.00 Rel Std Dev = 0.12  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 5.294  
 Std Dev = 0.02 Rel Std Dev = 0.30  
 Zero Order Coef = -169.27  
 First Order Coef = 2527.28  
 Second Order Coef = 38.49  
 Standard Deviation = 15.595988

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.067  
 Std Dev = 0.01 Rel Std Dev = 13.68  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.561  
 Std Dev = 0.01 Rel Std Dev = 0.61  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.762  
 Std Dev = 0.01 Rel Std Dev = 0.29  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 7.131  
 Std Dev = 0.02 Rel Std Dev = 0.33  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 10.305  
 Std Dev = 0.02 Rel Std Dev = 0.19  
 Zero Order Coef = -77.03  
 First Order Coef = 1233.28  
 Second Order Coef = 15.60  
 Standard Deviation = 15.816482

Sol Value = 0.080 g/210L \*\*\*  
 Fit value = 0.3810 mg/l %%%  
 Samples Taken = 4, Discarded = 1

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Sample #1	3234.00
Sample #2	3167.00
Sample #3	3288.00
Sample #4	3154.00

Average Result = 3203.0000  
 STD DEV = 73.8986  
 REL STD DEV = 2.307

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*

Sample #1	3363.00
Sample #2	3375.00
Sample #3	3391.00
Sample #4	3353.00

Average Result = 3373.0000  
 STD DEV = 19.0788  
 REL STD DEV = 0.566

CAL ADJUSTMENT  
 #80-00131 SP

\*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1019  
 3 um H2O Adjust (mg/l\*10,000) = 606  
 9 um H2O Adjust (mg/l\*10,000) = 436  
 \*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

# Post Cal Adjust Stability Checks #80-001131

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 03/16/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:56
Control Test	0.049	09:57
Air Blank	0.000	09:57
Control Test	0.049	09:58
Air Blank	0.000	09:58
Control Test	0.048	09:59
Air Blank	0.000	09:59
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

SP

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 03/16/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.077	09:47
Air Blank	0.000	09:48
Control Test	0.078	09:48
Air Blank	0.000	09:49
Control Test	0.078	09:49
Air Blank	0.000	09:50
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

SP

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 03/16/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.197	09:52
Air Blank	0.000	09:52
Control Test	0.197	09:53
Air Blank	0.000	09:53
Control Test	0.198	09:54
Air Blank	0.000	09:55
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

SP

Operator's Signature

VOLUSIA COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-001131  
 03/16/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.081	10:05
Air Blank	0.000	10:05
Control Test	0.080	10:06
Air Blank	0.000	10:06
Control Test	0.081	10:07
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

SP

Operator's Signature



# Post Cal Adjust Stability Checks #80-001131

WOLUSTIA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001131  
03/29/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	14:24
Control Test	0.049	14:24
Air Blank	0.000	14:25
Control Test	0.048	14:25
Air Blank	0.000	14:26
Control Test	0.048	14:27
Air Blank	0.000	14:27
Control Test Stats		
Average	0.0483	
Std Dev	0.0016	
Rel Std Dev(%)	1.1945	

MP5088

SP

Operator's Signature

WOLUSTIA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001131  
03/29/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	14:28
Control Test	0.079	14:29
Air Blank	0.000	14:29
Control Test	0.079	14:30
Air Blank	0.000	14:30
Control Test	0.078	14:31
Air Blank	0.000	14:32
Control Test Stats		
Average	0.0787	
Std Dev	0.0016	
Rel Std Dev(%)	0.7339	

MP5089

SP

Operator's Signature

WOLUSTIA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001131  
03/29/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	14:34
Control Test	0.196	14:35
Air Blank	0.000	14:36
Control Test	0.198	14:36
Air Blank	0.000	14:37
Control Test	0.197	14:37
Air Blank	0.000	14:38
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	

MP5090

SP

Operator's Signature

WOLUSTIA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001131  
03/29/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	14:40
Control Test	0.080	14:40
Air Blank	0.000	14:40
Control Test	0.081	14:41
Air Blank	0.000	14:41
Control Test	0.081	14:42
Air Blank	0.000	14:42
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

SP

Operator's Signature

Repeated post stability checks  
after suspecting issue w/  
previous simulators.