




INSTRUMENT PROCESSING SHEET

Agency Clermont Police DepartmentS/N 80-001174Florida Department of
Law EnforcementDate In 11/20/2020 DI Completion Date 11/25/2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>TDG</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Pelican 1620 Case with combination lock</u>	Quality Checks Performed By <u>TDG</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28663-68639</u> See "Notes" for Stability ☒ Stability Checks <u>11/30/2020</u> Lot # <u>exp</u>	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.12.09 14:04:22 -05'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>TL-201905A TL-05/14/2024</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>TL-201905B TL-05/14/2024</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>TL-201904D TL-04/30/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 1/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	TL-201905A TL-05/14/2024	0.080	MP4864	TL-201905B TL-05/14/2024	0.200	MP5097	TL-201904D TL-04/30/2024	0.080 DGS	N/A	AG003005 1/30/2022	Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>DERR</u> ☒ Lab Temp °C <u>22.37C</u> External Digital Therm. ID#: <u>300918</u> ☒ 34°C +/-2 Serial #: <u>SD3967</u> ☒ 34°C +/-2 Serial #: <u>SD3968</u> ☒ 34°C +/-2 Serial #: <u>SD3969</u>																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP4863	TL-201905A TL-05/14/2024																																																													
0.080	MP4864	TL-201905B TL-05/14/2024																																																													
0.200	MP5097	TL-201904D TL-04/30/2024																																																													
0.080 DGS	N/A	AG003005 1/30/2022																																																													
Calibration Adjustment Performed By <u>DERR</u> Barometric Pressure Gauge <u>1019</u> ID # <u>68639</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>20190</td> <td>04/06/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005</td> <td>01/30/2022</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	20060	02/10/2022	0.100	MP5099	20190	04/06/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	SD3967	201905A	05/14/2021	0.080	SD3968	201905B	05/14/2021	0.200	SD3969	201904D	04/30/2021	0.080 DGS	N/A	AG003005	01/30/2022	Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2020A</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	MP5095	N/A	N/A																																																												
0.040	MP5098	20060	02/10/2022																																																												
0.100	MP5099	20190	04/06/2022																																																												
0.200	MP5100	20160	03/18/2022																																																												
0.300	MP5101	20030	01/21/2022																																																												
0.080 DGS	N/A	08819080A1	06/05/2021																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050	SD3967	201905A	05/14/2021																																																												
0.080	SD3968	201905B	05/14/2021																																																												
0.200	SD3969	201904D	04/30/2021																																																												
0.080 DGS	N/A	AG003005	01/30/2022																																																												
Simulator	Serial Number																																																														
0.000	SD3965																																																														
Interferent	SD3966																																																														
0.050	SD3967																																																														
0.080	SD3968																																																														
0.200	SD3969																																																														
Notes/Suggested Service: <u>Optical calibration to bring values closer to nominal. DERR</u> <u>0.050 9/20 L - #202010A, 10/05/2022</u> <u>0.080 9/20 L - #202010B, 10/05/2022</u> <u>0.200 9/20 L - #202010D, 10/06/2022</u> <u>TDG 11/30/2020</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2020.11.30 <u>Michael D. Haggerty</u> 13:07:50 -05'00' Tech Review / Date 2020.12.09 <u>CR</u> 14:02:57 Admin Review / Date																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001174, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001174</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CLERMONT P.D.</u>	0.050 g/210 L	0.004
Calibration Date:	<u>11/25/2020</u>	0.080 g/210 L	0.005
Calibration Time:	<u>06:45</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/25/2020

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

2020.12.09
13:47:28
-05'00'

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 06:45

Date of Inspection: 11/25/2020

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.049	0.079	0.203	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.202	0.080

Standard Deviations	0.0005	0.0003	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

11/25/2020
Date

MX
2020.12.
09
13:47:52
-05'00'

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001174	Clermont Police Department	11/25/2020	DERR <i>DELL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ☒ CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/25/2020 Software: 8100.27 SN 80-001174 Test g/210L Time Air Blank 0.000 03:26 Control Test 0.047 03:27 Air Blank 0.000 03:28 Control Test 0.047 03:28 Air Blank 0.000 03:29 Control Test 0.048 03:29 Air Blank 0.000 03:30 Control Test Stats Average 0.0473 Std Dev 0.0006 Rel Std Dev(%) 1.2198 Operator's Signature <i>DELL</i>	0.077 to 0.083 ☒ CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/25/2020 Software: 8100.27 SN 80-001174 Test g/210L Time Air Blank 0.000 03:31 Control Test 0.078 03:32 Air Blank 0.000 03:32 Control Test 0.078 03:33 Air Blank 0.000 03:34 Control Test 0.078 03:34 Air Blank 0.000 03:35 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>DELL</i>	0.194 to 0.206 ☒ CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/25/2020 Software: 8100.27 SN 80-001174 Test g/210L Time Air Blank 0.000 03:36 Control Test 0.198 03:37 Air Blank 0.000 03:37 Control Test 0.199 03:38 Air Blank 0.000 03:39 Control Test 0.199 03:39 Air Blank 0.000 03:40 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2905 Operator's Signature <i>DELL</i>	0.077 to 0.083 ☒ CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/25/2020 Software: 8100.27 SN 80-001174 Test g/210L Time Air Blank 0.000 03:42 Control Test 0.079 03:42 Air Blank 0.000 03:42 Control Test 0.079 03:43 Air Blank 0.000 03:43 Control Test 0.079 03:44 Air Blank 0.000 03:44 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>DELL</i>

2020.12
09
14:02:0
5-05'00'

MX

CLERMONT P.D.
Intoxilizer - Alcohol Analyzer
Model 8000
SN 8J-001174
11/25/2020 02:43:23

```

Auto Calibration
Max Power Res Value = 26
Auto Range Res Value = 20

Sol Value = 0.00 g/210L ***
Tit Value = 0.000 mg/l ****
Samples Taken = 4, Discarded = 1
Run 10 = 12801, Sum 10 = 13845

<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 0.1340    (0.0000)
Sample #2 = 0.1230    (0.0250)
Sample #3 = 0.1420    (0.0370)
Sample #4 = 0.1330    (0.0820)

ug % Abs = 0.1327 (0.0480)
TD ORU = 0.0095 (0.0300)
EL STD DEV = 7.164 (62.604)

```

```

TO DEV = 0.0095 (0.0300)
EL STD DEV = 7.164 (62.604)
-----
<<<< CHANNEL 2 >>>>
Sample      % ABS      (% ABS Ref)
sample #1 = 0.1240 (0.0000)
sample #2 = 0.1280 (-0.0060)
sample #3 = 0.1280 (0.0020)
sample #4 = 0.1400 (0.0080)
%g % ABS = 0.1320 (0.0113)
TO DEV = 0.0069 (0.0070)
EL STD DEV = 5.249 (506.783)

```

Sample	% Abs	(% Abs)
Sample #1 =	3.7240	(-0.01)
Sample #2 =	3.7390	(0.01)
Sample #3 =	3.7320	(0.01)
Sample #4 =	3.7270	(0.01)
Avg % Abs =	3.7327	(0.0193)
STD DEV =	0.0060	(0.0176)
REL STD DEV =	0.161	(91.119)

```

<<<< CHANNEL 2 >>>>
Sample      % ABS      (% ABS Ref)
Sample #1 = 10.2880  (-0.0100)
Sample #2 = 10.2300  (-0.0530)
Sample #3 = 10.2430  (-0.0860)
Sample #4 = 10.2540  (-0.0820)
Avg % ABS = 10.2423  (-0.0770)
STD DEV = 0.0120  (-0.0123)
REL STD DEV = 0.117  (-15.959)

```

```

TD DEV = 0.0069 (0.0070)
EL STD DEV = 5.249 (526.783)

-----

)l Value = 0.040 g/210L ***
it value = 0.1905 ng/l 888%
amples Taken = 4, Discarded = 1
m lo = 12789, g m lo = 13839

<<<< CHANNEL 1 >>>>

Sample      % Aps      (% Aps Ref)
sample #1 = 0.9050      (-0.0150)
sample #2 = 0.8930      (0.0290)
sample #3 = 0.9190      (0.0380)
sample #4 = 0.9040      (0.0500)
g % Aps = 0.9031 (0.0507)
D DEV = 0.0131 (0.0191)
L STD DEV = 1.442 (37.673)

```

Optical Calibration
SN: 80-001174
Agency: Clermont PD
Date: 11/25/2020
Quadratic Fit: +/- 0.002g/2
By: DERR <i>DELL</i>

2020.12.09
13:48:44 -05'00'

MX

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001174	Clermont PD	11/20/2020	TDG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/20/2020 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:46 Control Test 0.048 14:46 Air Blank 0.000 14:47 Control Test 0.047 14:48 Air Blank 0.000 14:48 Control Test 0.048 14:49 Air Blank 0.000 14:49 Control Test Stats Average 0.0477 Std Dev 0.0006 Rel Std Dev(%) 1.2112 Operator's Signature	CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/20/2020 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:51 Control Test 0.078 14:52 Air Blank 0.000 14:52 Control Test 0.078 14:53 Air Blank 0.000 14:54 Control Test 0.078 14:54 Air Blank 0.000 14:55 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature	CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/20/2020 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:57 Control Test 0.198 14:57 Air Blank 0.000 14:58 Control Test 0.198 14:58 Air Blank 0.000 14:59 Control Test 0.197 15:00 Air Blank 0.000 15:00 Control Test Stats Average 0.1977 Std Dev 0.0006 Rel Std Dev(%) 0.2321 Operator's Signature	CLERMONT P.D. Intoxilyzer - Alcohol Analyzer Model 8000 11/20/2020 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:01 Control Test 0.082 15:01 Air Blank 0.000 15:02 Control Test 0.082 15:02 Air Blank 0.000 15:03 Control Test 0.081 15:03 Air Blank 0.000 15:04 Control Test Stats Average 0.0817 Std Dev 0.0006 Rel Std Dev(%) 0.7170 Operator's Signature

Comments: Stabilities failed. Will perform an optical calibration

MX

2020.12.09
13:48:26
-05'00'