



# INSTRUMENT PROCESSING SHEET

Agency Brevard CSOS/N 80-001264Florida Department of  
Law EnforcementDate In 11/03/2020 DI Completion Date 11/04/2020 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>TDG</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>TDG</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>129</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.136</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.476</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By <u>MH</u> Flow Column # <u>ATP 101</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>130/130</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP 106</u> 32 mm <u>0.136/0.144</u> (.139 - .169) 36 mm <u>0.152/0.156</u> (.156 - .190) 53 mm <u>0.238/0.234</u> (.228 - .278) 103 mm <u>0.496/0.492</u> (.447 - .547)																																																											
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.09 08:25:34 -05'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>TDC</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.37</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +-2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080	MP4864	201905B 05/14/2021																																																											
0.200	MP5097	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG003005 01/30/2022																																																											
<b>Calibration Adjustment</b> Performed By <u>MH</u> Barometric Pressure Gauge <u>1020</u> ID # <u>28663</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>20190</td> <td>04/06/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005</td> <td>01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	20060	02/10/2022	0.100	MP5099	20190	04/06/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	MP4863	201905A	05/14/2021	0.080	MP4864	201905B	05/14/2021	0.200	MP5097	201904D	04/30/2021	0.080 DGS	N/A	AG003005	01/30/2022	<b>Department Inspection</b> Performed By <u>MH</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	MP5095	N/A	N/A																																																										
0.040	MP5098	20060	02/10/2022																																																										
0.100	MP5099	20190	04/06/2022																																																										
0.200	MP5100	20160	03/18/2022																																																										
0.300	MP5101	20030	01/21/2022																																																										
0.080 DGS	N/A	08819080A1	06/05/2021																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050	MP4863	201905A	05/14/2021																																																										
0.080	MP4864	201905B	05/14/2021																																																										
0.200	MP5097	201904D	04/30/2021																																																										
0.080 DGS	N/A	AG003005	01/30/2022																																																										
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use _____ 2020.11.09 <i>Doral Soto</i> 11-05-2020 <i>CA</i> 08:24:20 Tech Review / Date      Admin Review / Date																																																												

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BREVARD COUNTY S.O.  
Time of Inspection: 15:55

Date of Inspection: 11/04/2020

Serial Number: 80-001264  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079

Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5


Remarks:



2020.11.  
09  
08:23:34  
-05'00'

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

11/04/2020  
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001264	Brevard CSO	11/04/2020	TDG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>BREVARD COUNTY S.O. Intoxilyzer - Alconol Analyzer Model: 8000 11/04/2020 SN 80-001264 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:46 Control Test 0.050 09:47 Air Blank 0.000 09:47 Control Test 0.050 09:48 Air Blank 0.000 09:49 Control Test 0.050 09:49 Air Blank 0.000 09:50 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature: <i>MG</i></p>	<p>BREVARD COUNTY S.O. Intoxilyzer - Alconol Analyzer Model: 8000 11/04/2020 SN 80-001264 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:53 Control Test 0.081 09:54 Air Blank 0.000 09:55 Control Test 0.080 09:55 Air Blank 0.000 09:56 Control Test 0.080 09:57 Air Blank 0.000 09:57 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187</p> <p>Operator's Signature: <i>MG</i></p>	<p>BREVARD COUNTY S.O. Intoxilyzer - Alconol Analyzer Model: 8000 11/04/2020 SN 80-001264 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:01 Control Test 0.201 10:02 Air Blank 0.000 10:02 Control Test 0.199 10:03 Air Blank 0.000 10:03 Control Test 0.199 10:04 Air Blank 0.000 10:05 Control Test Stats Average 0.1997 Std Dev 0.0012 Rel Std Dev(%) 0.5783</p> <p>Operator's Signature: <i>MG</i></p>	<p>BREVARD COUNTY S.O. Intoxilyzer - Alconol Analyzer Model: 8000 11/04/2020 SN 80-001264 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:16 Control Test 0.076 10:16 Air Blank 0.000 10:17 Control Test 0.077 10:17 Air Blank 0.000 10:18 Control Test 0.078 10:18 Air Blank 0.000 10:18 Control Test Stats Average 0.0770 Std Dev 0.0010 Rel Std Dev(%) 1.2987</p> <p>Operator's Signature: <i>MG</i></p>

**Comments:** Failed stabilities - Will perform an optical calibration. *MG 11/04/2020*



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001264, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001264</u>	UNCERTAINTY* ±
Owning Agency:	<u>BREVARD COUNTY S.O.</u>	0.050 g/ 210 L      0.004
Calibration Date:	<u>11/04/2020</u>	0.080 g/ 210 L      0.005
Calibration Time:	<u>15:55</u>	0.200 g/ 210 L      0.007
		0.080 g/ 210 L Dry Gas Control      0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/04/2020      Date  
Michael D. Haughey      MICHAEL D HAUGHEY,  
Department Inspector

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

2020.11.09 08:22:10 -05'00'

INTOXILYZER 8000  
Instrument Initialization  
12-23 11/04/2020

BREUARD COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
SN 80-001264  
11/04/2020 12:53:12

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12597, Sum Io = 13289

\*\*\*\* CHANNEL 1 \*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.890 (-0.0170)  
Sample #2 = 0.8470 (0.0260)  
Sample #3 = 0.8570 (0.0320)  
Sample #4 = 0.8520 (0.0520)  
Avg % Abs = 0.8520 (0.0367)  
STD DEV = 0.0050 (0.0136)  
REL STD DEV = 0.587 (37.128)

\*\*\*\* CHANNEL 2 \*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5340 (-0.0020)  
Sample #2 = 1.4760 (0.0630)  
Sample #3 = 1.5070 (0.0400)  
Sample #4 = 1.4910 (0.0710)  
Avg % Abs = 1.4913 (0.0580)  
STD DEV = 0.0155 (0.0161)  
REL STD DEV = 1.040 (27.747)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12587, Sum Io = 13279

\*\*\*\* CHANNEL 1 \*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 2.0020 (-0.0240)  
Sample #2 = 1.9760 (0.0260)  
Sample #3 = 1.9940 (0.0310)  
Sample #4 = 2.0000 (0.0410)  
Avg % Abs = 1.9900 (0.0327)  
STD DEV = 0.0125 (0.0076)  
REL STD DEV = 0.628 (23.380)

\*\*\*\* CHANNEL 2 \*\*\*\*  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5370 (-0.0260)  
Sample #2 = 3.5090 (0.0300)  
Sample #3 = 3.5090 (0.0290)  
Sample #4 = 3.5260 (0.0360)  
Avg % Abs = 3.5147 (0.0317)  
STD DEV = 0.0098 (0.0038)  
REL STD DEV = 0.279 (11.956)

\*\*\*\* AUTO CAL DATA \*\*\*\*  
\*\*\*\* CHANNEL 1 \*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.126  
Std Dev = 0.01 Rel Std Dev = 4.66  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.852  
Std Dev = 0.00 Rel Std Dev = 0.59  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.990  
Std Dev = 0.01 Rel Std Dev = 0.63  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.816  
Std Dev = 0.02 Rel Std Dev = 0.42  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.517  
Std Dev = 0.04 Rel Std Dev = 0.80  
Zero Order Coef = -263.82  
First Order Coef = 2450.96  
Second Order Coef = 33.11  
Standard Deviation = 45.251523

\*\*\*\* CHANNEL 2 \*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.128  
Std Dev = 0.00 Rel Std Dev = 1.97  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.491  
Std Dev = 0.02 Rel Std Dev = 1.04  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.515  
Std Dev = 0.01 Rel Std Dev = 0.28  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.747  
Std Dev = 0.05 Rel Std Dev = 0.79  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.731  
Std Dev = 0.07 Rel Std Dev = 0.69  
Zero Order Coef = -149.95  
First Order Coef = 1341.06  
Second Order Coef = 14.49  
Standard Deviation = 25.713838

\*\*\*\* CHANNEL 1 \*\*\*\*  
Sample #1 = 2921.00  
Sample #2 = 2947.00  
Sample #3 = 3063.00  
Sample #4 = 3002.00  
Average Result = 3004.0000  
STD DEV = 58.1259  
REL STD DEV = 1.932  
\*\*\*\*\*

\*\*\*\* CHANNEL 2 \*\*\*\*  
Sample #1 = 3252.00  
Sample #2 = 3302.00  
Sample #3 = 3324.00  
Sample #4 = 3322.00  
Average Result = 3316.0000  
STD DEV = 12.1655  
REL STD DEV = 0.367  
\*\*\*\*\*

Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1020  
3 um H2O Adjust (mg/l x 0.000) = 805  
9 um H2O Adjust (mg/l x 0.000) = 493  
\*\*\*\* AUTO CAL PASS

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 0.000 -0.0005  
0.040 0.040 0.0005  
0.100 0.100 0.0004  
0.200 0.201 -0.0007  
0.300 0.300 0.0003

Sol Value = 0.060 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\* CHANNEL 1 \*\*\*\*  
Sample #1 = 2921.00  
Sample #2 = 2947.00  
Sample #3 = 3063.00  
Sample #4 = 3002.00  
Average Result = 3004.0000  
STD DEV = 58.1259  
REL STD DEV = 1.932  
\*\*\*\*\*

\*\*\*\* CHANNEL 2 \*\*\*\*  
Sample #1 = 3252.00  
Sample #2 = 3302.00  
Sample #3 = 3324.00  
Sample #4 = 3322.00  
Average Result = 3316.0000  
STD DEV = 12.1655  
REL STD DEV = 0.367  
\*\*\*\*\*

Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1020  
3 um H2O Adjust (mg/l x 0.000) = 805  
9 um H2O Adjust (mg/l x 0.000) = 493  
\*\*\*\* AUTO CAL PASS

80-001264  
Cal Adj  
11/4/2020  
MLX  
2020.11  
09  
08:21:3  
6-05'00

<b>TYPE OF TEST</b>	<b>SERIAL NUMBER</b>	<b>AGENCY</b>	<b>DATE</b>	<b>PERFORMED BY</b>
Stabilities Test	80-001264	Brevard County SO	11/4/2020	MM

<b>0.05g/210L</b>	<b>0.08g/210L</b>	<b>0.20g/210L</b>	<b>DGS 0.08g/210L</b>
<b>0.047 to 0.053</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>	<b>0.194 to 0.206</b> <input checked="" type="checkbox"/>	<b>0.077 to 0.083</b> <input checked="" type="checkbox"/>

BREVARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:38
Control Test	0.049	13:39
Air Blank	0.000	13:39
Control Test	0.047	13:40
Air Blank	0.000	13:41
Control Test	0.048	13:41
Air Blank	0.000	13:42
Control Test Stats		
Average	0.0480	
Std Dev	0.0010	
Rel Std Dev(%)	2.0833	

*MM*  
 Operator's Signature

BREVARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:43
Control Test	0.079	13:44
Air Blank	0.000	13:44
Control Test	0.078	13:45
Air Blank	0.000	13:46
Control Test	0.079	13:46
Air Blank	0.000	13:47
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*MM*  
 Operator's Signature

BREVARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:48
Control Test	0.199	13:49
Air Blank	0.000	13:49
Control Test	0.198	13:50
Air Blank	0.000	13:51
Control Test	0.197	13:51
Air Blank	0.000	13:52
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

*MM*  
 Operator's Signature

BREVARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:54
Control Test	0.078	13:54
Air Blank	0.000	13:55
Control Test	0.080	13:55
Air Blank	0.000	13:55
Control Test	0.080	13:55
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0793	
Std Dev	0.0012	
Rel Std Dev(%)	1.4555	

*MM*  
 Operator's Signature

CO-001264

# Flow Calibration

11/4/2020

MPD

#2

BREWARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
    SORT(Diff) ) = 6.402  
 2: Rate (Liters/min) = 15  
    SORT(Diff) ) = 11.660  
 3: Rate (Liters/min) = 30  
    SORT(Diff) ) = 21.094  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 660  
 Rounded Intercept = -557112  
 Correlation = 0.99886

BREWARD COUNTY S.O.  
 Intoxilyzer - Alcohol Analyzer  
 Model: 8000 SN 80-001264  
 11/04/2020  
 Software: 8100.27

Flow Rate Calibration\*\*\*\*\*  
 1: Rate (Liters/min) = 5  
    SORT(Diff) ) = 6.555  
 2: Rate (Liters/min) = 15  
    SORT(Diff) ) = 11.574  
 3: Rate (Liters/min) = 30  
    SORT(Diff) ) = 20.879  
 Dependent Data Scale Factor = 100000 L/min  
 Independent Data Scale Factor = 256  
 Rounded Slope = 675  
 Rounded Intercept = -580884  
 Correlation = 0.99843

SD

2020.11.  
 09  
 08:19:55  
 -05'00'