



INSTRUMENT PROCESSING SHEET

Agency Tyndall AFB

S/N 80-001646

Florida Department of
Law Enforcement

Date In 10-26-2020 DI Completion Date 10-30-2020

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____		Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 215 ☒ Flow Verification (L/s) Flow Column # ATP-105 32 mm 0.164 (.139 - .169) 36 mm 0.175 (.156 - .190) 53 mm 0.250 (.228 - .278) 103 mm 0.519 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 28427 ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>201905A 05-14-2021</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>201905B 05-14-2021</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>201904D 04-30-2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	201905A 05-14-2021	0.080	MP5089	201905B 05-14-2021	0.200	MP5090	201904D 04-30-2021	0.080 DGS	N/A	AG931603 11-12-2021	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																								
Simulator	Serial #	Lot #/Exp																																									
0.050	MP5088	201905A 05-14-2021																																									
0.080	MP5089	201905B 05-14-2021																																									
0.200	MP5090	201904D 04-30-2021																																									
0.080 DGS	N/A	AG931603 11-12-2021																																									
Final Release Date FDLE Alcohol Testing Program		Digitally signed by FDLE Alcohol Testing Program Date: 2020.11.04 11:50:56 -05'00'		Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																							
Calibration Adjustment Performed By <u>IS</u> Barometric Pressure Gauge 1007 ID # 28421		Department Inspection Performed By <u>IS</u> Barometric Pressure ID# 30793 Gauge 1017 Instrument 1017 Mouth Alcohol Solution Lot # 2019-B Acetone Stock Solution Lot # 2019-A																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20190</td> <td>04-06-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06-05-2021</td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20190	04-06-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	08819080A1	06-05-2021	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	MP5091	N/A	N/A																																								
0.040	MP5082	20060	02-10-2022																																								
0.100	MP5083	20190	04-06-2022																																								
0.200	MP5084	20160	03-18-2022																																								
0.300	MP5085	20030	01-21-2022																																								
0.080 DGS	N/A	08819080A1	06-05-2021																																								
Simulator	Serial Number																																										
0.000	MP5086																																										
Interferent	MP5087																																										
0.050	MP5088																																										
0.080	MP5089																																										
0.200	MP5090																																										
☒ Post Calibration Adjustment Stability Checks		Attachments <table border="1"> <tbody> <tr> <td>☒ Form 41</td> <td>☒ Post-Stability Checks</td> </tr> <tr> <td>☒ Stability Checks</td> <td>☐ Flow Calibration</td> </tr> <tr> <td>☒ Calibration Certificate</td> <td>☐ Form 40</td> </tr> <tr> <td>☒ Calibration Adjustment</td> <td>☐ Other _____</td> </tr> </tbody> </table>				☒ Form 41	☒ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☒ Calibration Adjustment	☐ Other _____																														
☒ Form 41	☒ Post-Stability Checks																																										
☒ Stability Checks	☐ Flow Calibration																																										
☒ Calibration Certificate	☐ Form 40																																										
☒ Calibration Adjustment	☐ Other _____																																										
Notes/Suggested Service: Admin Review: corrected serial # on Post stability checks 11-04-2020 0.050 MP5088 0.080 MP5089 0.200 MP5090		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Michael D. Hughes 2020.11.02 15:31:03 -05'00' Tech Review / Date Admin Review 2020.11.04 11:50:12 -05'00'																																									

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TYNDALL AFB

Serial Number: 80-001646

Time of Inspection: 09:49

Date of Inspection: 10/30/2020

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#: 201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#: 201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG011102 Exp: 04/20/2022
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
Standard Deviations	0.0005	0.0003	0.0004	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

10/30/2020

Date


MX

CR 2020.11
04
11:49:4
4-05'00'

Stability Checks

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.047	10:25
Air Blank	0.000	10:25
Control Test	0.047	10:26
Air Blank	0.000	10:26
Control Test	0.048	10:27
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	


Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:29
Control Test	0.078	10:29
Air Blank	0.000	10:30
Control Test	0.078	10:30
Air Blank	0.000	10:31
Control Test	0.077	10:32
Air Blank	0.000	10:32
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Wet


Operator's Signature

TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
10/28/2020
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.198	10:34
Air Blank	0.000	10:35
Control Test	0.197	10:36
Air Blank	0.000	10:36
Control Test	0.198	10:37
Air Blank	0.000	10:37
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	


Operator's Signature


TYNDALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001646
10/28/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.082	10:39
Air Blank	0.000	10:39
Control Test	0.082	10:40
Air Blank	0.000	10:40
Control Test	0.082	10:41
Air Blank	0.000	10:41
Control Test Stats		
Average	0.0820	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry


Operator's Signature

MX
2020.11
04
11:49:2
2
-05'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001646, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001646</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>TYNDALL AFB</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/30/2020</u>	0.080 g/ 210 L 0.005
Calibration Time:	<u>09:49</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date 10/30/2020
Israel Soto
ISRAEL SOTO,
Department Inspector

Page 1 of 1

2020.11.
04
11:48:52
-05'00'

[Handwritten signature]

OPT. Bench Cal Adj.

TYNALL AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 88-001646
10/29/2020 12:47:07

Auto Calibration
Max Power Res Value = 100
Auto Range Res Value = 72

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12698, Sum Io = 13598

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.1000	(-0.0170)	
Sample #2 = 0.1070	(0.0050)	
Sample #3 = 0.0940	(0.0390)	
Sample #4 = 0.1110	(0.0390)	
Avg % Abs = 0.1040	(0.0273)	
STD DEV = 0.0089	(0.0193)	
REL STD DEV = 8.546	(70.784)	

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.0940	(0.0030)	
Sample #2 = 0.1060	(0.0060)	
Sample #3 = 0.1100	(0.0220)	
Sample #4 = 0.1310	(0.0160)	
Avg % Abs = 0.1163	(0.0147)	
STD DEV = 0.0127	(0.0081)	
REL STD DEV = 10.952	(55.111)	

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12685, Sum Io = 13591

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.8110	(-0.0350)	
Sample #2 = 0.8010	(-0.0250)	
Sample #3 = 0.8120	(0.0230)	
Sample #4 = 0.8110	(0.0160)	
Avg % Abs = 0.8080	(0.0113)	
STD DEV = 0.0061	(0.0146)	
REL STD DEV = 0.753	(128.573)	

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 1.4830	(-0.0200)	
Sample #2 = 1.4900	(-0.0170)	
Sample #3 = 1.4690	(0.0061)	
Sample #4 = 1.4780	(0.0000)	
Avg % Abs = 1.4790	(-0.0037)	
STD DEV = 0.0105	(0.0119)	
REL STD DEV = 0.712	(325.373)	

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12678, Sum Io = 13589

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 1.9090	(-0.0160)	
Sample #2 = 1.8890	(0.0270)	
Sample #3 = 1.8820	(0.0500)	
Sample #4 = 1.8470	(0.0630)	
Avg % Abs = 1.8727	(0.0467)	
STD DEV = 0.0225	(0.0182)	
REL STD DEV = 1.202	(39.064)	

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 3.5240	(0.0000)	
Sample #2 = 3.5110	(0.0130)	
Sample #3 = 3.5040	(0.0220)	
Sample #4 = 3.4910	(0.0380)	
Avg % Abs = 3.5020	(0.0243)	
STD DEV = 0.0101	(0.0127)	
REL STD DEV = 0.290	(52.037)	

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12667, Sum Io = 13582

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 3.6200	(-0.0360)	
Sample #2 = 3.5890	(0.0.00)	
Sample #3 = 3.6050	(-0.0060)	
Sample #4 = 3.5770	(0.0010)	
Avg % Abs = 3.5963	(0.00.7)	
STD DEV = 0.0140	(0.0080)	
REL STD DEV = 0.391	(481.248)	

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 6.7260	(-0.0190)	
Sample #2 = 6.7100	(0.0040)	
Sample #3 = 6.7050	(0.0020)	
Sample #4 = 6.7030	(-0.0010)	
Avg % Abs = 6.7060	(0.0017)	
STD DEV = 0.0036	(0.0025)	
REL STD DEV = 0.054	(150.997)	

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12665, Sum Io = 13581

<<<< CHANNEL 1 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 5.2070	(-0.0180)	
Sample #2 = 5.2000	(0.0340)	
Sample #3 = 5.2020	(0.0320)	
Sample #4 = 5.2080	(0.0340)	
Avg % Abs = 5.2033	(0.0333)	
STD DEV = 0.0042	(0.0012)	
REL STD DEV = 0.080	(3.464)	

<<<< CHANNEL 2 >>>>

Sample	% Abs	(% Abs Ref)
Sample #1 = 9.7050	(-0.0080)	
Sample #2 = 9.6850	(0.0260)	
Sample #3 = 9.6870	(0.0170)	
Sample #4 = 9.6720	(0.0320)	
Avg % Abs = 9.6813	(0.0250)	
STD DEV = 0.0081	(0.0075)	
REL STD DEV = 0.084	(30.199)	

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.104
Std Dev = 0.01 Rel Std Dev = 8.55
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.808
Std Dev = 0.01 Rel Std Dev = 0.75
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.873
Std Dev = 0.02 Rel Std Dev = 1.20
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.590
Std Dev = 0.01 Rel Std Dev = 0.39
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.203
Std Dev = 0.00 Rel Std Dev = 0.08
Zero Order Coef = -245.64
First Order Coef = 2595.65
Second Order Coef = 37.35
Standard Deviation = 26.727715

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.116
Std Dev = 0.01 Rel Std Dev = 10.95
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.479
Std Dev = 0.01 Rel Std Dev = 0.71
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.502
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.706
Std Dev = 0.00 Rel Std Dev = 0.05
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.681
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -138.05
First Order Coef = 1343.83
Second Order Coef = 14.98
Standard Deviation = 19.987957

MX

OPT. Bench Cal Adj

Post Stability Checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0005
0.040	0.039	0.0006
0.100	0.100	0.0003
0.200	0.201	-0.0007
0.300	0.300	0.0003

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0004
0.040	0.040	0.0005
0.100	0.100	0.0002
0.200	0.200	-0.0005
0.300	0.300	0.0002


Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3022.00
 Sample #2 = 3041.00
 Sample #3 = 3028.00
 Sample #4 = 3057.00
 Average Result = 3042.0000
 STD DEV = 14.5258
 REL STD DEV = 0.478

 ***** CHANNEL 2
 Sample #1 = 3353.00
 Sample #2 = 3350.00
 Sample #3 = 3374.00
 Sample #4 = 3360.00
 Average Result = 3364.6667
 STD DEV = 8.0829
 REL STD DEV = 0.240

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1006
 3 um H2O Adjust (mg/l)*10,000 = 767
 9 um H2O Adjust (mg/l)*10,000 = 445
 **** AUTO CAL PASS


TYNDALL AFB
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001646
 10/30/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:48
Control Test	0.048	07:48
Air Blank	0.000	07:49
Control Test	0.048	07:50
Air Blank	0.000	07:50
Control Test	0.049	07:51
Air Blank	0.000	07:51
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1545	


 Operator's Signature

TYNDALL AFB
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001646
 10/30/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:57
Control Test	0.199	07:58
Air Blank	0.000	07:59
Control Test	0.200	07:59
Air Blank	0.000	08:00
Control Test	0.199	08:01
Air Blank	0.000	08:01
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	


 Operator's Signature

TYNDALL AFB
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001646
 10/30/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:52
Control Test	0.078	07:53
Air Blank	0.000	07:54
Control Test	0.079	07:54
Air Blank	0.000	07:55
Control Test	0.079	07:56
Air Blank	0.000	07:56
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet


 Operator's Signature

TYNDALL AFB
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001646
 10/30/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:02
Control Test	0.080	08:03
Air Blank	0.000	08:03
Control Test	0.080	08:03
Air Blank	0.000	08:04
Control Test	0.080	08:04
Air Blank	0.000	08:05
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry


 Operator's Signature

2020.11.
 04
 11:44:03
 -05'00"

MX