

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI

Serial Number: 80-003408

Time of Inspection: 11:44

Date of Inspection: 07/13/2020

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.081	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.081	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079

Standard Deviations	0.0004	0.0004	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

M8
BK 2020.07.17
11:56:52
-04'00"

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

07/13/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-003408	Florida Highway Patrol	07/13/2020	<i>HL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053 ☒	0.077 to 0.083 ☒	0.194 to 0.206 ☒	0.077 to 0.083 ☒																																																																																																																																				
FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 07/13/2020 Software: 8100.27 SN 80-003408 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>09:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>David Hays</i>	Air Blank	0.000	09:05	Control Test	0.049	09:05	Air Blank	0.000	09:06	Control Test	0.049	09:07	Air Blank	0.000	09:07	Control Test	0.049	09:08	Air Blank	0.000	09:08	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 07/13/2020 Software: 8100.27 SN 80-003408 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr> <tr><td>Control Test</td><td>0.081</td><td>09:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </table> Operator's Signature <i>David Hays</i>	Air Blank	0.000	09:10	Control Test	0.080	09:11	Air Blank	0.000	09:11	Control Test	0.081	09:12	Air Blank	0.000	09:12	Control Test	0.080	09:13	Air Blank	0.000	09:14	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 07/13/2020 Software: 8100.27 SN 80-003408 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:15</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.199</td><td>09:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </table> Operator's Signature <i>David Hays</i>	Air Blank	0.000	09:15	Control Test	0.198	09:15	Air Blank	0.000	09:16	Control Test	0.199	09:17	Air Blank	0.000	09:17	Control Test	0.198	09:18	Air Blank	0.000	09:18	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 07/13/2020 Software: 8100.27 SN 80-003408 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>David Hays</i>	Air Blank	0.000	09:20	Control Test	0.079	09:21	Air Blank	0.000	09:21	Control Test	0.079	09:21	Air Blank	0.000	09:22	Control Test	0.079	09:22	Air Blank	0.000	09:23	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	09:05																																																																																																																																					
Control Test	0.049	09:05																																																																																																																																					
Air Blank	0.000	09:06																																																																																																																																					
Control Test	0.049	09:07																																																																																																																																					
Air Blank	0.000	09:07																																																																																																																																					
Control Test	0.049	09:08																																																																																																																																					
Air Blank	0.000	09:08																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0490																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	09:10																																																																																																																																					
Control Test	0.080	09:11																																																																																																																																					
Air Blank	0.000	09:11																																																																																																																																					
Control Test	0.081	09:12																																																																																																																																					
Air Blank	0.000	09:12																																																																																																																																					
Control Test	0.080	09:13																																																																																																																																					
Air Blank	0.000	09:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0803																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7187																																																																																																																																						
Air Blank	0.000	09:15																																																																																																																																					
Control Test	0.198	09:15																																																																																																																																					
Air Blank	0.000	09:16																																																																																																																																					
Control Test	0.199	09:17																																																																																																																																					
Air Blank	0.000	09:17																																																																																																																																					
Control Test	0.198	09:18																																																																																																																																					
Air Blank	0.000	09:18																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1983																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2911																																																																																																																																						
Air Blank	0.000	09:20																																																																																																																																					
Control Test	0.079	09:21																																																																																																																																					
Air Blank	0.000	09:21																																																																																																																																					
Control Test	0.079	09:21																																																																																																																																					
Air Blank	0.000	09:22																																																																																																																																					
Control Test	0.079	09:22																																																																																																																																					
Air Blank	0.000	09:23																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

MA BK



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003408, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003408</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/13/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:44</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/13/2020

Date

DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

MA

2020.07.17
11:57:41
-04'00'

BK