



# INSTRUMENT PROCESSING SHEET

Agency Lakeland Police DepartmentS/N 80-003945

Florida Department of Law Enforcement

Date In 06/26/2020 DI Completion Date 7/21/20 Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>RAW</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>3 feet missing from instrument.</u>	<b>Quality Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>193</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
<b>Final Release Date</b> FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.07.24 08:03:02 -04'00'	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1018</u></td> <td><u>201905A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>SD3933</u></td> <td><u>201905B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>G2078</u></td> <td><u>201904D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>	0.080	<u>SD3933</u>	<u>201905B</u> <u>05-14-2021</u>	0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>SP</u> <input checked="" type="checkbox"/> Lab Temp °C <u>21.17</u> External Digital Therm. ID#: <u>300505</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5088</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5089</u> <input checked="" type="checkbox"/> 34°C +- .2 Serial #: <u>MP5090</u>																								
Simulator	Serial #	Lot #/Exp																																							
0.050	<u>SD1018</u>	<u>201905A</u> <u>05-14-2021</u>																																							
0.080	<u>SD3933</u>	<u>201905B</u> <u>05-14-2021</u>																																							
0.200	<u>G2078</u>	<u>201904D</u> <u>04-30-2021</u>																																							
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																							
<b>Calibration Adjustment</b> Performed By <u>SP</u> Barometric Pressure Gauge <u>1015</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5091</u></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td><u>MP5082</u></td> <td><u>20060</u></td> <td><u>2-10-22</u></td> </tr> <tr> <td>0.100</td> <td><u>MP5083</u></td> <td><u>20190</u></td> <td><u>4-6-22</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5084</u></td> <td><u>20160</u></td> <td><u>3-18-22</u></td> </tr> <tr> <td>0.300</td> <td><u>MP5085</u></td> <td><u>20030</u></td> <td><u>1-21-22</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>08819080A</u></td> <td><u>6-5-21</u></td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks	Simulator	Serial Number	Lot Number	Expiration	0.000	<u>MP5091</u>	N/A	N/A	0.040	<u>MP5082</u>	<u>20060</u>	<u>2-10-22</u>	0.100	<u>MP5083</u>	<u>20190</u>	<u>4-6-22</u>	0.200	<u>MP5084</u>	<u>20160</u>	<u>3-18-22</u>	0.300	<u>MP5085</u>	<u>20030</u>	<u>1-21-22</u>	0.080 DGS	N/A	<u>08819080A</u>	<u>6-5-21</u>	<b>Department Inspection</b> Performed By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1015</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																						
0.000	<u>MP5091</u>	N/A	N/A																																						
0.040	<u>MP5082</u>	<u>20060</u>	<u>2-10-22</u>																																						
0.100	<u>MP5083</u>	<u>20190</u>	<u>4-6-22</u>																																						
0.200	<u>MP5084</u>	<u>20160</u>	<u>3-18-22</u>																																						
0.300	<u>MP5085</u>	<u>20030</u>	<u>1-21-22</u>																																						
0.080 DGS	N/A	<u>08819080A</u>	<u>6-5-21</u>																																						
Simulator	Serial Number																																								
0.000	<u>MP5086</u>																																								
Interferent	<u>MP5087</u>																																								
0.050	<u>MP5088</u>																																								
0.080	<u>MP5089</u>																																								
0.200	<u>MP5090</u>																																								
Notes/Suggested Service: <u>Added missing feet. SP</u> <u>Conducted cal adjustment due to barometric pressure difference &gt;1 %. SP</u>	<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																								
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
David Eliezer Reyes Rivera <small>Digitally signed by David Eliezer Reyes Rivera Date: 2020.07.23 14:10:47 -04'00'</small> <u>Brett Kirkland</u> 2020.07.24 07:59:55 -04'00' Tech Review / Date _____ Admin Review / Date _____																																									



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003945, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003945</u>	UNCERTAINTY* ±	
Owning Agency:	<u>LAKELAND PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/21/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:10</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/21/2020 Date  
Shayla Platt  
SHAYLA D PLATT,  
Department Inspector

FDLE/ATP Form 69 April 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 11:10

Date of Inspection: 07/21/2020

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.079

Standard Deviations	0.0004	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT  
Signature and Printed Name

07/21/2020  
Date

# stability checks

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003945  
06/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:20
Control Test	0.047	14:20
Air Blank	0.000	14:21
Control Test	0.048	14:22
Air Blank	0.000	14:22
Control Test	0.047	14:23
Air Blank	0.000	14:23
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	



Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003945  
06/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:25
Control Test	0.078	14:26
Air Blank	0.000	14:27
Control Test	0.078	14:27
Air Blank	0.000	14:28
Control Test	0.078	14:29
Air Blank	0.000	14:29
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003945  
06/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:34
Control Test	0.201	14:35
Air Blank	0.000	14:36
Control Test	0.201	14:36
Air Blank	0.000	14:37
Control Test	0.201	14:38
Air Blank	0.000	14:38
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003945  
06/26/2020  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.081	14:40
Air Blank	0.000	14:41
Control Test	0.079	14:41
Air Blank	0.000	14:42
Control Test	0.080	14:42
Air Blank	0.000	14:43
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

Dry



Operator's Signature

DERR Digitally signed  
by DERR  
Date: 2020.07.23  
14:08:33 -04'00'

BK 2020.07.24  
08:01:14  
-04'00'

LAKELAND PD  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000  
 07/21/2026  
 SN 80-003945  
 07:32:46  
 Auto Calibration  
 Max Power Res Value = 34  
 Auto Range Res Value = 20

<<<< CHANNEL 2 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.4930 (-0.0060)  
 Sample #2 = 1.4810 (0.0110)  
 Sample #3 = 1.4790 (0.0150)  
 Sample #4 = 1.4930 (0.0100)  
 Avg % Abs = 1.4843 (0.0120)  
 STD DEV = 0.0076 (0.0026)  
 REL STD DEV = 0.510 (22.048)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
 <<<< CHANNEL 1 >>>>  
 Sol Val = 0.000 mg/l or 0.000 g/210L  
 % Abs = 0.074  
 Std Dev = 0.02 Rel Std Dev = 23.41  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 0.733  
 Std Dev = 0.00 Rel Std Dev = 0.68  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 1.716  
 Std Dev = 0.01 Rel Std Dev = 0.61  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 3.312  
 Std Dev = 0.01 Rel Std Dev = 0.36  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 4.838  
 Std Dev = 0.02 Rel Std Dev = 0.31  
 Zero Order Coef = -197.38  
 First Order Coef = 2826.49  
 Second Order Coef = 34.33  
 Standard Deviation = 12.504560

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0002  
 0.040 0.040 0.0002  
 0.100 0.100 0.0002  
 0.200 0.200 -0.0003  
 0.300 0.300 0.0001

Sol Value = 0.100 g/210L \*\*\*  
 Fit Value = 0.4762 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11419, Sum Io = 14272  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 1.7300 (-0.0140)  
 Sample #2 = 1.7040 (0.0150)  
 Sample #3 = 1.7240 (0.0090)  
 Sample #4 = 1.7190 (0.0290)  
 Avg % Abs = 1.7157 (0.0177)  
 STD DEV = 0.0104 (0.0103)  
 REL STD DEV = 0.607 (58.094)

Sol Value = 0.300 g/210L \*\*\*  
 Fit Value = 1.4286 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11416, Sum Io = 14266  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 4.8530 (-0.0046)  
 Sample #2 = 4.8390 (0.0220)  
 Sample #3 = 4.8220 (0.0240)  
 Sample #4 = 4.8520 (0.0180)  
 Avg % Abs = 4.8377 (0.0213)  
 STD DEV = 0.0150 (0.0031)  
 REL STD DEV = 0.311 (14.321)

Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 3177.00  
 Sample #2 = 3144.00  
 Sample #3 = 3147.00  
 Sample #4 = 3133.00  
 Average Result = 3141.3333  
 STD DEV = 7.3711  
 REL STD DEV = 0.235  
 \*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3356.00  
 Sample #2 = 3340.00  
 Sample #3 = 3368.00  
 Sample #4 = 3349.00  
 Average Result = 3352.3333  
 STD DEV = 14.2945  
 REL STD DEV = 0.426  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1015  
 3 UT H2O Adjust (mg/l\*10,000) = 668  
 9 UT H2O Adjust (mg/l\*10,000) = 457  
 \*\*\*\*\* AUTO CAL PASS

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0003  
 0.040 0.040 0.0004  
 0.100 0.100 0.0001  
 0.200 0.200 -0.0003  
 0.300 0.300 0.0001

Sol Value = 0.040 g/210L \*\*\*  
 Fit Value = 0.1905 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11422, Sum Io = 14272  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 0.7340 (0.0010)  
 Sample #2 = 0.7360 (0.0060)  
 Sample #3 = 0.7280 (0.0160)  
 Sample #4 = 0.7330 (0.0220)  
 Avg % Abs = 0.7330 (0.0147)  
 STD DEV = 0.0050 (0.0081)  
 REL STD DEV = 0.682 (55.111)

Sol Value = 0.200 g/210L \*\*\*  
 Fit Value = 0.9524 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 Sum Io = 11417, Sum Io = 14267  
 <<<< CHANNEL 1 >>>>  
 Sample % Abs (% Abs Ref)  
 Sample #1 = 3.3520 (-0.0070)  
 Sample #2 = 3.3070 (0.0120)  
 Sample #3 = 3.3260 (0.0050)  
 Sample #4 = 3.3020 (0.0170)  
 Avg % Abs = 3.3117 (0.0113)  
 STD DEV = 0.0127 (0.0060)  
 REL STD DEV = 0.382 (53.186)

Sol Value = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.099  
 Std Dev = 0.03 Rel Std Dev = 27.94  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.484  
 Std Dev = 0.01 Rel Std Dev = 0.51  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.530  
 Std Dev = 0.02 Rel Std Dev = 0.52  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.732  
 Std Dev = 0.01 Rel Std Dev = 0.29  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 9.752  
 Std Dev = 0.02 Rel Std Dev = 0.21  
 Zero Order Coef = -117.68  
 First Order Coef = 1326.92  
 Second Order Coef = 15.31  
 Standard Deviation = 14.733530

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0003  
 0.040 0.040 0.0004  
 0.100 0.100 0.0001  
 0.200 0.200 -0.0003  
 0.300 0.300 0.0001

Sol Value = 0.0000 mg/l or 0.000 g/210L  
 % Abs = 0.099  
 Std Dev = 0.03 Rel Std Dev = 27.94  
 Sol Val = 0.1905 mg/l or 0.040 g/210L  
 % Abs = 1.484  
 Std Dev = 0.01 Rel Std Dev = 0.51  
 Sol Val = 0.4762 mg/l or 0.100 g/210L  
 % Abs = 3.530  
 Std Dev = 0.02 Rel Std Dev = 0.52  
 Sol Val = 0.9524 mg/l or 0.200 g/210L  
 % Abs = 6.732  
 Std Dev = 0.01 Rel Std Dev = 0.29  
 Sol Val = 1.4286 mg/l or 0.300 g/210L  
 % Abs = 9.752  
 Std Dev = 0.02 Rel Std Dev = 0.21  
 Zero Order Coef = -117.68  
 First Order Coef = 1326.92  
 Second Order Coef = 15.31  
 Standard Deviation = 14.733530

Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 3177.00  
 Sample #2 = 3144.00  
 Sample #3 = 3147.00  
 Sample #4 = 3133.00  
 Average Result = 3141.3333  
 STD DEV = 7.3711  
 REL STD DEV = 0.235  
 \*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3356.00  
 Sample #2 = 3340.00  
 Sample #3 = 3368.00  
 Sample #4 = 3349.00  
 Average Result = 3352.3333  
 STD DEV = 14.2945  
 REL STD DEV = 0.426  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1015  
 3 UT H2O Adjust (mg/l\*10,000) = 668  
 9 UT H2O Adjust (mg/l\*10,000) = 457  
 \*\*\*\*\* AUTO CAL PASS

Sol Value = 0.080 g/210L \*\*\*  
 Fit Value = 0.3810 mg/l \*\*\*\*  
 Samples Taken = 4, Discarded = 1  
 \*\*\*\*\* CHANNEL 1  
 Sample #1 = 3177.00  
 Sample #2 = 3144.00  
 Sample #3 = 3147.00  
 Sample #4 = 3133.00  
 Average Result = 3141.3333  
 STD DEV = 7.3711  
 REL STD DEV = 0.235  
 \*\*\*\*\*  
 \*\*\*\*\* CHANNEL 2  
 Sample #1 = 3356.00  
 Sample #2 = 3340.00  
 Sample #3 = 3368.00  
 Sample #4 = 3349.00  
 Average Result = 3352.3333  
 STD DEV = 14.2945  
 REL STD DEV = 0.426  
 \*\*\*\*\*  
 Dry Gas H2O Adjust Results \*\*\*\*\*  
 Barometric Pressure = 1015  
 3 UT H2O Adjust (mg/l\*10,000) = 668  
 9 UT H2O Adjust (mg/l\*10,000) = 457  
 \*\*\*\*\* AUTO CAL PASS

Solution Stats Quadratic Fit Chan 2  
 Act Fit Residual  
 g/210L g/210L g/210L  
 0.000 0.000 -0.0003  
 0.040 0.040 0.0004  
 0.100 0.100 0.0001  
 0.200 0.200 -0.0003  
 0.300 0.300 0.0001

*Cal Adjustment*  
*#80-003945 8P*

# Post Cal Adjust Stability Checks

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN: 80-033945  
07/21/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:33
Control Test	0.059	08:34
Air Blank	0.000	08:34
Control Test	0.049	08:35
Air Blank	0.000	08:36
Control Test	0.049	08:36
Air Blank	0.000	08:37
Control Test Stats	0.0493	
Average	0.0006	
Std Dev	1.1703	
Rel Std Dev(%)		

SP

Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN: 80-033945  
07/21/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:28
Control Test	0.080	08:29
Air Blank	0.000	08:30
Control Test	0.079	08:31
Air Blank	0.000	08:31
Control Test	0.079	08:32
Air Blank	0.000	08:32
Control Test Stats	0.0793	
Average	0.0006	
Std Dev	0.7277	
Rel Std Dev(%)		

SP

Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN: 80-033945  
07/21/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:39
Control Test	0.200	08:40
Air Blank	0.000	08:40
Control Test	0.200	08:41
Air Blank	0.000	08:42
Control Test	0.199	08:42
Air Blank	0.000	08:43
Control Test Stats	0.1997	
Average	0.0006	
Std Dev	0.2892	
Rel Std Dev(%)		

SP

Operator's Signature

LAKELAND PD  
Intoxilyzer - Alcohol Analyzer  
Model: 8000 SN: 80-033945  
07/21/2020  
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	08:25
Control Test	0.081	08:25
Air Blank	0.000	08:25
Control Test	0.080	08:26
Air Blank	0.000	08:26
Control Test	0.080	08:27
Air Blank	0.000	08:27
Control Test Stats	0.0803	
Average	0.0006	
Std Dev	0.7187	
Rel Std Dev(%)		

DAS

SP

Operator's Signature

DERR  
Digitally signed by  
DERR  
Date: 2020.07.23  
14:06:25 -04'00'

2020.07.23  
4:08:01.59  
BK  
-04'00'