



INSTRUMENT PROCESSING SHEET

Agency Florida Wildlife Conservation Commission s/N 80-005248Florida Department of
Law EnforcementDate In 11/3/2020 DI Completion Date 11/3/2020 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ Final Release Date FDLE Alcohol Testing Program	Quality Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>152</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>238</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>23.37C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	SD3967	201905A 05/14/2021																																																											
0.080	SD3968	201905B 05/14/2021																																																											
0.200	SD3969	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG003005 01/30/2022																																																											
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1023</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date <u>11-04-2020</u> Admin Review / Date <u>2020.11.04 11:30:27</u> <div style="text-align: right;"> <u>2020.11.04 11:30:27</u> <u>-0500</u> </div>	Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3965																																																												
Interferent	SD3966																																																												
0.050	SD3967																																																												
0.080	SD3968																																																												
0.200	SD3969																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____																																																													



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005248, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005248</u>	UNCERTAINTY * \pm	
Owning Agency:	<u>FISH & WILDLIFE CC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/03/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>23:38</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

[Signature]
2020.11.04
11:27:32
2-05'00"

11/03/2020

Date

[Signature]
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 April 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 23:38

Date of Inspection: 11/03/2020

Serial Number: 80-005248
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.080	0.204	0.079
0.000	0.048	0.080	0.205	0.080
0.000	0.049	0.080	0.205	0.079
0.000	0.049	0.080	0.205	0.079
0.000	0.049	0.080	0.204	0.080
0.000	0.049	0.080	0.203	0.079
0.000	0.049	0.080	0.204	0.079
0.000	0.049	0.080	0.203	0.079
0.000	0.049	0.080	0.204	0.079
0.000	0.049	0.080	0.204	0.079

Standard Deviations	0.0004	0.0000	0.0007	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

DS

CP 2020.11.
04
11:26:54
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

11/03/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005248	Florida Wildlife Conservation Commission	11/3/2020	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>FISH & WILDLIFE CC Intoxilyzer - Alconol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-005248</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>18:58</td></tr><tr><td>Control Test</td><td>0.047</td><td>18:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>18:59</td></tr><tr><td>Control Test</td><td>0.047</td><td>19:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:01</td></tr><tr><td>Control Test</td><td>0.048</td><td>19:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0473</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2198</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	18:58	Control Test	0.047	18:59	Air Blank	0.000	18:59	Control Test	0.047	19:00	Air Blank	0.000	19:01	Control Test	0.048	19:01	Air Blank	0.000	19:02	Control Test Stats			Average	0.0473		Std Dev	0.0006		Rel Std Dev(%)	1.2198		<div>FISH & WILDLIFE CC Intoxilyzer - Alconol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-005248</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:03</td></tr><tr><td>Control Test</td><td>0.078</td><td>19:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:05</td></tr><tr><td>Control Test</td><td>0.078</td><td>19:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:05</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	19:03	Control Test	0.078	19:04	Air Blank	0.000	19:05	Control Test	0.078	19:05	Air Blank	0.000	19:05	Control Test	0.079	19:07	Air Blank	0.000	19:07	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<div>FISH & WILDLIFE CC Intoxilyzer - Alconol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-005248</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:13</td></tr><tr><td>Control Test</td><td>0.200</td><td>19:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:15</td></tr><tr><td>Control Test</td><td>0.199</td><td>19:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:16</td></tr><tr><td>Control Test</td><td>0.202</td><td>19:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:17</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2003</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7625</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	19:13	Control Test	0.200	19:14	Air Blank	0.000	19:15	Control Test	0.199	19:15	Air Blank	0.000	19:16	Control Test	0.202	19:17	Air Blank	0.000	19:17	Control Test Stats			Average	0.2003		Std Dev	0.0015		Rel Std Dev(%)	0.7625		<div>FISH & WILDLIFE CC Intoxilyzer - Alconol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-005248</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:18</td></tr><tr><td>Control Test</td><td>0.081</td><td>19:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:19</td></tr><tr><td>Control Test</td><td>0.080</td><td>19:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:20</td></tr><tr><td>Control Test</td><td>0.080</td><td>19:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	19:18	Control Test	0.081	19:19	Air Blank	0.000	19:19	Control Test	0.080	19:20	Air Blank	0.000	19:20	Control Test	0.080	19:20	Air Blank	0.000	19:21	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	18:58																																																																																																																																																	
Control Test	0.047	18:59																																																																																																																																																	
Air Blank	0.000	18:59																																																																																																																																																	
Control Test	0.047	19:00																																																																																																																																																	
Air Blank	0.000	19:01																																																																																																																																																	
Control Test	0.048	19:01																																																																																																																																																	
Air Blank	0.000	19:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0473																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2198																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:03																																																																																																																																																	
Control Test	0.078	19:04																																																																																																																																																	
Air Blank	0.000	19:05																																																																																																																																																	
Control Test	0.078	19:05																																																																																																																																																	
Air Blank	0.000	19:05																																																																																																																																																	
Control Test	0.079	19:07																																																																																																																																																	
Air Blank	0.000	19:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:13																																																																																																																																																	
Control Test	0.200	19:14																																																																																																																																																	
Air Blank	0.000	19:15																																																																																																																																																	
Control Test	0.199	19:15																																																																																																																																																	
Air Blank	0.000	19:16																																																																																																																																																	
Control Test	0.202	19:17																																																																																																																																																	
Air Blank	0.000	19:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:18																																																																																																																																																	
Control Test	0.081	19:19																																																																																																																																																	
Air Blank	0.000	19:19																																																																																																																																																	
Control Test	0.080	19:20																																																																																																																																																	
Air Blank	0.000	19:20																																																																																																																																																	
Control Test	0.080	19:20																																																																																																																																																	
Air Blank	0.000	19:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC

Time of Inspection: 17:44

Date of Inspection: 11/03/2020

Serial Number: 80-005248

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

DS

SR 2020.11.
04
11:25:50
-05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes RIVERA

DAVID E REYES-RIVERA

Signature and Printed Name

11/03/2020
Date