INSTRUMENT PROCESSING SHEET

FDLE Florida Department of Law Enforcement

Agency Escambia County Sheriff's Office S/N 80-006435

Date In 01/14/2020 DI Completion Date 1/16/20 ☒Ship ☐P/U ☐H/D ☐CMI ☐EE

Intake Performed By RAW

- ☑ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

☑ Case	☑ Handle
☑ Keyboard	☑ Dry Gas Shelf
☑ Feet	☑ Breath Tube
☑ Ports	☑ Screws Tight

Other Equipment/ Accessories:

☐ Power cord	☐ Printer Cable
☑ Static Bag	☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE

JAN 22 2020

Alcohol Testing Program

Quality Checks Performed By DS

- ☒ Breath Tube Screen
- ☒ Replace External O-Rings
- ☒ Instrument Set Up Verified
- ☒ R-Value 186
- ☒ Flow Verification (L/s)

Flow Column # ATP-105

32 mm	0.148	(.139 - .169)
36 mm	0.167	(.156 - .190)
53 mm	0.242	(.228 - .278)
103 mm	0.531	(.447 - .547)

☒ Barometric Pressure Check

Gauge ID # ~~10~~ 28427

☒ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5088	201905A 05-14-2021
0.080	MP5089	201905B 05-14-2021
0.200	MP5090	201904D 04-30-2021
0.080 DGS	N/A	AG916501 06-14-2021

Flow Calibration Performed By _____

Flow Column # _____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value _____
- ☐ Post Calibration Verification (L/s)

Flow Column # _____

32 mm		(.139 - .169)
36 mm		(.156 - .190)
53 mm		(.228 - .278)
103 mm		(.447 - .547)

Maintenance Performed By _____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other _____

Temperature Checks Performed By SP

- ☑ Lab Temp °C 21.2
- External Digital Therm. ID#: 300505
- ☑ 34°C +-.2 Serial #: SD1012
- ☑ 34°C +-.2 Serial #: DR1279
- ☑ 34°C +-.2 Serial #: ~~SD1013~~ MP5090 SP

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure ID# 28421

Gauge 1025 Instrument 1022

Mouth Alcohol Solution Lot # 2018-B

Acetone Stock Solution Lot # 2019-A

Simulator	Serial Number
0.000	G2408
Interferent	G2882
0.050	SD1012
0.080	DR1279
0.200	SP ~~SD1013~~ MP5090

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☐ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: _____

- ☒ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ☒ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ☒ **Conduct an Agency Inspection Before Evidentiary Use**

PJM 1/21/20 Brett Kirkland 1/21/2020

Tech Review / Date Admin Review / Date

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006435 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-006435
Owning Agency:	ESCAMBIA COUNTY SO
Calibration Date:	01/16/2020
Calibration Time:	13:15

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.005
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/16/2020
Date

Shayla Platt
SHAYLA D PLATT,
Department Inspector

BK 1/21/2020

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 13:15
Date of Inspection: 01/16/2020
Serial Number: 80-006435
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.049	0.078	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081
Standard Deviations	0.0000	0.0003	0.0003	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

RJM
TSK
1/21/2020

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
Signature and Printed Name

SHAYLA D PLATT

01/16/2020
Date

FDLE/ATP Form 41 –Revised August 2005

Stability checks

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:52
Control Test	0.049	12:53
Air Blank	0.000	12:53
Control Test	0.049	12:54
Air Blank	0.000	12:54
Control Test	0.049	12:55
Air Blank	0.000	12:56
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:57
Control Test	0.080	12:58
Air Blank	0.000	12:58
Control Test	0.080	12:59
Air Blank	0.000	12:59
Control Test	0.080	13:00
Air Blank	0.000	13:00
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:04
Control Test	0.200	13:04
Air Blank	0.000	13:05
Control Test	0.200	13:06
Air Blank	0.000	13:06
Control Test	0.201	13:07
Air Blank	0.000	13:07
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006435
01/14/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:08
Control Test	0.082	13:08
Air Blank	0.000	13:09
Control Test	0.081	13:09
Air Blank	0.000	13:10
Control Test	0.081	13:10
Air Blank	0.000	13:11
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

Dry

Operator's Signature

PJM
BK
1/21/2020