



INSTRUMENT PROCESSING SHEET

Agency Escambia County

S/N 80-006436

Florida Department of Law Enforcement

Date In 3/12/2020

DI Completion Date 3/29/20

Ship P/U H/D CMI EE

Intake Performed By <u>DP</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____		Quality Checks Performed By <u>JS</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>220</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.292</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2020.04.09 10:27:26 -04'00'		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>SD1021</u></td> <td><u>2019 05A</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.080</td> <td><u>DR1275</u></td> <td><u>2019 05B</u> <u>05-14-2021</u></td> </tr> <tr> <td>0.200</td> <td><u>SD1011</u></td> <td><u>2019 04D</u> <u>04-30-2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG931603</u> <u>11-12-2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>SD1021</u>	<u>2019 05A</u> <u>05-14-2021</u>	0.080	<u>DR1275</u>	<u>2019 05B</u> <u>05-14-2021</u>	0.200	<u>SD1011</u>	<u>2019 04D</u> <u>04-30-2021</u>	0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																								
Simulator	Serial #	Lot #/Exp																																									
0.050	<u>SD1021</u>	<u>2019 05A</u> <u>05-14-2021</u>																																									
0.080	<u>DR1275</u>	<u>2019 05B</u> <u>05-14-2021</u>																																									
0.200	<u>SD1011</u>	<u>2019 04D</u> <u>04-30-2021</u>																																									
0.080 DGS	N/A	<u>AG931603</u> <u>11-12-2021</u>																																									
Calibration Adjustment Performed By _____ Barometric Pressure Gauge ID # _____		Department Inspection Performed By <u>SP</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1017</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>MP5086</u></td> </tr> <tr> <td>Interferent</td> <td><u>MP5087</u></td> </tr> <tr> <td>0.050</td> <td><u>MP5088</u></td> </tr> <tr> <td>0.080</td> <td><u>MP5089</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5090</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>MP5086</u>	Interferent	<u>MP5087</u>	0.050	<u>MP5088</u>	0.080	<u>MP5089</u>	0.200	<u>MP5090</u>
Simulator	Serial Number	Lot Number	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	<u>MP5086</u>																																										
Interferent	<u>MP5087</u>																																										
0.050	<u>MP5088</u>																																										
0.080	<u>MP5089</u>																																										
0.200	<u>MP5090</u>																																										
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																					
Simulator	Serial Number	Lot Number	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Notes/Suggested Service: _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
_____ _____ _____ _____ _____		<table border="1"> <tr> <td><u>Michael D. Hagley</u> 2020.04.07 14:13:32 -04'00'</td> <td><u>Brett Kirkland</u> 2020.04.09 10:23:16 -04'00'</td> </tr> <tr> <td>Tech Review / Date</td> <td>Admin Review / Date</td> </tr> </table>		<u>Michael D. Hagley</u> 2020.04.07 14:13:32 -04'00'	<u>Brett Kirkland</u> 2020.04.09 10:23:16 -04'00'	Tech Review / Date	Admin Review / Date																																				
<u>Michael D. Hagley</u> 2020.04.07 14:13:32 -04'00'	<u>Brett Kirkland</u> 2020.04.09 10:23:16 -04'00'																																										
Tech Review / Date	Admin Review / Date																																										



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006436, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006436</u>	UNCERTAINTY* ±	
Owning Agency:	<u>ESCAMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/29/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>13:03</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

03/29/2020

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 January 2020

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

BK

MH

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 13:03

Date of Inspection: 03/29/2020

Serial Number: 80-006436
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.048	0.078	0.200	0.078
0.000	0.047	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.078	0.200	0.078
0.000	0.047	0.078	0.199	0.078
0.000	0.047	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.079	0.199	0.077

Standard Deviations	0.0004	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

MH

BK 2020.04.09 10:24:35 -0400

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT
Signature and Printed Name

03/29/2020
Date

Stability Checks

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/13/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:20
Control Test	0.046	14:21
Air Blank	0.000	14:21
Control Test	0.047	14:22
Air Blank	0.000	14:23
Control Test	0.046	14:23
Air Blank	0.000	14:24
Control Test Stats		
Average	0.0463	
Std Dev	0.0006	
Rel Std Dev(%)	1.2461	

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/13/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:25
Control Test	0.078	14:26
Air Blank	0.000	14:26
Control Test	0.078	14:27
Air Blank	0.000	14:28
Control Test	0.078	14:28
Air Blank	0.000	14:29
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/13/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:31
Control Test	0.200	14:31
Air Blank	0.000	14:32
Control Test	0.200	14:33
Air Blank	0.000	14:33
Control Test	0.199	14:34
Air Blank	0.000	14:34
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

wet

Operator's Signature

Operator's Signature

Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/13/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.078	14:37
Air Blank	0.000	14:38
Control Test	0.078	14:38
Air Blank	0.000	14:38
Control Test	0.079	14:39
Air Blank	0.000	14:39
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

dry

MA

Operator's Signature

stability checks

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:18
Control Test	0.048	06:19
Air Blank	0.000	06:19
Control Test	0.048	06:20
Air Blank	0.000	06:21
Control Test	0.049	06:21
Air Blank	0.000	06:22
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

MP5088



Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:23
Control Test	0.078	06:24
Air Blank	0.000	06:24
Control Test	0.079	06:25
Air Blank	0.000	06:26
Control Test	0.079	06:26
Air Blank	0.000	06:27
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

MP5089
wet



Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:35
Control Test	0.079	06:35
Air Blank	0.000	06:36
Control Test	0.078	06:36
Air Blank	0.000	06:37
Control Test	0.079	06:37
Air Blank	0.000	06:37
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DRY



Operator's Signature

ESCAMBIA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006436
 03/29/2020
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	06:29
Control Test	0.200	06:29
Air Blank	0.000	06:30
Control Test	0.200	06:31
Air Blank	0.000	06:31
Control Test	0.200	06:32
Air Blank	0.000	06:32
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MP5090



Operator's Signature

*Repeated stability
 checks after
 suspecting issue
 w/ previous
 simulators.*

MA

BK

2020.04.
 09
 10:25:36
 -04'00"