



INSTRUMENT PROCESSING SHEET

Agency Manatee County SO

S/N 80-006634

Florida Department of Law Enforcement

Date In 04/16/2020 DI Completion Date 04/16/2020

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ <b>Final Release Date</b> <div style="text-align: center;"> <p><b>FDLE</b></p> <p>APR 30 2020</p> <p>Alcohol Testing Program</p> </div>	<b>Quality Checks</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>223</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11/12/2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG931603 11/12/2021	<b>Flow Calibration</b> Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Temperature Checks</b> Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.36</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/-2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4863	201905A 05/14/2021																																																											
0.080	MP4864	201905B 05/14/2021																																																											
0.200	MP5097	201904D 04/30/2021																																																											
0.080 DGS	N/A	AG931603 11/12/2021																																																											
<b>Calibration Adjustment</b> Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1016</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD1014																																																												
Interferent	SD1015																																																												
0.050	MP4863																																																												
0.080	MP4864																																																												
0.200	MP5097																																																												
Notes/Suggested Service: <u>E-mailed</u> <div style="border: 1px solid black; padding: 5px; display: inline-block;"> <input checked="" type="checkbox"/> <b>APPROVED</b> 4/17/2020         </div>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: center; margin-top: 10px;"> <p><i>SRP 4/27/20 Brett Kirkland 4/28/2020</i></p> </div>																																																												
	Tech Review / Date _____ Admin Review / Date _____																																																												

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 12:37

Date of Inspection: 04/16/2020

Serial Number: 80-006634  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080

Standard Deviations	0.0003	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Michael D Haughey*

MICHAEL D HAUGHEY

Signature and Printed Name

04/16/2020  
Date

*SP BK 4/28/2020*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006634	Manatee County SO	04/16/2020	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>


<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006634 04/16/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:41</p> <p>Control Test 0.050 09:42</p> <p>Air Blank 0.000 09:42</p> <p>Control Test 0.049 09:43</p> <p>Air Blank 0.000 09:43</p> <p>Control Test 0.050 09:44</p> <p>Air Blank 0.000 09:45</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006634 04/16/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:47</p> <p>Control Test 0.080 09:48</p> <p>Air Blank 0.000 09:48</p> <p>Control Test 0.080 09:49</p> <p>Air Blank 0.000 09:49</p> <p>Control Test 0.080 09:50</p> <p>Air Blank 0.000 09:51</p> <p>Control Test Stats</p> <p>Average 0.0800</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006634 04/16/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:53</p> <p>Control Test 0.201 09:54</p> <p>Air Blank 0.000 09:54</p> <p>Control Test 0.200 09:55</p> <p>Air Blank 0.000 09:55</p> <p>Control Test 0.199 09:56</p> <p>Air Blank 0.000 09:57</p> <p>Control Test Stats</p> <p>Average 0.2000</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 0.5000</p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006634 04/16/2020 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:00</p> <p>Control Test 0.080 10:00</p> <p>Air Blank 0.000 10:01</p> <p>Control Test 0.080 10:01</p> <p>Air Blank 0.000 10:01</p> <p>Control Test 0.080 10:02</p> <p>Air Blank 0.000 10:02</p> <p>Control Test Stats</p> <p>Average 0.0800</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>
---	---	---	---

SP  
BSK  
4/28/2020

MX  
Operator's Signature

MX  
Operator's Signature

MX  
Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006634, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006634</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/16/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>12:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2020  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

04/16/2020 Date   
MICHAEL D HAUGHEY,  
Department Inspector

*SP BK 4/28/2020*