



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop LS/N 80-006762Florida Department of
Law EnforcementDate In 10/27/2020 DI Completion Date 11/4/2020☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Tank holder was off</u> <u>Lose screws on dry gas</u> <u>regulator</u> Final Release Date FDLE Alcohol Testing Program		Quality Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>229</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 1/30/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	SD3967	201905A 05/14/2021	0.080	SD3968	201905B 05/14/2021	0.200	SD3969	201904D 04/30/2021	0.080 DGS	N/A	AG003005 1/30/2022	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By <u>DERR</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other Attached tank holder/ Temperature Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.87C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>	
Simulator	Serial #	Lot #/Exp																		
0.050	SD3967	201905A 05/14/2021																		
0.080	SD3968	201905B 05/14/2021																		
0.200	SD3969	201904D 04/30/2021																		
0.080 DGS	N/A	AG003005 1/30/2022																		

Calibration Adjustment Performed By <u>DERR</u> Barometric Pressure Gauge <u>1019</u> ID # <u>68639</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5095</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>20060</td> <td>02/10/2022</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>20190</td> <td>04/06/2022</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>20160</td> <td>03/18/2022</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>20030</td> <td>01/21/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08819080A1</td> <td>06/05/2021</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3967</td> <td>201905A</td> <td>05/14/2021</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> <td>201905B</td> <td>05/14/2021</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> <td>201904D</td> <td>04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005</td> <td>01/30/2022</td> </tr> </tbody> </table> Notes/Suggested Service: <u>Optical calibration to bring</u> <u>values closer to nominal. Continued from Mainte-</u> <u>nance section: Secured lose dry gas regulator.</u> <u>Required a second optical calibration to bring</u> <u>values closer to nominal.</u>				Simulator	Serial Number	Lot Number	Expiration	0.000	MP5095	N/A	N/A	0.040	MP5098	20060	02/10/2022	0.100	MP5099	20190	04/06/2022	0.200	MP5100	20160	03/18/2022	0.300	MP5101	20030	01/21/2022	0.080 DGS	N/A	08819080A1	06/05/2021	Simulator	Serial Number	Lot Number	Expiration	0.050	SD3967	201905A	05/14/2021	0.080	SD3968	201905B	05/14/2021	0.200	SD3969	201904D	04/30/2021	0.080 DGS	N/A	AG003005	01/30/2022	Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3965</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>SD3967</td> </tr> <tr> <td>0.080</td> <td>SD3968</td> </tr> <tr> <td>0.200</td> <td>SD3969</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before <u>2020-11-05</u> <u>Doral Soto</u> 11-04-2020 <u>09:58:4</u> Tech Review / Date Admin Review / Date				Simulator	Serial Number	0.000	SD3965	Interferent	SD3966	0.050	SD3967	0.080	SD3968	0.200	SD3969
Simulator	Serial Number	Lot Number	Expiration																																																																
0.000	MP5095	N/A	N/A																																																																
0.040	MP5098	20060	02/10/2022																																																																
0.100	MP5099	20190	04/06/2022																																																																
0.200	MP5100	20160	03/18/2022																																																																
0.300	MP5101	20030	01/21/2022																																																																
0.080 DGS	N/A	08819080A1	06/05/2021																																																																
Simulator	Serial Number	Lot Number	Expiration																																																																
0.050	SD3967	201905A	05/14/2021																																																																
0.080	SD3968	201905B	05/14/2021																																																																
0.200	SD3969	201904D	04/30/2021																																																																
0.080 DGS	N/A	AG003005	01/30/2022																																																																
Simulator	Serial Number																																																																		
0.000	SD3965																																																																		
Interferent	SD3966																																																																		
0.050	SD3967																																																																		
0.080	SD3968																																																																		
0.200	SD3969																																																																		



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006762, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006762</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FHP TROOP L</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/04/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>00:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

2020.11
05
09:57:0
7 -05'00'

11/04/2020 Date
David Reyna Lopez Rivera
DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L

Serial Number: 80-006762

Time of Inspection: 00:28

Date of Inspection: 11/04/2020

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.047	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.047	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080

Standard Deviations	0.0004	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

DS

CR

2020.11.05
09:56:15 -05'00'

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

11/04/2020
Date

0.28

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities 2	80-006762	Florida Highway Patrol Troop L	11/3/2020	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-006762</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>20:45</td></tr><tr><td>Control Test</td><td>0.047</td><td>20:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:46</td></tr><tr><td>Control Test</td><td>0.048</td><td>20:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:47</td></tr><tr><td>Control Test</td><td>0.048</td><td>20:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:48</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0477</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	20:45	Control Test	0.047	20:45	Air Blank	0.000	20:46	Control Test	0.048	20:47	Air Blank	0.000	20:47	Control Test	0.048	20:48	Air Blank	0.000	20:48	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-006762</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>20:50</td></tr><tr><td>Control Test</td><td>0.077</td><td>20:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:51</td></tr><tr><td>Control Test</td><td>0.078</td><td>20:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:52</td></tr><tr><td>Control Test</td><td>0.079</td><td>20:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	20:50	Control Test	0.077	20:50	Air Blank	0.000	20:51	Control Test	0.078	20:52	Air Blank	0.000	20:52	Control Test	0.079	20:53	Air Blank	0.000	20:53	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-006762</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>20:55</td></tr><tr><td>Control Test</td><td>0.197</td><td>20:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:56</td></tr><tr><td>Control Test</td><td>0.198</td><td>20:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:57</td></tr><tr><td>Control Test</td><td>0.197</td><td>20:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>20:58</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1973</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	20:55	Control Test	0.197	20:55	Air Blank	0.000	20:56	Control Test	0.198	20:57	Air Blank	0.000	20:57	Control Test	0.197	20:58	Air Blank	0.000	20:58	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27 SN 80-006762</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>21:00</td></tr><tr><td>Control Test</td><td>0.080</td><td>21:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:00</td></tr><tr><td>Control Test</td><td>0.080</td><td>21:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:01</td></tr><tr><td>Control Test</td><td>0.080</td><td>21:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>21:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	21:00	Control Test	0.080	21:00	Air Blank	0.000	21:00	Control Test	0.080	21:01	Air Blank	0.000	21:01	Control Test	0.080	21:02	Air Blank	0.000	21:02	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	20:45																																																																																																																																																	
Control Test	0.047	20:45																																																																																																																																																	
Air Blank	0.000	20:46																																																																																																																																																	
Control Test	0.048	20:47																																																																																																																																																	
Air Blank	0.000	20:47																																																																																																																																																	
Control Test	0.048	20:48																																																																																																																																																	
Air Blank	0.000	20:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	20:50																																																																																																																																																	
Control Test	0.077	20:50																																																																																																																																																	
Air Blank	0.000	20:51																																																																																																																																																	
Control Test	0.078	20:52																																																																																																																																																	
Air Blank	0.000	20:52																																																																																																																																																	
Control Test	0.079	20:53																																																																																																																																																	
Air Blank	0.000	20:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2821																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	20:55																																																																																																																																																	
Control Test	0.197	20:55																																																																																																																																																	
Air Blank	0.000	20:56																																																																																																																																																	
Control Test	0.198	20:57																																																																																																																																																	
Air Blank	0.000	20:57																																																																																																																																																	
Control Test	0.197	20:58																																																																																																																																																	
Air Blank	0.000	20:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	21:00																																																																																																																																																	
Control Test	0.080	21:00																																																																																																																																																	
Air Blank	0.000	21:00																																																																																																																																																	
Control Test	0.080	21:01																																																																																																																																																	
Air Blank	0.000	21:01																																																																																																																																																	
Control Test	0.080	21:02																																																																																																																																																	
Air Blank	0.000	21:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5690 (0.0010)
Sample #2 = 1.5500 (0.0000)
Sample #3 = 1.5440 (0.0110)
Sample #4 = 1.5450 (0.0020)
Avg % Abs = 1.5463 (0.0043)
STD DEV = 0.0032 (0.0059)
REL STD DEV = 0.208 (135.218)
FHP Troop L
Intoxilyzer - Alcohol Analyzer
SN 80-006762
Model 8000
11/03/2020 19:47:04

Auto Calibration
Max Power Res Value = 47
Auto Range Res Value = 34

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12683, Sum Io = 13244
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1050 (-0.0140)
Sample #2 = 0.1100 (-0.0320)
Sample #3 = 0.0810 (-0.0090)
Sample #4 = 0.1100 (0.0000)
Avg % Abs = 0.1003 (-0.0137)
STD DEV = 0.0167 (0.0165)
REL STD DEV = 16.688 (120.750)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (-0.0060)
Sample #2 = 0.1270 (-0.0120)
Sample #3 = 0.1260 (-0.0040)
Sample #4 = 0.1390 (-0.0020)
Avg % Abs = 0.1307 (-0.0060)
STD DEV = 0.0072 (0.0053)
REL STD DEV = 5.536 (68.192)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12683, Sum Io = 13245
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8680 (-0.0150)
Sample #2 = 0.8230 (0.0120)
Sample #3 = 0.8250 (0.0100)
Sample #4 = 0.8210 (0.0220)
Avg % Abs = 0.8230 (0.0147)
STD DEV = 0.0020 (0.0064)
REL STD DEV = 0.243 (43.635)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8440 (-0.0030)
Sample #2 = 6.8370 (0.0040)
Sample #3 = 6.8400 (0.0090)
Sample #4 = 6.8370 (0.0120)
Avg % Abs = 6.8380 (0.0083)
STD DEV = 0.0017 (0.0040)
REL STD DEV = 0.025 (48.497)
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12674, Sum Io = 13241
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3440 (-0.0030)
Sample #2 = 5.3510 (-0.0100)
Sample #3 = 5.3290 (0.0120)
Sample #4 = 5.3520 (-0.0060)
Avg % Abs = 5.3440 (-0.0013)
STD DEV = 0.0130 (0.0117)
REL STD DEV = 0.243 (878.920)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.8290 (-0.0100)
Sample #2 = 9.8170 (0.0080)
Sample #3 = 9.8110 (0.0140)
Sample #4 = 9.8140 (-0.0020)
Avg % Abs = 9.8140 (0.0067)
STD DEV = 0.0030 (0.0081)
REL STD DEV = 0.031 (121.244)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12674, Sum Io = 13242
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6760 (-0.0080)
Sample #2 = 3.7230 (-0.0310)
Sample #3 = 3.7240 (-0.0240)
Sample #4 = 3.7060 (0.0020)
Avg % Abs = 3.7177 (-0.0177)
STD DEV = 0.0101 (0.0174)
REL STD DEV = 0.272 (98.421)

Optical Calibration 2	
SN:	80-006762
Agency:	FHP Troop L
Date:	11/3 /2020
Quadratic Fit:	+/- 0.002g/210L
By:	DERR

Auto CAL DATA *****
Channel 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.100
Std Dev = 0.02 Rel Std Dev = 16.69
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.823
Std Dev = 0.00 Rel Std Dev = 0.24
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.950
Std Dev = 0.00 Rel Std Dev = 0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.718
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.344
Std Dev = 0.01 Rel Std Dev = 0.24
Zero Order Coef = -201.10
First Order Coef = 2444.33
Second Order Coef = 49.23
Standard Deviation = 44.518257

Channel 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.131
Std Dev = 0.01 Rel Std Dev = 5.54
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.546
Std Dev = 0.00 Rel Std Dev = 0.21
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.605
Std Dev = 0.01 Rel Std Dev = 0.17
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.838
Std Dev = 0.00 Rel Std Dev = 0.03
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.814
Std Dev = 0.00 Rel Std Dev = 0.03
Zero Order Coef = -158.00
First Order Coef = 1298.43
Second Order Coef = 17.58
Standard Deviation = 14.469942

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0019
0.040 0.039 0.0013
0.100 0.100 0.0002
0.200 0.201 -0.0019
0.300 0.300 0.0004

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0003
0.100 0.100 0.0002
0.200 0.200 -0.0004
0.300 0.300 0.0002

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
Channel 1
Sample #1 = 3094.00
Sample #2 = 3084.00
Sample #3 = 3043.00
Sample #4 = 3065.00
Average Result = 3064.0000
STD DEV = 20.5183
REL STD DEV = 0.670

Channel 2
Sample #1 = 3274.00
Sample #2 = 3258.00
Sample #3 = 3263.00
Sample #4 = 3272.00
Average Result = 3264.3333
STD DEV = 7.0946
REL STD DEV = 0.217
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1020
3 um H2O Adjust (mg/l*10,000) = 745
9 um H2O Adjust (mg/l*10,000) = 545
Auto CAL PASS

02

2020.11.05
09:54:56
-05'00"

2020.11.0

5

09:54:26

02

09:54:26

Type of Test	Serial Number	Agency	Date	Time	Performed By
Post Stabilities	80-006762	Florida Highway Patrol Troop L	11/3/2020	-05'00"	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27</div> <div>Test g/210L Time Air Blank 0.000 19:22 Control Test 0.046 19:23 Air Blank 0.000 19:24 Control Test 0.046 19:24 Air Blank 0.000 19:25 Control Test 0.047 19:26 Air Blank 0.000 19:26 Control Test Status Average 0.0463 Std Dev 0.0106 Rel Std Dev(%) 1.2461</div>	<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27</div> <div>Test g/210L Time Air Blank 0.000 19:28 Control Test 0.077 19:29 Air Blank 0.000 19:30 Control Test 0.077 19:30 Air Blank 0.000 19:31 Control Test 0.077 19:31 Air Blank 0.000 19:32 Control Test Status Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000</div>	<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27</div> <div>Test g/210L Time Air Blank 0.000 19:34 Control Test 0.196 19:35 Air Blank 0.000 19:35 Control Test 0.196 19:36 Air Blank 0.000 19:36 Control Test 0.197 19:37 Air Blank 0.000 19:37 Control Test Status Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.2941</div>	<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 11/03/2020 Software: 8100.27</div> <div>Test g/210L Time Air Blank 0.000 19:39 Control Test 0.080 19:39 Air Blank 0.000 19:39 Control Test 0.080 19:40 Air Blank 0.000 19:40 Control Test 0.079 19:41 Air Blank 0.000 19:41 Control Test Status Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</div>
<div>Operator's Signature </div>	<div>Operator's Signature </div>	<div>Operator's Signature </div>	<div>Operator's Signature </div>

2020.11.0
5 09:53:25
-05'00"

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006762	Florida Highway Patrol Troop L	10/27/2020	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2020 Software: 8100.27</p> <p>SN 80-006762</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:27 Control Test 0.045 14:28 Air Blank 0.000 14:29 Control Test 0.046 14:29 Air Blank 0.000 14:30 Control Test 0.046 14:30 Air Blank 0.000 14:31 Control Test Stats Average 0.0457 Std Dev 0.0006 Rel Std Dev(%) 1.2643</p>	<p>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2020 Software: 8100.27</p> <p>SN 80-006762</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:32 Control Test 0.075 14:33 Air Blank 0.000 14:33 Control Test 0.076 14:34 Air Blank 0.000 14:34 Control Test 0.077 14:35 Air Blank 0.000 14:36 Control Test Stats Average 0.0760 Std Dev 0.0010 Rel Std Dev(%) 1.3158</p>	<p>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2020 Software: 8100.27</p> <p>SN 80-006762</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:36 Control Test 0.193 14:37 Air Blank 0.000 14:38 Control Test 0.193 14:38 Air Blank 0.000 14:39 Control Test 0.194 14:40 Air Blank 0.000 14:40 Control Test Stats Average 0.1933 Std Dev 0.0006 Rel Std Dev(%) 0.2986</p>	<p>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 10/27/2020 Software: 8100.27</p> <p>SN 80-006762</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:42 Control Test 0.078 14:42 Air Blank 0.000 14:43 Control Test 0.079 14:43 Air Blank 0.000 14:44 Control Test 0.078 14:44 Air Blank 0.000 14:44 Control Test Stats Average 0.0783 Std Dev 0.0006 Rel Std Dev(%) 0.7370</p>