



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade Police Department s/N 80-007084Florida Department of Law Enforcement Date In 08/03/2020 DI Completion Date 08/05/2020 Ship P/U H/D CMI EE

Intake Performed By <u>DERR</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____	Quality Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>173</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
---	--	--

Final Release Date

FDLE
Alcohol
Testing
Program

Digitally signed by FDLE Alcohol Testing Program
Date: 2020.08.07 11:56:29 -04'00'

Simulator	Serial #	Lot #/Exp
0.050	SD3967	201905A 05/14/2021
0.080	SD3968	201905B 05/14/2021
0.200	SD3969	201904D 04/30/2021
0.080 DGS	N/A	AG003005 1/30/2022

Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	Temperature Checks Performed By <u>DERR</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.89C</u> External Digital Therm. ID#: <u>300918</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3967</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3968</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>SD3969</u>
---	--

Calibration Adjustment Performed By <u>DERR</u> Barometric Pressure Gauge <u>1015</u> ID # <u>68639</u>	Department Inspection Performed By <u>DERR</u> Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2019B</u> Acetone Stock Solution Lot # <u>2019A</u>
---	---

Simulator	Serial Number	Lot Number	Expiration
0.000	MP5095	N/A	N/A
0.040	MP5098	20060	02/10/2022
0.100	MP5099	20190	04/06/2022
0.200	MP5100	20160	03/18/2022
0.300	MP5101	20030	01/21/2022
0.080 DGS	N/A	08819080A1	06/05/2021

Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050	SD3967	201905A	05/14/2021
0.080	SD3968	201905B	05/14/2021
0.200	SD3969	201904D	04/30/2021
0.080 DGS	N/A	AG003005	01/30/2022

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment	<input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
--	--

Notes/Suggested Service: Optical calibration to bring values closer to nominal.

Technical review correction; Optical calibration date DERR 08/05/2020

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

2020.08.06 08:57:30 -04'00' Michael D. Haughey 2020.08.07 11:50:18 -04'00' Brett Kirkland

Tech Review / Date _____ Admin Review / Date _____

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 11:10

Date of Inspection: 08/05/2020

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.047	0.077	0.198	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.077	0.198	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.047	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079

Standard Deviations	0.0005	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

MX
BK 2020.08.07
11:51:17 -04'00"

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera

Signature and Printed Name

08/05/2020
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-007084	Miami-Dade Police Department	8/5/2020	<i>AKK</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/05/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:16</td></tr> <tr><td>Control Test</td><td>0.047</td><td>09:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:17</td></tr> <tr><td>Control Test</td><td>0.047</td><td>09:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:18</td></tr> <tr><td>Control Test</td><td>0.047</td><td>09:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0470</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>AKK</i></p>	Test	g/210L	Time	Air Blank	0.000	09:16	Control Test	0.047	09:17	Air Blank	0.000	09:17	Control Test	0.047	09:18	Air Blank	0.000	09:18	Control Test	0.047	09:19	Air Blank	0.000	09:20	Control Test Stats			Average	0.0470		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/05/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:21</td></tr> <tr><td>Control Test</td><td>0.077</td><td>09:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:24</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>AKK</i></p>	Test	g/210L	Time	Air Blank	0.000	09:21	Control Test	0.077	09:22	Air Blank	0.000	09:22	Control Test	0.079	09:23	Air Blank	0.000	09:24	Control Test	0.078	09:24	Air Blank	0.000	09:25	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/05/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:26</td></tr> <tr><td>Control Test</td><td>0.197</td><td>09:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:27</td></tr> <tr><td>Control Test</td><td>0.196</td><td>09:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5076</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>AKK</i></p>	Test	g/210L	Time	Air Blank	0.000	09:26	Control Test	0.197	09:26	Air Blank	0.000	09:27	Control Test	0.196	09:28	Air Blank	0.000	09:28	Control Test	0.198	09:29	Air Blank	0.000	09:30	Control Test Stats			Average	0.1970		Std Dev	0.0010		Rel Std Dev(%)	0.5076		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007084 08/05/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>AKK</i></p>	Test	g/210L	Time	Air Blank	0.000	09:31	Control Test	0.080	09:31	Air Blank	0.000	09:32	Control Test	0.080	09:32	Air Blank	0.000	09:32	Control Test	0.079	09:33	Air Blank	0.000	09:33	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:16																																																																																																																																																	
Control Test	0.047	09:17																																																																																																																																																	
Air Blank	0.000	09:17																																																																																																																																																	
Control Test	0.047	09:18																																																																																																																																																	
Air Blank	0.000	09:18																																																																																																																																																	
Control Test	0.047	09:19																																																																																																																																																	
Air Blank	0.000	09:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0470																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:21																																																																																																																																																	
Control Test	0.077	09:22																																																																																																																																																	
Air Blank	0.000	09:22																																																																																																																																																	
Control Test	0.079	09:23																																																																																																																																																	
Air Blank	0.000	09:24																																																																																																																																																	
Control Test	0.078	09:24																																																																																																																																																	
Air Blank	0.000	09:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2821																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:26																																																																																																																																																	
Control Test	0.197	09:26																																																																																																																																																	
Air Blank	0.000	09:27																																																																																																																																																	
Control Test	0.196	09:28																																																																																																																																																	
Air Blank	0.000	09:28																																																																																																																																																	
Control Test	0.198	09:29																																																																																																																																																	
Air Blank	0.000	09:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5076																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:31																																																																																																																																																	
Control Test	0.080	09:31																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test	0.080	09:32																																																																																																																																																	
Air Blank	0.000	09:32																																																																																																																																																	
Control Test	0.079	09:33																																																																																																																																																	
Air Blank	0.000	09:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		

2020.08.07
11:52:07
-04'00'
AKK BK

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
08/05/2020
SN 80-07084
08:20:31

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9630 (0.0150)
Sample #2 = 6.9910 (-0.0050)
Sample #3 = 7.0000 (0.0140)
Sample #4 = 7.0010 (0.0160)
Avg % Abs = 6.9973 (0.0083)
STD DEV = 0.0055 (0.0116)
REL STD DEV = 0.079 (139.083)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9630 (0.0150)
Sample #2 = 6.9910 (-0.0050)
Sample #3 = 7.0000 (0.0140)
Sample #4 = 7.0010 (0.0160)
Avg % Abs = 6.9973 (0.0083)
STD DEV = 0.0055 (0.0116)
REL STD DEV = 0.079 (139.083)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5250 (0.0200)
Sample #2 = 1.5500 (0.0130)
Sample #3 = 1.5380 (0.0230)
Sample #4 = 1.5850 (0.0240)
Avg % Abs = 1.5577 (0.0200)
STD DEV = 0.0244 (0.0061)
REL STD DEV = 1.568 (30.414)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5250 (0.0200)
Sample #2 = 1.5500 (0.0130)
Sample #3 = 1.5380 (0.0230)
Sample #4 = 1.5850 (0.0240)
Avg % Abs = 1.5577 (0.0200)
STD DEV = 0.0244 (0.0061)
REL STD DEV = 1.568 (30.414)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5250 (0.0200)
Sample #2 = 1.5500 (0.0130)
Sample #3 = 1.5380 (0.0230)
Sample #4 = 1.5850 (0.0240)
Avg % Abs = 1.5577 (0.0200)
STD DEV = 0.0244 (0.0061)
REL STD DEV = 1.568 (30.414)

Auto Calibration
Max Power Res Value = 91
Auto Range Res Value = 69
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12667, Sum Io = 13081

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1470 (-0.0050)
Sample #2 = 0.1590 (0.0540)
Sample #3 = 0.1390 (0.0820)
Sample #4 = 0.1930 (0.0830)
Avg % Abs = 0.1637 (0.0730)
STD DEV = 0.0273 (0.0165)
REL STD DEV = 16.681 (22.551)

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0190 (0.0000)
Sample #2 = 2.0480 (0.0110)
Sample #3 = 1.9930 (0.0550)
Sample #4 = 2.0350 (0.0640)
Avg % Abs = 2.0253 (0.0433)
STD DEV = 0.0287 (0.0284)
REL STD DEV = 1.419 (65.448)

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0190 (0.0000)
Sample #2 = 2.0480 (0.0110)
Sample #3 = 1.9930 (0.0550)
Sample #4 = 2.0350 (0.0640)
Avg % Abs = 2.0253 (0.0433)
STD DEV = 0.0287 (0.0284)
REL STD DEV = 1.419 (65.448)

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0190 (0.0000)
Sample #2 = 2.0480 (0.0110)
Sample #3 = 1.9930 (0.0550)
Sample #4 = 2.0350 (0.0640)
Avg % Abs = 2.0253 (0.0433)
STD DEV = 0.0287 (0.0284)
REL STD DEV = 1.419 (65.448)

Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 2.0190 (0.0000)
Sample #2 = 2.0480 (0.0110)
Sample #3 = 1.9930 (0.0550)
Sample #4 = 2.0350 (0.0640)
Avg % Abs = 2.0253 (0.0433)
STD DEV = 0.0287 (0.0284)
REL STD DEV = 1.419 (65.448)

Solution Status Quadratic Fit Chan 2
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.001 -0.0006
0.040 0.039 0.0006
0.100 0.099 0.0006
0.200 0.201 -0.0010
0.300 0.300 0.0004
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2693.00
Sample #2 = 2690.00
Sample #3 = 2718.00
Sample #4 = 2695.00
Average Result = 2701.0000
STD DEV = 14.9332
REL STD DEV = 0.553

***** CHANNEL 2 *****
Sample #1 = 3326.00
Sample #2 = 3335.00
Sample #3 = 3363.00
Sample #4 = 3300.00
Average Result = 3332.6667
STD DEV = 31.5647
REL STD DEV = 0.947

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1015
3 um H2O Adjust (mg/l x 10,000) = 1108
9 um H2O Adjust (mg/l x 10,000) = 477
***** AUTO CAL PASS *****

Channel 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.164
Std Dev = 0.03 Rel Std Dev = 16.68
Sol Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 0.923
Std Dev = 0.02 Rel Std Dev = 2.12
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 2.025
Std Dev = 0.03 Rel Std Dev = 1.42
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.854
Std Dev = 0.02 Rel Std Dev = 0.39
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.551
Std Dev = 0.00 Rel Std Dev = 0.06
Zero Order Coef = -388.31
First Order Coef = 2453.58
Second Order Coef = 33.72
Standard Deviation = 32.722984
***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.148
Std Dev = 0.03 Rel Std Dev = 21.44
Sol Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 1.558
Std Dev = 0.02 Rel Std Dev = 1.57
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.655
Std Dev = 0.01 Rel Std Dev = 0.34
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.997
Std Dev = 0.01 Rel Std Dev = 0.08
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.029
Std Dev = 0.01 Rel Std Dev = 0.13
Zero Order Coef = -159.39
First Order Coef = 1280.95
Second Order Coef = 15.70
Standard Deviation = 37.004757

Channel 2 >>>>
Sol Val = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12605, Sum Io = 13046
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.5530 (0.0000)
Sample #2 = 5.5520 (0.0090)
Sample #3 = 5.5470 (0.0220)
Sample #4 = 5.5530 (0.0320)
Avg % Abs = 5.5507 (0.0210)
STD DEV = 0.0032 (0.0115)
REL STD DEV = 0.058 (54.917)

Channel 2 >>>>
Sol Val = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12614, Sum Io = 13054
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.8310 (-0.0100)
Sample #2 = 3.8400 (-0.0030)
Sample #3 = 3.8700 (0.0060)
Sample #4 = 3.8520 (0.0240)
Avg % Abs = 3.8540 (0.0090)
STD DEV = 0.0151 (0.0137)
REL STD DEV = 0.392 (152.753)

Channel 2 >>>>
Sol Val = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12614, Sum Io = 13054
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.8310 (-0.0100)
Sample #2 = 3.8400 (-0.0030)
Sample #3 = 3.8700 (0.0060)
Sample #4 = 3.8520 (0.0240)
Avg % Abs = 3.8540 (0.0090)
STD DEV = 0.0151 (0.0137)
REL STD DEV = 0.392 (152.753)

Channel 2 >>>>
Sol Val = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12614, Sum Io = 13054
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.8310 (-0.0100)
Sample #2 = 3.8400 (-0.0030)
Sample #3 = 3.8700 (0.0060)
Sample #4 = 3.8520 (0.0240)
Avg % Abs = 3.8540 (0.0090)
STD DEV = 0.0151 (0.0137)
REL STD DEV = 0.392 (152.753)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Channel 2 >>>>
Sol Val = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12637, Sum Io = 13066
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9100 (-0.0160)
Sample #2 = 0.9010 (0.0240)
Sample #3 = 0.9390 (0.0180)
Sample #4 = 0.9280 (0.0390)
Avg % Abs = 0.9227 (0.0270)
STD DEV = 0.0196 (0.0108)

Optical Calibration
SN: 80-007084
Agency: Miami-Dade PD
Date: 08/05/2021- DERR 8/5/20
Quadratic Fit: +/- 0.002g/210L
By: DERR

Solution Status Quadratic Fit Chan 1
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0000
0.100 0.099 0.0009
0.200 0.201 -0.0009
0.300 0.300 0.0003

Solution Status Quadratic Fit Chan 1
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0000
0.100 0.099 0.0009
0.200 0.201 -0.0009
0.300 0.300 0.0003

Solution Status Quadratic Fit Chan 1
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0000
0.100 0.099 0.0009
0.200 0.201 -0.0009
0.300 0.300 0.0003

Solution Status Quadratic Fit Chan 1
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0000
0.100 0.099 0.0009
0.200 0.201 -0.0009
0.300 0.300 0.0003

Solution Status Quadratic Fit Chan 1
Rct Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0003
0.040 0.040 0.0000
0.100 0.099 0.0009
0.200 0.201 -0.0009
0.300 0.300 0.0003

MAX

2020.08.07
11:52:53
-04:00
BK

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007084	Miami-Dade Police Department	08/03/2020	<i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/03/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.046</td><td>11:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.047</td><td>11:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0470</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.1277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:10	Control Test	0.048	11:10	Air Blank	0.000	11:11	Control Test	0.046	11:12	Air Blank	0.000	11:12	Control Test	0.047	11:13	Air Blank	0.000	11:13	Control Test Stats			Average	0.0470		Std Dev	0.0010		Rel Std Dev(%)	2.1277		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/03/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:16	Control Test	0.078	11:17	Air Blank	0.000	11:18	Control Test	0.077	11:18	Air Blank	0.000	11:19	Control Test	0.078	11:20	Air Blank	0.000	11:20	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/03/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:22	Control Test	0.198	11:23	Air Blank	0.000	11:24	Control Test	0.198	11:24	Air Blank	0.000	11:25	Control Test	0.198	11:25	Air Blank	0.000	11:26	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 08/03/2020 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:28	Control Test	0.079	11:28	Air Blank	0.000	11:29	Control Test	0.079	11:29	Air Blank	0.000	11:30	Control Test	0.078	11:30	Air Blank	0.000	11:31	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:10																																																																																																																																																	
Control Test	0.048	11:10																																																																																																																																																	
Air Blank	0.000	11:11																																																																																																																																																	
Control Test	0.046	11:12																																																																																																																																																	
Air Blank	0.000	11:12																																																																																																																																																	
Control Test	0.047	11:13																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0470																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	2.1277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.078	11:17																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.077	11:18																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.078	11:20																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test	0.198	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.198	11:24																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test	0.198	11:25																																																																																																																																																	
Air Blank	0.000	11:26																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.079	11:28																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.079	11:29																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test	0.078	11:30																																																																																																																																																	
Air Blank	0.000	11:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature	<i>[Signature]</i> Operator's Signature																																																																																																																																																

MAX
BK
2020.08.07
11:54:08
-04'00'



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007084, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007084</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/05/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:10</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full,
without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/05/2020 Date
DAVID E REYES-RIVERA, Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

MX



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade PD

S/N 80-007084

Florida Department of Law Enforcement

Date In 06/03/2020

DI Completion Date 06/04/2020

Ship P/U H/D CMI EE

Intake Performed By <u>MH</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>179</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
Final Release Date <div style="text-align: center;"> <p>FDLE</p> <p>JUN 29 2020</p> <p>Alcohol Testing Program</p> </div>		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP4863</td> <td>201905A 05/14/2021</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> <td>201905B 05/14/2021</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> <td>201904D 04/30/2021</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG003005 01/30/2022</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP4863	201905A 05/14/2021	0.080	MP4864	201905B 05/14/2021	0.200	MP5097	201904D 04/30/2021	0.080 DGS	N/A	AG003005 01/30/2022	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																		
0.050	MP4863	201905A 05/14/2021																		
0.080	MP4864	201905B 05/14/2021																		
0.200	MP5097	201904D 04/30/2021																		
0.080 DGS	N/A	AG003005 01/30/2022																		
		Temperature Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.17</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5097</u>																		

Calibration Adjustment Performed By _____ ID # _____ Barometric Pressure Gauge _____ ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>68639</u> Gauge <u>1014</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-A</u>	
Simulator	Serial Number
0.000	SD1014
Interferent	SD1015
0.050	MP4863
0.080	MP4864
0.200	MP5097
Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Diagnostic</u>	

Notes/Suggested Service: E-mailed
Agency reported failed DSP test.
Ran diagnostic test along with having instrument
come into ready mode 3 times. No issues.
Passed 6 out of 6 diagnostic tests.

APPROVED 6/4/20

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

SR 6/25/20 Brett Kirkland 6/29/2020
 Tech Review / Date Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 11:34

Date of Inspection: 06/04/2020

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.078	0.200	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.078
Standard Deviations	0.0005	0.0005	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

06/04/2020
Date

*SP BK
6/29/2020*

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-007684	Miami-Dade PD	06/04/2020	MX

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

VIS

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
06/04/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.049	09:51
Air Blank	0.000	09:52
Control Test	0.049	09:53
Air Blank	0.000	09:53
Control Test	0.050	09:54
Air Blank	0.000	09:54
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
06/04/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.079	09:57
Air Blank	0.000	09:58
Control Test	0.078	09:59
Air Blank	0.000	09:59
Control Test	0.079	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
06/04/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:01
Control Test	0.200	10:02
Air Blank	0.000	10:03
Control Test	0.198	10:03
Air Blank	0.000	10:04
Control Test	0.199	10:05
Air Blank	0.000	10:05
Control Test Stats		
Average	0.1990	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
06/04/2020
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:06
Control Test	0.078	10:06
Air Blank	0.000	10:07
Control Test	0.079	10:07
Air Blank	0.000	10:07
Control Test	0.079	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	


SP BK 6/29/2020

MX
Operator's Signature

MX
Operator's Signature

MX
Operator's Signature

MX
Operator's Signature



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007084, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007084</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/04/2020</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:34</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/04/2020 Date

 MICHAEL D HAUGHEY,
 Department Inspector

FDLE/ATP Form 69 April 2020
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

SP
 BK 6/29/2020

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 09:07

Date of Inspection: 06/04/2020

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
COMPLIANCE NOT DETERMINED. AI NOT CONDUCTED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D HAUGHEY

Signature and Printed Name

06/04/2020
Date

SP
13K
6/29/2020

MIAMI-DADE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007084
06/04/2020
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

MX

SP

BR
6/29/2020



INSTRUMENT PROCESSING SHEET

Agency Miami Dade PD

S/N 80-007084

Florida Department of Law Enforcement

Date In 02/05/2020 DI Completion Date 02/06/2020

Ship P/U H/D CMI EE

Intake Performed By <u>JMB</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____		Quality Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>171</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.257</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks		Flow Calibration Performed By _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																											
Final Release Date FDLE FEB 24 2020 Alcohol Testing Program		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td><u>MP4863</u></td> <td><u>201905A</u> <u>05/14/2021</u></td> </tr> <tr> <td>0.080</td> <td><u>MP4864</u></td> <td><u>201905B</u> <u>05/14/2021</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5097</u></td> <td><u>201904D</u> <u>04/30/2021</u></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>AG916501</u> <u>06/14/2021</u></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	<u>MP4863</u>	<u>201905A</u> <u>05/14/2021</u>	0.080	<u>MP4864</u>	<u>201905B</u> <u>05/14/2021</u>	0.200	<u>MP5097</u>	<u>201904D</u> <u>04/30/2021</u>	0.080 DGS	N/A	<u>AG916501</u> <u>06/14/2021</u>	Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Temperature Checks Performed By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.50</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5097</u>																																												
Simulator	Serial #	Lot #/Exp																																																													
0.050	<u>MP4863</u>	<u>201905A</u> <u>05/14/2021</u>																																																													
0.080	<u>MP4864</u>	<u>201905B</u> <u>05/14/2021</u>																																																													
0.200	<u>MP5097</u>	<u>201904D</u> <u>04/30/2021</u>																																																													
0.080 DGS	N/A	<u>AG916501</u> <u>06/14/2021</u>																																																													
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By <u>MH</u> Barometric Pressure ID# <u>28663</u> Gauge <u>1011</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2019-B</u> Acetone Stock Solution Lot # <u>2019-1</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td><u>SD3965</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD3966</u></td> </tr> <tr> <td>0.050</td> <td><u>MP4863</u></td> </tr> <tr> <td>0.080</td> <td><u>MP4864</u></td> </tr> <tr> <td>0.200</td> <td><u>MP5097</u></td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	<u>SD3965</u>	Interferent	<u>SD3966</u>	0.050	<u>MP4863</u>	0.080	<u>MP4864</u>	0.200	<u>MP5097</u>
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number	Lot Number	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	<u>SD3965</u>																																																														
Interferent	<u>SD3966</u>																																																														
0.050	<u>MP4863</u>																																																														
0.080	<u>MP4864</u>																																																														
0.200	<u>MP5097</u>																																																														
Notes/Suggested Service: <u>Tech Review: Added check marks in final box</u> <input checked="" type="checkbox"/> APPROVED <u>02/06/2020</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <u>SP 2/21/20</u> <u>Keith Kirkland</u> <u>2/24/2020</u> Tech Review / Date Admin Review / Date																																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 11:48

Date of Inspection: 02/06/2020

Serial Number: 80-007084
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201905A Exp: 05/14/2021	0.08g/210L Test (g/210L) Lot#:201905B Exp: 05/14/2021	0.20g/210L Test (g/210L) Lot#:201904D Exp: 04/30/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG916501 Exp: 06/14/2021
0.000	0.048	0.081	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.079
0.000	0.048	0.081	0.200	0.079
0.000	0.048	0.080	0.200	0.078
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.080	0.200	0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP
13K
2/24/2020

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MICHAEL D HAUGHEY

Signature and Printed Name

02/06/2020
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-007084	Miami Dade PD	06/06/2020	MK

0.05g/210L

0.047 to 0.053

0.08g/210L

0.077 to 0.083

0.20g/210L

0.194 to 0.206

DGS 0.08g/210L

0.077 to 0.083

MIRMI-DROE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
02/06/2020
Software: 8100.27
SN 80-007084

Test g/210L Time

Air Blank 0.000 09:54
Control Test 0.048 09:55
Air Blank 0.000 09:56
Control Test 0.047 09:56
Air Blank 0.000 09:57
Control Test 0.048 09:58
Control Test Stats 0.000 09:58
Average 0.0477
Std Dev 0.0006
Rel Std Dev(x) 1.2112

MIRMI-DROE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
02/06/2020
Software: 8100.27
SN 80-007094

Test g/210L Time

Air Blank 0.000 10:00
Control Test 0.079 10:00
Air Blank 0.000 10:01
Control Test 0.080 10:01
Air Blank 0.000 10:02
Control Test 0.080 10:02
Air Blank 0.000 10:03
Control Test 0.080 10:03
Control Test Stats 0.000 10:03
Average 0.0797
Std Dev 0.0006
Rel Std Dev(x) 0.7247

MIRMI-DROE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
02/06/2020
Software: 8100.27
SN 80-007084

Test g/210L Time

Air Blank 0.000 10:05
Control Test 0.199 10:06
Air Blank 0.000 10:06
Control Test 0.200 10:06
Air Blank 0.000 10:07
Control Test 0.200 10:07
Air Blank 0.000 10:08
Control Test 0.200 10:08
Air Blank 0.000 10:09
Control Test Stats 0.000 10:09
Average 0.1997
Std Dev 0.0006
Rel Std Dev(x) 0.2892

MIRMI-DROE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
02/06/2020
Software: 8100.27
SN 80-007084

Test g/210L Time

Air Blank 0.000 10:10
Control Test 0.078 10:11
Air Blank 0.000 10:11
Control Test 0.079 10:11
Air Blank 0.000 10:12
Control Test 0.079 10:12
Air Blank 0.000 10:13
Control Test Stats 0.000 10:13
Average 0.0787
Std Dev 0.0006
Rel Std Dev(x) 0.7339


MK
Operator's Signature

MK
Operator's Signature

MK
Operator's Signature

MK
Operator's Signature

SP
TSK
2/24/2020



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007084, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007084</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>02/06/2020</u>	0.080 g/210 L	0.005
Calibration Time:	<u>11:48</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/06/2020 Date *[Signature]* MICHAEL D HAUGHEY,

Department Inspector

SP 15K 2/24/2020