

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT RICHEY PD
Time of Inspection: 07:48

Date of Inspection: 08/14/2022

Serial Number: 80-000967
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.078		
0.000	0.048	0.079		
0.000	0.048	0.079		

Number of Simulators Used: 5

Remarks: IMPROPER KEY STROKE WHEN REVIEWING THE .08 ALCOHOL REFERENCE RESULTS, WHICH BYPASSED THE .20 TESTING AND FINAL DIAGNOSTIC CHECK.

The above instrument complies (~~X~~) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JEFFREY W MCCONNELL

Signature and Printed Name

08/14/2022

Date