



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Seminole CSO

Instrument Serial Number: 80-000998

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Keith Betham
Date of Inspection: 6/11/2021
Time of Inspection: 23:27:35
Agency Inspection Discrepancy: [X] Erroneous Information
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[X] Lot Number [] Expiration Date for 0.20 g/ 210L [X] Alcohol Reference Solution [] Dry Gas Standard is [X] Incorrect [] Expired.

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date:
[] Cylinder Change Records Date:
[] Control Test Records Date:
[] Diagnostic Check Records Date:
Comments: The lot number for the 0.20 g/210 L Alcohol Reference Solution was mistyped and contains an extra "0". Please see below for corrective action.

CORRECTIVE ACTION
[X] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 9/24/2021.
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other:

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.08.24 11:13:04 -04'00'

Signature of Alcohol Testing Program Staff Member

8/24/2021 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.
Time of Inspection: 23:27

Date of Inspection: 06/11/2021

Serial Number: 80-000998
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1433248 Exp: 01/06/2024
0.000	0.049	0.080	0.198	0.081
0.000	0.049	0.080	0.198	0.082
0.000	0.049	0.080	0.199	0.081

Number of Simulators Used: 5

Remarks: *Correct lot# .02 202005A hb 06/11/2021*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name
 KEITH D BETHAM

 06/11/2021
 Date