

(Amended)

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA PD

Time of Inspection: 21:01

Date of Inspection: 07/01/2021

Serial Number: 80-001000

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A4 Exp: 01/05/2022
0.000	0.046	0.080	0.000 / 0.196	0.077
0.000	0.046	0.080	0.000 / 0.196	0.077
0.000	0.047	0.081	0.000 / 0.196	0.077

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance.

Disconnected, simulator from instrument. Checked all parts and re-connected on 0.20 test. Worked correctly afterwards.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES H FRAZIER

Signature and Printed Name

07/01/2021
Date

Amended
**Florida Department of Law Enforcement
 Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA PD
 Time of Inspection: 22:13

Date of Inspection: 03/18/2021

Serial Number: 80-001000
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A4 Exp: 01/05/2022
0.000	0.048	0.079	/ 0.199	0.077
0.000	0.048	0.079	/ 0.203	0.077
0.000	0.048	0.078	/ 0.202	0.076

Number of Simulators Used: 5

Remarks:
 20: Ambient Fail.

see attached note.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Christopher Bradshaw #75 Bradshaw

 CHRISTOPHER B BRADSHAW
 Signature and Printed Name

03/18/2021
 Date