

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC P.D.
Time of Inspection: 14:48

Date of Inspection: 09/27/2021

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A4 Exp: 01/05/2022
0.000	0.046	0.078	0.201	0.079
0.000	0.047	0.078	0.203	0.079
0.000	0.047	0.079	0.202	0.080

Number of Simulators Used: 4

Remarks:

A F / M A: Range Exceeded.

To much mouth alcohol used during test

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #12

ZACHARY A GIBSON

Signature and Printed Name

09/27/2021
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Indialantic PD

Instrument Serial Number: 80-001006

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Zachary Gibson
Date of Inspection: 09-27-2021
Time of Inspection: 14:48
Agency Inspection Discrepancy: [X] Procedural
Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
Lot Number [] Expiration Date for ___g/ 210L [] Alcohol Reference Solution [] Dry Gas Standard is [] Incorrect [] Expired.
[X] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [X] REASON for repeating the following test(s); OR the [X] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[X] Alcohol Free Subject Test [X] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments: _____

CORRECTIVE ACTION
[X] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by _____ (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Israel Soto

Digitally signed by Israel Soto
Date: 2021.11.15 10:57:32 -05'00'

Signature of Alcohol Testing Program Staff Member

11/15/2021
Date