

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 10:23

Date of Inspection: 02/04/2021

Serial Number: 80-001045  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:201902C Exp: 02/20/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09820080A2 Exp: 06/05/2022
0.000	0.049	0.078	0.184 / 0.195	0.083
0.000	0.049	0.078	0.187 / 0.196	0.082
0.000	0.049	0.079	0.191 / 0.197	0.082

Number of Simulators Used: 5


**Remarks:**

20: Control Outside Tolerance.SIM NOT A CORRECT TEMP RERAN.

*REMOVED FROM EVID USB BY PTF 2-4-2021 DUE TO WATER LEAK  
N&AIL WORK STOPPED.*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

ROBERT D LIGHTSEY

02/04/2021  
Date