

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE ALFRED PD

Time of Inspection: 22:45

Date of Inspection: 04/18/2021

Serial Number: 80-001048

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A5 Exp: 01/05/2022
0.000	0.047	0.074 / 0.077	0.198	0.080
0.000	0.048	0.074 / 0.077	0.200	0.079
0.000	0.048	0.075 / 0.078	0.200	0.079

Number of Simulators Used: 4


Remarks:

08: Control Outside Tolerance.

** HAD TO REPEAT .08 DUE TO SOLUTION NOT BEING HOT ENOUGH. ay*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


#25 CAPD
 _____ ANTHONY W GETTLE
 Signature and Printed Name

04/18/2021
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Lake Alfred PD

Instrument Serial Number: 80-001048

AGENCY INSPECTION DATA REVIEW		
Agency Inspector: Anthony Gettle	Date of Inspection: 04/18/2021	Time of Inspection: 22:45:50
Agency Inspection Discrepancy: <input type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input checked="" type="checkbox"/> Other (Missing Required Information)		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for ____g/ 210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input checked="" type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input checked="" type="checkbox"/> 0.08 g/210L Test <input type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: The 0.08 g/210 L Test was repeated. The reason the test was repeated, as well as any corrective actions taken prior to repeating the test, must be included.
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 07/08/2021 (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by ____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2021.06.08 12:19:56 -04'00'

Signature of Alcohol Testing Program Staff Member

6/8/2021 Date