

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 22:08

Date of Inspection: 01/31/2021

Serial Number: 80-001075
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

~~WIT. Y. LORENZO~~

WIT. TPR. R.E. ADDISON

Y. L.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

YOSNIEL LORENZO

Signature and Printed Name

01/31/2021
Date

Amended

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI

Time of Inspection: 22:16

Date of Inspection: 01/31/2021

Serial Number: 80-001110

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.080	0.199	0.082
0.000	0.049	0.080	0.199	0.082

Number of Simulators Used: 5

Remarks:

WIT. Y. LORENZO

WIT. TPR R.E. ADDISON *J. L.*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

YOSNIEL LORENZO

Signature and Printed Name

01/31/2021
Date

Amended

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP MIAMI
Time of Inspection: 22:24

Date of Inspection: 01/31/2021

Serial Number: 80-001120
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
0.000	0.048	0.078	0.198	0.082
0.000	0.048	0.079	0.200	0.081
0.000	0.048	0.078	0.200	0.081

Number of Simulators Used: 5

Remarks:

~~WIT. Y. LORENZO~~

WIT. TPR. R.E. ADDISON J. L.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

YOSNIEL LORENZO

Signature and Printed Name

01/31/2021
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 21:39

Date of Inspection: 01/31/2021

Serial Number: 80-001121
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1114319 Exp: 03/27/2022
0.000	0.048	0.079	0.199	0.082
0.000	0.048	0.079	0.199	0.082
0.000	0.048	0.080	0.200	0.082

Number of Simulators Used: 5

Remarks:

WIT.

WITNESS: OFC. Y. LORENZO - REA

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RAYMOND E ADDISON

Signature and Printed Name

01/31/2021
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TRP E MIAMI

Time of Inspection: 01:35

Date of Inspection: 02/01/2021

Serial Number: 80-006626

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:1114319 Exp: 03/27/2022	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.047	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		
0.000	0.048	0.000 / 0.000		

Number of Simulators Used: 5

Remarks:

08: Control Outside:HOSE ADJUSTED NEW TANK; FAILED SECOND . FAILED SECONSNON-compliance:FAILED SECO
ND 08 GAS OFFLINE FOR MAINTENANCE.

08 GAS ATTACHED INSTEAD OF 08 SOLUTION SECOND AI PERFORMED
AND INSPECTION PASSED: R4A

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ray R. E. Addison

RAYMOND E ADDISON

Signature and Printed Name

02/01/2021
Date