Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.
Time of Inspection: 11:34 Date of Inspection: 12/16/2021

Serial Number: 80-001345
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049	0.080	0.195	
0.000	0.049	0.081	0.195	
0.000	0.049	0.080	0.196	

Number of Simulators Used: 5

Remarks:

The instrument skipped the dry gas ~~check~~ test and post-inspection diagnostic check due to pressing escape key too many times after reviewing prior test results.

Inadvertently selected "complies" when test was not in compliance due to aborted inspection.

The above instrument complies (X) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CODY L WADE
Signature and Printed Name

12/16/2021
Date