

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 17:05

Date of Inspection: 10/12/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
/ 0.000	0.050	0.080	0.199	0.079
/ 0.000	0.050	0.081	0.200	0.078
/ 0.000	0.050	0.080	0.199	0.078

Number of Simulators Used: 4

Remarks:

00: Ambient Fail.

AMENDED 12/22/2021

PURGE FAIL AT BEGINNING OF ALCOHOL FREE TEST

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DEAN BALDWIN

Signature and Printed Name

10/12/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 09:52

Date of Inspection: 06/25/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.049	0.080	0.200	0.000 / 0.079
0.000	0.049	0.080	0.201	0.000 / 0.079
0.000	0.048	0.080	0.202	0.000 / 0.079

Number of Simulators Used: 4


Remarks:

08: FAILED TO CONNECT .08 GAS HOSE.

Exposed Mouth Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DEAN BALDWIN

06/25/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 16:52

Date of Inspection: 05/24/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.050	0.080	0.201	0.079
0.000	0.051	0.080	0.202	0.079
0.000	0.050	0.081	0.204	0.079


Number of Simulators Used: 4

Remarks:

Expanded Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DEAN BALDWIN
Signature and Printed Name

05/24/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 17:14

Date of Inspection: 04/26/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.049	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.081	0.202	0.079


Number of Simulators Used: 4

Remarks:

Expend Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DEAN BALDWIN

04/26/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 16:47

Date of Inspection: 03/24/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219080A3 Exp: 01/05/2022
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.081	0.201	0.079


Number of Simulators Used: 4

Remarks:

Exhaled Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DEAN BALDWIN

03/24/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD
Time of Inspection: 17:46

Date of Inspection: 02/24/2021

Serial Number: 80-005477
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:30219080A3 Exp: 01/05/2022
0.000	0.050	0.080	0.201	0.079
0.000	0.051	0.081	0.201	0.080
0.000	0.051	0.081	0.201	0.080

Number of Simulators Used: 4

Remarks:

Expired Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DEAN BALDWIN
Signature and Printed Name

02/24/2021
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DADE CITY PD

Time of Inspection: 16:42

Date of Inspection: 01/19/2021

Serial Number: 80-005477

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#: 201908B Exp: 08/07/2021	0.20g/210L Test (g/210L) Lot#: 202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 30219030A3 Exp: 01/05/2022
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.081	0.202	0.080
0.000	0.050	0.081	0.202	0.081

Number of Simulators Used: 4

Remarks:

Expired Mouth Alcohol Used

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 DEAN BALDWIN
Signature and Printed Name

01/19/2021
Date