Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 09:00

Date of Inspection: 02/25/2021

Serial Number: 80-006046 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201910D Exp: 10/22/2021	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1249794 Exp: 12/13/2022
0.000	0.047	0.076	0.195	0.079
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

NO ISSUES

The above instrument complies (X) does not comply (

) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Cont

WILLIAM T SMITH
Signature and Printed Name

02/25/2021

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL Time of Inspection: 23:52

Date of Inspection: 11/28/2021

Serial Number: 80-006046 Software: 8100.27

	YES	NO
Check or Test		
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1506449 Exp: 06/10/2024
/ 0.000	0.048	0.078	0.193	0.079
/ 0.000	0.049	0.078	0.193	0.080
/ 0.000	0.049	0.077	0.194	0.080

Number of Simulators Used: 5

Remarks: 00: Ambient Fail.

Air Blank Successful after a Second/repeated Past

The above instrument complies (X) does not comply (

) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

	TTTTTTN TONTT	
	WILLIAM T SMITH	
Signature and Pr.	inted Name	

11/28/2021 Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Florida Highway Patrol

Time of Inspection: Date of Inspection: 12/30/2021

Serial Number: 80-006046 Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		x
Diagnostic Check (Pre-Inspection): OK		x
Alcohol Free Subject Test: 0.000		x
Mouth Alcohol Test: Slope Not Met		x
Interferent Detect Test: Interferent Detect		x
Diagnostic Check (Post-Inspection): OK		x

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used:

Remarks: Unable to do monthly inspection due to EEPROM failure. Will send to FDLE to evaluate prior to possible repair.

The above instrument complies (_____) does not comply (_____) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANTHONY N. DOBOSIEWICZ

Signature and Printed Name

12/30/2021 Date

FDLE/ATP Form 40 - March 2004 Ref. 11D-8.006