

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 05:15

Date of Inspection: 12/30/2021

Serial Number: 80-006637  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   | Yes |    |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202108D<br>Exp: 08/17/2023 | 0.20g/210L Test (g/210L)<br>Lot#:202106B<br>Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:1433249<br>Exp: 01/08/2024 |
|----------------------------|---|---|---|--|
| 0.000                      | 0.049   | 0.083   | 0.200   | 0.080  |
| 0.000                      | 0.049   | 0.082   | 0.200   | 0.080  |
| 0.000                      | 0.049   | 0.083   | 0.201   | 0.081  |

Number of Simulators Used: 5


Remarks:

Time-Date changed. A F / M A: NO MOUTH ALCOHOL.

Needed to put more Mouth Alcohol in mouth, as it diluted prior to providing sample.

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

ANTHONY N DOBOSIEWICZ

12/30/2021  
Date

Amended

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 23:50

Date of Inspection: 09/28/2021

Serial Number: 80-006637  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

|                            |   |   |   |  |
|----------------------------|---|---|---|--|
| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202101C<br>Exp: 01/12/2023 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:1433249<br>Exp: 01/08/2024 |
| 0.000                      | 0.046   | 0.076   | 0.194   | 0.080  |
| 0.000                      | 0.047   | 0.078   | 0.200   | 0.079  |
| 0.000                      | 0.047   | 0.079   | 0.200   | 0.079  |

Number of Simulators Used: 5

Remarks:

MOUTH ALCOHOL RANGE EXCEEDED/NO SLOPE


MOUTH ALCOHOL RANGE EXCEEDED/NO SLOPE

No issues on repeated mouth alcohol test

WS

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

WILLIAM T SMITH

09/28/2021  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 09:25

Date of Inspection: 02/25/2021

Serial Number: 80-006637  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201910D<br>Exp: 10/22/2021 | 0.08g/210L Test (g/210L)<br>Lot#:202007A<br>Exp: 07/21/2022 | 0.20g/210L Test (g/210L)<br>Lot#:202005A<br>Exp: 05/12/2022 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:915310<br>Exp: 03/07/2021 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.077   | 0.198   | 0.082   |
| 0.000                      | 0.048   | 0.078   | 0.200   | 0.081   |
| 0.000                      | 0.047   | 0.078   | 0.198   | 0.082   |

Number of Simulators Used: 5

Remarks:  
NO ISSUES

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

02/25/2021  
Date