

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SHERIFF'S OFFICE

Serial Number: 80-006937

Time of Inspection: 16:30

Date of Inspection: 02/26/2021

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 0

Remarks: I COMPLETED THE FEBRUARY AGENCY INSPECTION ON 02/26/2021. I COULD NOT PRINT THE COMPLETED AGENCY INSPECTION (FORM 40) DUE TO ERROR DISPLAYED "FORM NOT FOUND"

DURING THE UPLOAD TO THE F.D.L.E. SERVER, THE AGENCY INSPECTION FOR FEBRUARY 2021 WAS LOST.

I TOOK INSTRUMENT TO F.D.L.E. ON 03/05/2021 AND DEPARTMENT INSPECTOR ISRAEL SOTO WAS ALSO UNABLE TO RETRIEVE THE FEBRUARY 2021 INSPECTION.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

03/10/2021

Date