

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FHP Instrument Serial Number: 80-007442

AGENCY INSPECTION DATA REVIEW								
Agency Inspector: Sgt. Kari Hilliard			Date of Inspection: 11	/22/2021	Time of Inspection: 20:22			
Age	Agency Inspection Discrepancy: ☐ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information ☐ Procedural ☐ Other							
	Agency Inspection Not Conducted or Re	ecords regarding Ag	ency Inspection have no	ot been uploaded	d.			
	□ Lot Number □Expiration Date for <u>g</u> / 210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired.							
X (FDLE/ATP Form 39 states in part, "If a Remarks section of FDLE/ATP Form 40 test(s); OR the ☐ Possible Cause and ☐ Alcohol Free Subject Test ☐ ☐ 0.05 g/210L Test ☑ ☐ .08 g/210L Dry Gas Standard Test	Agency Inspection l Corrective Action Ta	Report – Intoxilyzer 8000	D. The ⊠EASC et(s) was not rec est □ Int	ON for repeating the following			
	□ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." □ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. □ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. □ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.							
	The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.							
	Other:							
ОТЬ	IER ELECTRONIC DATA REVIEW							
	Login Records Date:	Comments: RFI detect wa trigg	ered with no comment.	Inspector's cam	era microphone set off			
	Cylinder Change Records Date:	detector.		•				
	Control Test Records Date:							
	Diagnostic Check Records Date:							
COL	RRECTIVE ACTION							
	Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by (Date). Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required							
5	ignature of Alcohol Testing Program Staf	f Member		12/22/202 Date	<u>21</u>			

Umended

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 20:22

Date of Inspection: 11/22/2021

Serial Number: 80-007442

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03521080A3 Exp: 03/05/2023
0.000	0.046	0.078	0.195	/ 0.081
0.000	0.047	0.078	0.197	/ 0.080
0.000	0.048	0.078	0.197	/ 0.081

Number of Simulators Used: 5

Remarks:

08: RFI Detect.

Camera Mic KRAN 12/22/21

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

KARI A HILLIARD

11/22/2021