

Graham, Thomas

From: Gutschow, Taylor
Sent: Wednesday, June 16, 2021 10:15 AM
To: Pinsky, David
Cc: Graham, Thomas
Subject: 80-001064
Attachments: SKM_C754e21061608252.pdf

Please add this Form 40 to the record for 80-001064. The May 2021 Agency Inspection was not performed due to the agency not having the necessary Alcohol Reference Solutions on hand. The AI has been advised on what to do if this should happen again in the future.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Jason Engel <EngelJ@CityofNewPortRichey.org>
Sent: Wednesday, June 16, 2021 8:11 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Subject: RE: Intoxilizer Inspections

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Thank you for the help.

Regards, Jason

From: Gutschow, Taylor [<mailto:TaylorGutschow@fdle.state.fl.us>]
Sent: Monday, June 14, 2021 10:25 AM
To: Jason Engel <EngelJ@CityofNewPortRichey.org>
Subject: RE: Intoxilizer Inspections

Hi,

Please fill out this blank Form 40, include a comment that the inspection was unable to be performed due to a lack of Alcohol Reference Solutions on hand, then return it to me with your signature at the bottom. If you find yourself in a similar situation in the future, please reach out to your local agencies to see if you can borrow their solutions. I'd also advise doing inspections earlier in the month rather than the end of the month to avoid these last-minute situations. Missing an agency inspection on an instrument with evidential breath tests does violate Chapter 11D-8 and could jeopardize the affected breath tests in court.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: Jason Engel <EngelJ@CityofNewPortRichey.org>
Sent: Sunday, May 30, 2021 6:44 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Subject: Intoxilizer Inspections

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

Hello Taylor,

After last month's inspection I forgot to order more reference solutions for the inspections so I will not be able to do an inspection for May. I know the instrument will be down until I complete the inspection. Is there anything special I need to do now that I will be missing a month? I've never missed a month before so I'm unsure. Thank you Jason.

Regards,