



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Pasco CSO

Instrument Serial Number: 80-003347

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Barry Nixon
Date of Inspection: 5/05/2021
Time of Inspection: 15:53:13
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [x] Erroneous Information
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[x] Lot Number [] Expiration Date for 0.05 g/ 210L [x] Alcohol Reference Solution [] Dry Gas Standard is [x] Incorrect [] Expired.
[] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [] REASON for repeating the following test(s); OR the [] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments:
The lot number for the 0.05 g/210 L Alcohol Reference Solution is incorrect (missing a number).

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 9/24/2021.
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Signature of Alcohol Testing Program Staff Member

8/24/2021
Date