



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

November 21, 2021

Taylor Gutschow
Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

REF: 11-21- 2021 Inspection

Dear Mr. Gutschow,

Agency Inspection results for the month of November, 2021 inspection. Inspections for the month of November, 2021 instruments 80-007505 passed inspection without any issues. Instrument 80-007507 had a purge fail and ambient fail after the mouth alcohol sample. Inspection was aborted and the trashcan under the counter that contained my spent mouth alcohol was removed from the room. Inspection was re-ran and passed with no issues.

Sgt. S. Abeln
Agency Inspector


Cc: File



Highlands County Sheriff's Office

Sheriff Paul Blackman

400 S. Eucalyptus Street
Sebring, Florida 33870
863-402-7200
www.highlandssheriff.org



Dually Accredited

December 13, 2021

Taylor Gutschow
Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

REF: 12-13- 2021 Inspection

Dear Mr. Gutschow,

Agency Inspection results for the month of November, 2021 inspection. Inspections for the month of November, 2021 instruments 80-007505 passed inspection without any issues. Instrument 80-007507 had a purge fail and ambient fail after the mouth alcohol sample. Inspection was aborted and the trashcan under the counter that contained my spent mouth alcohol was removed from the room. Inspection was re-ran and again had a purge and ambient failure. I took the instrument out of service and packaged it up and will be shipping it out to you.

Sgt. S. Abeln
Agency Inspector

A handwritten signature in black ink, appearing to read "S. Abeln".

Cc: File