FDLE
Florida Department of Law Enforcement

INSTRUMENT PROCESSING SHEET

Agency Pinellas County Sheriff's Office S/N 80-000888

Date In 1/23/2021 DI Completion Date 1/25/2021 ☐Ship ☑P/U ☐H/D ☐CMI ☐EE

Intake Performed By DERR

☑ Annual
☐ Registration
☐ Return from CMI / EE

Visual Inspection:
☑ Case ☑ Handle
☑ Keyboard ☑ Dry Gas Shelf
☑ Feet ☑ Breath Tube
☑ Ports ☑ Screws Tight

Other Equipment/ Accessories:
☐ Power cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes: _____

Final Release Date

FDLE Alcohol Testing Program

Digitally signed by FDLE Alcohol Testing Program
Date: 2021.01.28 11:52:14 -05'00'

Quality Checks Performed By DERR

☑ Breath Tube Screen
☑ Replace External O-Rings
☑ Instrument Set Up Verified
☑ R-Value 221
☑ Flow Verification (L/s)
Flow Column # ATP 104
32 mm 0.142 (.139 - .169)
36 mm 0.167 (.156 - .190)
53 mm 0.238 (.228 - .278)
103 mm 0.515 (.447 - .547)
☑ Barometric Pressure Check
Gauge ID # 28663
☑ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	SD3967	202010A
		10/05/2022
0.080	SD3968	202010B
		10/05/2022
0.200	SD3969	202010D
		10/06/2022
0.080 DGS	N/A	AG003005
		01/30/2022

Flow Calibration Performed By _____

Flow Column # _____
☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm
☐ R-Value _____
☐ Post Calibration Verification (L/s)
Flow Column # _____
32 mm _____ (.139 - .169)
36 mm _____ (.156 - .190)
53 mm _____ (.228 - .278)
103 mm _____ (.447 - .547)

Maintenance Performed By _____

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Breath Tube Replacement
☐ Other _____

Temperature Checks Performed By DERR

☑ Lab Temp °C 22.68C
External Digital Therm. ID#: 300918
☑ 34°C +-.2 Serial #: SD3967
☑ 34°C +-.2 Serial #: SD3968
☑ 34°C +-.2 Serial #: SD3969

Calibration Adjustment Performed By _____

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection Performed By DERR

Barometric Pressure ID# 68639
Gauge 1017 Instrument 1018
Mouth Alcohol Solution Lot # 2020-A
Acetone Stock Solution Lot # 2020-A

Simulator	Serial Number
0.000	SD3965
Interferent	SD3966
0.050	SD3967
0.080	SD3968
0.200	SD3969

Attachments

☑ Form 41	☐ Post-Stability Checks
☑ Stability Checks	☐ Flow Calibration
☑ Calibration Certificate	☑ Form 40
☐ Calibration Adjustment	☐ Other _____

Notes/Suggested Service: _____

☑ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
☑ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**
☑ **Conduct an Agency Inspection Before Evidentiary Use**

Israel Soto 01-26-2021
Tech Review / Date

2021.01.28 11:51:49 -05'00'
Admin Review / Date

15

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000888 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000888
Owning Agency:	PINELLAS COUNTY SO
Calibration Date:	01/25/2021
Calibration Time:	13:43

	UNCERTAINTY* ±
0.050 g/ 210 L	0.005
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,
without written approval of the Florida Department of
Law Enforcement Alcohol Testing Program.

01/25/2021
Date

DAVID E REYES-RIVERA,
Department Inspector

2021.01.28 11:51:34 -05'00'

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 13:43

Date of Inspection: 01/25/2021

Serial Number: 80-000888
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG003005 Exp: 01/30/2022
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079

Standard Deviations	0.0008	0.0006	0.0006	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

15

2021.01.28 11:51:15 -05'00'

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera

DAVID E REYES-RIVERA
Signature and Printed Name

01/25/2021
Date

FDLE/ATP Form 41 –Revised August 2005

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000888	Pinellas County Sheriff's Office	01/25/2021	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000888
01/25/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:11
Control Test	0.049	10:12
Air Blank	0.000	10:13
Control Test	0.048	10:13
Air Blank	0.000	10:14
Control Test	0.048	10:15
Air Blank	0.000	10:15

Control Test Stats

Average	0.0483
Std Dev	0.0016
Rel Std Dev(%)	3.3145

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000888
01/25/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:34
Control Test	0.078	10:34
Air Blank	0.000	10:35
Control Test	0.077	10:35
Air Blank	0.000	10:36
Control Test	0.077	10:37
Air Blank	0.000	10:37

Control Test Stats

Average	0.0773
Std Dev	0.0026
Rel Std Dev(%)	3.7468

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000888
01/25/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.197	10:26
Air Blank	0.000	10:27
Control Test	0.196	10:28
Air Blank	0.000	10:28
Control Test	0.195	10:29
Air Blank	0.000	10:30

Control Test Stats

Average	0.1960
Std Dev	0.0010
Rel Std Dev(%)	0.5102

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000888
01/25/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:44
Control Test	0.078	10:44
Air Blank	0.000	10:45
Control Test	0.079	10:45
Air Blank	0.000	10:45
Control Test	0.079	10:46
Air Blank	0.000	10:46

Control Test Stats

Average	0.0787
Std Dev	0.0006
Rel Std Dev(%)	0.7339

Operator's Signature

15

2021.01.28 11:50:58 -05'00'

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 09:02 Date of Inspection: 01/25/2021
Serial Number: 80-000888
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

JS

2021.01.28 11:50:42 -05'00'

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rep Rivera
Signature and Printed Name

DAVID E REYES-RIVERA

01/25/2021
Date

FDLE/ATP Form 40 – March 2004