



INSTRUMENT PROCESSING SHEET

Agency Charlotte CSOS/N 80-000946

Florida Department of Law Enforcement

Date In 10/22/2021DI Completion Date 10/28/2021 Ship P/U H/D CMI EE

Intake	By TDG	Quality Checks	By TDG	Date 10/28/2021	Flow Calibration	By	Date																																																														
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Exhaust cover broke off.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>154</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP5092</td> <td>202010A</td> </tr> <tr> <td></td> <td>10/05/2022</td> </tr> <tr> <td>0.080</td> <td rowspan="2">MP4864</td> <td>202010B</td> </tr> <tr> <td></td> <td>10/05/2022</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5094</td> <td>202010D</td> </tr> <tr> <td></td> <td>10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705</td> </tr> <tr> <td></td> <td></td> <td>09/23/2022</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A		10/05/2022	0.080	MP4864	202010B		10/05/2022	0.200	MP5094	202010D		10/06/2022	0.080 DGS	N/A	AG026705			09/23/2022	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>Maintenance</th> <th>By TDG</th> </tr> </thead> <tbody> <tr> <td> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other Replaced exhaust cover </td> <td></td> </tr> </tbody> </table>					Maintenance	By TDG	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other Replaced exhaust cover																																	
Simulator	Serial #	Lot #/Exp																																																																			
0.050	MP5092	202010A																																																																			
		10/05/2022																																																																			
0.080	MP4864	202010B																																																																			
		10/05/2022																																																																			
0.200	MP5094	202010D																																																																			
		10/06/2022																																																																			
0.080 DGS	N/A	AG026705																																																																			
		09/23/2022																																																																			
Maintenance	By TDG																																																																				
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other Replaced exhaust cover																																																																					
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>DI Temp. Checks</th> <th>By TDG</th> </tr> </thead> <tbody> <tr> <td> <input checked="" type="checkbox"/> Lab Temp °C <u>381198</u> External Digital Therm. ID#: <u>20.34</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5094</u> </td> <td></td> </tr> </tbody> </table>			DI Temp. Checks	By TDG	<input checked="" type="checkbox"/> Lab Temp °C <u>381198</u> External Digital Therm. ID#: <u>20.34</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5094</u>																																																														
DI Temp. Checks	By TDG																																																																				
<input checked="" type="checkbox"/> Lab Temp °C <u>381198</u> External Digital Therm. ID#: <u>20.34</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5092</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5094</u>																																																																					
Calibration Adjustment		By		Department Inspection				By TDG																																																													
Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					Barometric Pressure ID# <u>28199</u> Gauge <u>1009</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr style="background-color: #e0e0e0;"> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3963</td> </tr> <tr> <td>Interferent</td> <td>SD1017</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD3963	Interferent	SD1017	0.050	MP5092	0.080	MP4864	0.200	MP5094		
Simulator	Serial #	Lot #	Expiration																																																																		
0.000		N/A	N/A																																																																		
0.040																																																																					
0.100																																																																					
0.200																																																																					
0.300																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial #	Lot #	Expiration																																																																		
0.050																																																																					
0.080																																																																					
0.200																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial Number																																																																				
0.000	SD3963																																																																				
Interferent	SD1017																																																																				
0.050	MP5092																																																																				
0.080	MP4864																																																																				
0.200	MP5094																																																																				
Notes/Suggested Service: <u>On 10/28/2021, I replaced the damaged exhaust cover and spoke to Al Michael Kern about sending us a replacement. (TDG)</u>				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																																	
		David Eliezer Reyes-Rivera		<small>Digitally signed by David Eliezer Reyes-Rivera Date: 2021.11.02 11:05:21 -0400</small>				2021.11.02 11:54:59																																																													
		Tech Review / Date		Admir. Review / Date																																																																	

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000946	Charlotte CSO	10/28/2021	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053</p> <p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000946 10/28/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:08 Control Test 0.049 12:09 Air Blank 0.000 12:10 Control Test 0.049 12:11 Air Blank 0.000 12:12 Control Test 0.049 12:12</p> <p>Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MG</i></p>	<p>0.077 to 0.083</p> <p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000946 10/28/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:16 Control Test 0.078 12:16 Air Blank 0.000 12:17 Control Test 0.078 12:18 Air Blank 0.000 12:19 Control Test 0.078 12:19</p> <p>Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MG</i></p>	<p>0.194 to 0.206</p> <p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000946 10/28/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:20 Control Test 0.198 12:21 Air Blank 0.000 12:22 Control Test 0.198 12:23 Air Blank 0.000 12:24 Control Test 0.199 12:24</p> <p>Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911</p> <p>Operator's Signature <i>MG</i></p>	<p>0.077 to 0.083</p> <p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000946 10/28/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:25 Control Test 0.079 12:26 Air Blank 0.000 12:26 Control Test 0.079 12:26 Air Blank 0.000 12:27 Control Test 0.079 12:27</p> <p>Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MG</i></p>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 14:32

Date of Inspection: 10/28/2021

Serial Number: 80-000946
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#: 202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#: 202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG026705 Exp: 09/23/2022
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.078

Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

TAYLOR D GUTSCHOW

10/28/2021
 Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000946, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000946</u>	UNCERTAINTY* ±
Owning Agency:	<u>CHARLOTTE COUNTY SO</u>	0.050 g/ 210 L 0.005
Calibration Date:	<u>10/28/2021</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:32</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/28/2021

Date

TAYLOR D GUTSCHOW,
 Department Inspector

FDLE/ATP Form 69 January 2021
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality