



**INSTRUMENT PROCESSING SHEET**

Agency Coconut Creek Police Department

S/N 80-001044

Florida Department of Law Enforcement

Date In 4/30/2021 DI Completion Date 5/4/2021

Ship  P/U  H/D  CMI  EE

Intake	By DERR	Quality Checks	By DER	Date	5/4/2021	Flow Calibration	By	Date																																																													
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>253</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP5092</td> <td>202010A</td> </tr> <tr> <td></td> <td>10/05/2022</td> </tr> <tr> <td>0.080</td> <td rowspan="2">MP5093</td> <td>202010B</td> </tr> <tr> <td></td> <td>10/05/2022</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5094</td> <td>202010D</td> </tr> <tr> <td></td> <td>10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG026705</td> </tr> <tr> <td></td> <td></td> <td>09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202010A		10/05/2022	0.080	MP5093	202010B		10/05/2022	0.200	MP5094	202010D		10/06/2022	0.080 DGS	N/A	AG026705			09/23/2022				Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
Simulator	Serial #	Lot #/Exp																																																																			
0.050	MP5092	202010A																																																																			
		10/05/2022																																																																			
0.080	MP5093	202010B																																																																			
		10/05/2022																																																																			
0.200	MP5094	202010D																																																																			
		10/06/2022																																																																			
0.080 DGS	N/A	AG026705																																																																			
		09/23/2022																																																																			
<b>Calibration Adjustment</b>						<b>Department Inspection</b>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____						Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1016</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1017</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD1017	Interferent	SD3966	0.050	MP5092	0.080	MP5093	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																																		
0.000		N/A	N/A																																																																		
0.040																																																																					
0.100																																																																					
0.200																																																																					
0.300																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial #	Lot #	Expiration																																																																		
0.050																																																																					
0.080																																																																					
0.200																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial Number																																																																				
0.000	SD1017																																																																				
Interferent	SD3966																																																																				
0.050	MP5092																																																																				
0.080	MP5093																																																																				
0.200	MP5094																																																																				
<b>Attachments</b>						<b>Instrument Compliance</b>																																																															
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment						<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																															
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use						<u>2021.05.</u>																																																															
Israel Soto <small>Digitally signed by Israel Soto Date: 2021.05.04 12:10:04 -0400</small>						<u>04</u> <u>13:38:37</u> Tech Review / Date      Admin Review / Date <u>04/00</u>																																																															



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001044, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001044</u>	UNCERTAINTY* ±	
Owning Agency:	<u>COCONUT CREEK PD</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>05/04/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are iraceable to NIST. Thermometer temperatures are checked with NIST traceable Eurechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/04/2021

Date

**DAVID E REYES-RIVERA,**  
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 08:29

Date of Inspection: 05/04/2021

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.048	0.081	0.202	0.080
0.000	0.048	0.080	0.202	0.080
0.000	0.049	0.081	0.201	0.080
0.000	0.050	0.080	0.202	0.079
0.000	0.049	0.081	0.201	0.080
0.000	0.050	0.081	0.201	0.080
0.000	0.049	0.081	0.201	0.080
0.000	0.050	0.081	0.201	0.080


Standard Deviations	0.0007	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:





The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA  
 \_\_\_\_\_  
 Signature and Printed Name

05/04/2021  
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001044	Coconut Creek Police Department	5/4/2021	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p><b>0.047 to 0.053</b> <input checked="" type="checkbox"/></p> <p>COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 05/04/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:49</p> <p>Control Test 0.049 06:50</p> <p>Air Blank 0.000 06:50</p> <p>Control Test 0.048 06:51</p> <p>Air Blank 0.000 06:51</p> <p>Control Test 0.048 06:52</p> <p>Air Blank 0.000 06:52</p> <p>Control Test Stats</p> <p>Average 0.0483</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1945</p> <p>Operator's Signature </p>	<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 05/04/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:54</p> <p>Control Test 0.081 06:55</p> <p>Air Blank 0.000 06:55</p> <p>Control Test 0.080 06:56</p> <p>Air Blank 0.000 06:56</p> <p>Control Test 0.080 06:57</p> <p>Air Blank 0.000 06:57</p> <p>Control Test Stats</p> <p>Average 0.0803</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7187</p> <p>Operator's Signature </p>	<p><b>0.194 to 0.206</b> <input checked="" type="checkbox"/></p> <p>COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 05/04/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 06:58</p> <p>Control Test 0.200 06:59</p> <p>Air Blank 0.000 07:00</p> <p>Control Test 0.201 07:00</p> <p>Air Blank 0.000 07:01</p> <p>Control Test 0.201 07:01</p> <p>Air Blank 0.000 07:02</p> <p>Control Test Stats</p> <p>Average 0.2007</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2877</p> <p>Operator's Signature </p>	<p><b>0.077 to 0.083</b> <input checked="" type="checkbox"/></p> <p>COCONUT CREEK PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001044 05/04/2021 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:03</p> <p>Control Test 0.080 07:04</p> <p>Air Blank 0.000 07:04</p> <p>Control Test 0.079 07:04</p> <p>Air Blank 0.000 07:05</p> <p>Control Test 0.080 07:05</p> <p>Air Blank 0.000 07:06</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>