



INSTRUMENT PROCESSING SHEET

Agency Clay County Sheriff's Office

S/N 80-001071

Florida Department of
Law Enforcement

Date In 11/19/2021 DI Completion Date 11/22/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Agency included note that instrument had a modem failure and could not upload records. Inspection records uploaded with no issues and mutiple Diagnostic checks done that were all resulted OK.-RAW 11/22/21</u>	Quality Checks By <u>RAW</u> Date <u>11/22/21</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>184</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-22</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B 10-05-22</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-22</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-23</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-22	0.080	MP5089	202010B 10-05-22	0.200	MP5090	202010D 10-06-22	0.080 DGS	N/A	AG113403 05-14-23	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ DI Temp. Checks By <u>RAW</u> ☒ Lab Temp °C <u>21.79</u> External Digital Therm. ID#: <u>381189</u> ☒ 34°C +/-2 Serial #: <u>MP5088</u> ☒ 34°C +/-2 Serial #: <u>MP5089</u> ☒ 34°C +/-2 Serial #: <u>MP5090</u>																																															
Simulator	Serial #	Lot #/Exp																																																														
0.050	MP5088	202010A 10-05-22																																																														
0.080	MP5089	202010B 10-05-22																																																														
0.200	MP5090	202010D 10-06-22																																																														
0.080 DGS	N/A	AG113403 05-14-23																																																														
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Forgot to replace small O-ring on .05g/210L and got lower value. Replaced O-ring and got nominal values for .05 solution after repeated test. -RAW, 11/22/21</u>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>RAW</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1013</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> Attachments <table style="width:100%;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Israel Soto "Digitally signed by Israel Soto Date: 2021.11.23 09:01:58 +05'00" Tech Review / Date _____ <u>2021.11.23</u> <u>09:20:36</u> Admin Review / Date _____		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																													
0.000		N/A	N/A																																																													
0.040																																																																
0.100																																																																
0.200																																																																
0.300																																																																
0.080 DGS	N/A																																																															
Simulator	Serial #	Lot #	Expiration																																																													
0.050																																																																
0.080																																																																
0.200																																																																
0.080 DGS	N/A																																																															
Simulator	Serial Number																																																															
0.000	MP5086																																																															
Interferent	MP5087																																																															
0.050	MP5088																																																															
0.080	MP5089																																																															
0.200	MP5090																																																															
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 14:25

Date of Inspection: 11/22/2021

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.079	0.200	0.079
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.200	0.080

Standard Deviations	0.0004	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 RICHARD A WILLIAMS
Signature and Printed Name

11/22/2021
Date

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
11/22/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:04
Control Test	0.035	12:05
Air Blank	0.000	12:05
Control Test	0.049	12:06
Air Blank	0.000	12:07
Control Test	0.049	12:07
Air Blank	0.000	12:08
Control Test Stats		
Average	0.0443	
Std Dev	0.0081	
Rel Std Dev(%)	18.2321	

.05
#1

RAW

Operator's Signature

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
11/22/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:09
Control Test	0.048	12:10
Air Blank	0.000	12:10
Control Test	0.049	12:11
Air Blank	0.000	12:11
Control Test	0.049	12:12
Air Blank	0.000	12:13
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

.05
#2

RAW

Operator's Signature

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
11/22/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.078	12:16
Air Blank	0.000	12:17
Control Test	0.079	12:17
Air Blank	0.000	12:18
Control Test	0.079	12:18
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

.08
wet

RAW

Operator's Signature

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
11/22/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:20
Control Test	0.200	12:21
Air Blank	0.000	12:21
Control Test	0.199	12:22
Air Blank	0.000	12:23
Control Test	0.200	12:23
Air Blank	0.000	12:24
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

.20

RAW

Operator's Signature

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
11/22/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:33
Control Test	0.081	12:34
Air Blank	0.000	12:34
Control Test	0.081	12:34
Air Blank	0.000	12:35
Control Test	0.080	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

.08
DGS

RAW

Operator's Signature

Stability Checks

80-001071

11/22/21

RAW



This font is not yet supported

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001071, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001071</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>11/22/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams

Digitally signed by Richard A
Williams
Date: 2021.11.22 17:22:51 -05'00'

11/22/2021

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1



INSTRUMENT PROCESSING SHEET

Agency Clay County SO

S/N 80-001071

Florida Department of
Law Enforcement

Date In 01-11-2021 DI Completion Date 01-13-2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>IS</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☐ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ 	Quality Checks Performed By <u>IS</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>177</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> <td>202010B 10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG011102 04-20-2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-2022	0.080	MP5086	202010B 10-05-2022	0.200	MP5090	202010D 10-06-2022	0.080 DGS	N/A	AG011102 04-20-2022	Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By <u>IS</u> ☒ Lab Temp °C <u>21.11</u> External Digital Therm. ID#: <u>300505</u> ☒ 34°C +/-2 Serial #: <u>MP5088</u> ☒ 34°C +/-2 Serial #: <u>MP5086</u> ☒ 34°C +/-2 Serial #: <u>G2078</u>																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5088	202010A 10-05-2022																																																
0.080	MP5086	202010B 10-05-2022																																																
0.200	MP5090	202010D 10-06-2022																																																
0.080 DGS	N/A	AG011102 04-20-2022																																																
Final Release Date FDLE Alcohol Testing Program Digitally signed by FDLE Alcohol Testing Program Date: 2021.01.16 11:31:09 -05'00'	Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ 		Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial Number	Lot Number	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection Performed By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1022</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G11621</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>G2078</td> </tr> </tbody> </table> Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u>		Simulator	Serial Number	0.000	G11621	Interferent	MP5087	0.050	MP5088	0.080	MP5086	0.200	G2078	☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2021.01.13 <u>Michael P. Haughey</u> 15:24:14 -05'00' Tech Review / Date 2021.01.1 <u>SR</u> 6 11:30:47 Admin Review																																				
Simulator	Serial Number																																																	
0.000	G11621																																																	
Interferent	MP5087																																																	
0.050	MP5088																																																	
0.080	MP5086																																																	
0.200	G2078																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO
Time of Inspection: 11:43

Date of Inspection: 01/13/2021

Serial Number: 80-001071
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG011102 Exp: 04/20/2022
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.081
0.000	0.051	0.080	0.198	0.081
0.000	0.051	0.080	0.199	0.081
0.000	0.050	0.080	0.199	0.081

Standard Deviations	0.0007	0.0000	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

01/13/2021
Date

MX

CR 2021.01.
16
11:30:33
-05'00'

Stability Checks

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:08
Control Test	0.049	16:08
Air Blank	0.000	16:09
Control Test	0.049	16:09
Air Blank	0.000	16:10
Control Test	0.049	16:11
Air Blank	0.000	16:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:12
Control Test	0.079	16:13
Air Blank	0.000	16:13
Control Test	0.079	16:14
Air Blank	0.000	16:15
Control Test	0.079	16:15
Air Blank	0.000	16:16
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:18
Control Test	0.200	16:18
Air Blank	0.000	16:19
Control Test	0.199	16:19
Air Blank	0.000	16:20
Control Test	0.199	16:21
Air Blank	0.000	16:21
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	

wet

Operator's Signature

Operator's Signature

Operator's Signature

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001071
01/11/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:22
Control Test	0.081	16:23
Air Blank	0.000	16:23
Control Test	0.080	16:24
Air Blank	0.000	16:24
Control Test	0.080	16:24
Air Blank	0.000	16:25
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry

Operator's Signature

MX

2021.01.
16
11:30:20
-05'00"



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001071, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001071</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/13/2021</u>	0.080 g/ 210 L	0.005
Calibration Time:	<u>11:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/13/2021

Israel Soto

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 April 2020
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

[Signature]
2021.01
.16
11:29:59
-05'00'

Return Material Authorization

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: R. Patrone on 8/17/2020

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001071

Bill To Address:

Ship to Address:

FDLE-Tallahassee

Reason for Return:

Instrument will not power on. PLEASE UPLOAD ANY DATA PRIOR TO REPAIR.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Richard Patrone

Phone #: _____ Email: rpatrone@claysheriff.com

ATP Contact Name: Shayla Platt ATP Email: shaylaplatt@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency Clay County SO

S/N 80-001071

Florida Department of
Law Enforcement

Date In 08-26-2020 DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake Performed By <u>KAW</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ Final Release Date _____ _____ _____	Quality Checks Performed By _____ ☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Flow Calibration Performed By _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance Performed By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ Temperature Checks Performed By _____ ☐ Lab Temp °C _____ External Digital Therm. ID#: _____ ☐ 34°C +- .2 Serial #: _____ ☐ 34°C +- .2 Serial #: _____ ☐ 34°C +- .2 Serial #: _____																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Calibration Adjustment Performed By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection Performed By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> Attachments ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000																																																													
Interferent																																																													
0.050																																																													
0.080																																																													
0.200																																																													

MX