



# INSTRUMENT PROCESSING SHEET

Agency Pasco CSOS/N 80-001117

Florida Department of Law Enforcement

Date In 5/7/2021DI Completion Date 5/11/2021 Ship  P/U  H/D  CMI  EE

Intake	By TDG	Quality Checks	By TDG	Date 5/11/2021	Flow Calibration	By	Date																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks																																																																	
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP4863</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5097</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A AG026705 09/23/2022</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202010A 10/05/2022	0.080	202010B 10/05/2022	0.200	MP5097	202010D 10/06/2022	0.080 DGS	N/A AG026705 09/23/2022																																																				
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP4863	202010A 10/05/2022																																																																	
0.080		202010B 10/05/2022																																																																	
0.200	MP5097	202010D 10/06/2022																																																																	
0.080 DGS		N/A AG026705 09/23/2022																																																																	
				Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																															
				<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																															
				<b>DI Temp. Checks</b> By <u>MH</u> <input checked="" type="checkbox"/> Lab Temp °C <u>22.28</u> External Digital Therm. ID#: <u>300504</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4863</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP4864</u> <input checked="" type="checkbox"/> 34°C +/- .2 Serial #: <u>MP5097</u>																																																															
<b>Calibration Adjustment</b> By _____				<b>Department Inspection</b> By <u>TDG</u>																																																															
Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1017</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2020-A</u> Acetone Stock Solution Lot # <u>2020-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>SD1015</td> </tr> <tr> <td>0.050</td> <td>MP4863</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5097</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	SD1014	Interferent	SD1015	0.050	MP4863	0.080	MP4864	0.200	MP5097
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	SD1014																																																																		
Interferent	SD1015																																																																		
0.050	MP4863																																																																		
0.080	MP4864																																																																		
0.200	MP5097																																																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																															
				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																															
				Richard A Williams <small>Digitally signed by Richard A Williams Date: 2021.05.13 10:57:15 -0400</small> Tech Review / Date _____ Admin Review / Date <u>2021.05.14 09:30:08</u>																																																															

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 10:01

Date of Inspection: 05/11/2021

Serial Number: 80-001117  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


**Remarks:**

AI BYPASSED. COMPLIANCE NOT DETERMINED.

*Not determined*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


TAYLOR D GUTSCHOW

---

Signature and Printed Name

05/11/2021  
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1117	Pasco CSO	05/11/2021	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																												
<p>0.047 to 0.053</p> <p>Inadvertently printed to the external printer. Values were within the acceptable range. Will be attached, MG 5/11/2021</p>	<p>0.077 to 0.083</p> <p>PASCO COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 05/11/2021 Software: 8100.27</p> <p>SN 80-001117</p> <table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	11:29	Control Test	0.079	11:30	Air Blank	0.000	11:30	Control Test	0.079	11:31	Air Blank	0.000	11:32	Control Test	0.078	11:32	Air Blank	0.000	11:33	Control Test Stats			Average	0.0787		Std Dev	0.0005		Rel. Std Dev(%)	0.7339		<p>0.194 to 0.206</p> <p>PASCO COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 05/11/2021 Software: 8100.27</p> <p>SN 80-001117</p> <table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.5812</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	11:43	Control Test	0.200	11:44	Air Blank	0.000	11:44	Control Test	0.198	11:45	Air Blank	0.000	11:46	Control Test	0.198	11:46	Air Blank	0.000	11:47	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel. Std Dev(%)	0.5812		<p>0.077 to 0.083</p> <p>PASCO COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 05/11/2021 Software: 8100.27</p> <p>SN 80-001117</p> <table border="1"> <tr><th>Test</th><th>g/210L</th><th>Time</th></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Test	g/210L	Time	Air Blank	0.000	11:48	Control Test	0.079	11:48	Air Blank	0.000	11:49	Control Test	0.079	11:49	Air Blank	0.000	11:50	Control Test	0.079	11:50	Air Blank	0.000	11:50	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:29																																																																																																													
Control Test	0.079	11:30																																																																																																													
Air Blank	0.000	11:30																																																																																																													
Control Test	0.079	11:31																																																																																																													
Air Blank	0.000	11:32																																																																																																													
Control Test	0.078	11:32																																																																																																													
Air Blank	0.000	11:33																																																																																																													
Control Test Stats																																																																																																															
Average	0.0787																																																																																																														
Std Dev	0.0005																																																																																																														
Rel. Std Dev(%)	0.7339																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:43																																																																																																													
Control Test	0.200	11:44																																																																																																													
Air Blank	0.000	11:44																																																																																																													
Control Test	0.198	11:45																																																																																																													
Air Blank	0.000	11:46																																																																																																													
Control Test	0.198	11:46																																																																																																													
Air Blank	0.000	11:47																																																																																																													
Control Test Stats																																																																																																															
Average	0.1987																																																																																																														
Std Dev	0.0012																																																																																																														
Rel. Std Dev(%)	0.5812																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:48																																																																																																													
Control Test	0.079	11:48																																																																																																													
Air Blank	0.000	11:49																																																																																																													
Control Test	0.079	11:49																																																																																																													
Air Blank	0.000	11:50																																																																																																													
Control Test	0.079	11:50																																																																																																													
Air Blank	0.000	11:50																																																																																																													
Control Test Stats																																																																																																															
Average	0.0790																																																																																																														
Std Dev	0.0000																																																																																																														
Rel. Std Dev(%)	0.0000																																																																																																														
	<p>Operator's Signature</p> <p><i>MG</i></p>	<p>Operator's Signature</p> <p><i>MG</i></p>	<p>Operator's Signature</p> <p><i>MG</i></p>																																																																																																												

**Comments:**

PASCO COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001117  
05/11/2021  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:17
Control Test	0.049	11:18
Air Blank	0.000	11:18
Control Test	0.049	11:19
Air Blank	0.000	11:20
Control Test	0.049	11:20
Air Blank	0.000	11:21
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

---

Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO  
Time of Inspection: 14:14

Date of Inspection: 05/11/2021

Serial Number: 80-001117  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG026705 Exp: 09/23/2022
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.048	0.079	0.200	0.079
0.000	0.048	0.079	0.200	0.079
0.000	0.047	0.080	0.200	0.078
0.000	0.047	0.080	0.200	0.079


Standard Deviations	0.0010	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:


The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

05/11/2021  
 Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001117, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001117</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>05/11/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:14</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

05/11/2021

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 January 2021  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality