



INSTRUMENT PROCESSING SHEET

Agency Eglin AFB

S/N 80-001155

Florida Department of
Law Enforcement

Date In 7/1/2021

DI Completion Date 7/1/2021

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>IS</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>IS</u> Date <u>07-01-2021</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>215</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A 10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B 10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D 10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603 11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202010A 10-05-2022	0.080	MP5089	202010B 10-05-2022	0.200	MP5090	202010D 10-06-2022	0.080 DGS	N/A	AG931603 11-12-2021	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ DI Temp. Checks By <u>IS</u> ☒ Lab Temp °C <u>22.17</u> External Digital Therm. ID#: <u>569598</u> ☒ 34°C +/-2 Serial #: <u>MP5088</u> ☒ 34°C +/-2 Serial #: <u>MP5089</u> ☒ 34°C +/-2 Serial #: <u>MP5090</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP5088	202010A 10-05-2022																																																											
0.080	MP5089	202010B 10-05-2022																																																											
0.200	MP5090	202010D 10-06-2022																																																											
0.080 DGS	N/A	AG931603 11-12-2021																																																											
Calibration Adjustment By <u>IS</u> Barometric Pressure Gauge <u>1015</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>20060</td> <td>02-10-2022</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>20420</td> <td>09-22-2022</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>20160</td> <td>03-18-2022</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>20030</td> <td>01-21-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08121080A1</td> <td>05-05-2023</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202010A</td> <td>10-05-2022</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202010B</td> <td>10-05-2022</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202010D</td> <td>10-06-2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG931603</td> <td>11-12-2021</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	20060	02-10-2022	0.100	MP5083	20420	09-22-2022	0.200	MP5084	20160	03-18-2022	0.300	MP5085	20030	01-21-2022	0.080 DGS	N/A	08121080A1	05-05-2023	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202010A	10-05-2022	0.080	MP5089	202010B	10-05-2022	0.200	MP5090	202010D	10-06-2022	0.080 DGS	N/A	AG931603	11-12-2021	Department Inspection By <u>IS</u> Barometric Pressure ID# <u>28427</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-A</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP5091	N/A	N/A																																																										
0.040	MP5082	20060	02-10-2022																																																										
0.100	MP5083	20420	09-22-2022																																																										
0.200	MP5084	20160	03-18-2022																																																										
0.300	MP5085	20030	01-21-2022																																																										
0.080 DGS	N/A	08121080A1	05-05-2023																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP5088	202010A	10-05-2022																																																										
0.080	MP5089	202010B	10-05-2022																																																										
0.200	MP5090	202010D	10-06-2022																																																										
0.080 DGS	N/A	AG931603	11-12-2021																																																										
Simulator	Serial Number																																																												
0.000	MP5086																																																												
Interferent	MP5087																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5090																																																												
Notes/Suggested Service: <u>Pressure on instrument differed by more than 1% compared to barometer, performed optical bench calibration adjust. IS</u> _____ _____ _____ _____ _____ _____	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____ ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Richard A Williams Digitally signed by Richard A Williams Date: 2021.07.01 15:31:58 -0400 01 15:43:41 04:00' Tech Review / Date _____ Admin Review / Date _____																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EGLIN AFB
Time of Inspection: 14:07

Date of Inspection: 07/01/2021

Serial Number: 80-001155
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG931603 Exp: 11/12/2021
0.000	0.050	0.079	0.202	0.079
0.000	0.051	0.080	0.202	0.080
0.000	0.050	0.079	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.051	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.080	0.202	0.080
Standard Deviations	0.0004	0.0004	0.0000	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

07/01/2021
Date

stability checks

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:10
Control Test	0.050	11:11
Air Blank	0.000	11:11
Control Test	0.050	11:12
Air Blank	0.000	11:12
Control Test	0.049	11:13
Air Blank	0.000	11:14
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.080	11:15
Air Blank	0.000	11:16
Control Test	0.079	11:16
Air Blank	0.000	11:17
Control Test	0.078	11:18
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.199	11:22
Air Blank	0.000	11:22
Control Test	0.198	11:23
Air Blank	0.000	11:24
Control Test	0.198	11:24
Air Blank	0.000	11:25
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

wet


Operator's Signature


Operator's Signature


Operator's Signature

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.080	11:27
Air Blank	0.000	11:28
Control Test	0.080	11:28
Air Blank	0.000	11:29
Control Test	0.080	11:29
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry


Operator's Signature

Optical Bench Cal. Adj.

Handwritten signature

EQUIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021 11:31:00

Auto Calibration
Max Power Res Value = 79
Auto Range Res Value = 62

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12711, 9um Io = 13009

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1180 (-0.0080)
Sample #2 = 0.0910 (0.0500)
Sample #3 = 0.0940 (0.0710)
Sample #4 = 0.0880 (0.0610)
Avg % Abs = 0.0910 (0.0607)
STD DEV = 0.0030 (0.0105)
REL STD DEV = 3.297 (17.314)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1660 (-0.0030)
Sample #2 = 0.1250 (0.0430)
Sample #3 = 0.1390 (0.0370)
Sample #4 = 0.1330 (0.0380)
Avg % Abs = 0.1323 (0.0393)
STD DEV = 0.0070 (0.0132)
REL STD DEV = 5.308 (8.173)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12698, 9um Io = 13003

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8320 (-0.0130)
Sample #2 = 0.7930 (0.0260)
Sample #3 = 0.8270 (0.0070)
Sample #4 = 0.7960 (0.0380)
Avg % Abs = 0.8053 (0.0237)
STD DEV = 0.0188 (0.0156)
REL STD DEV = 2.337 (66.047)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5800 (-0.0040)
Sample #2 = 1.5150 (0.0460)
Sample #3 = 1.5350 (0.0410)
Sample #4 = 1.5060 (0.0640)
Avg % Abs = 1.5187 (0.0503)
STD DEV = 0.0148 (0.0121)
REL STD DEV = 0.977 (24.033)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12694, 9um Io = 12999

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.8760 (-0.0090)
Sample #2 = 1.8530 (0.0290)
Sample #3 = 1.8280 (0.0460)
Sample #4 = 1.8260 (0.0410)
Avg % Abs = 1.8357 (0.0387)
STD DEV = 0.0150 (0.0087)
REL STD DEV = 0.820 (22.595)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6200 (0.0080)
Sample #2 = 3.5800 (0.0510)
Sample #3 = 3.5440 (0.0680)
Sample #4 = 3.5610 (0.0560)
Avg % Abs = 3.5617 (0.0583)
STD DEV = 0.0180 (0.0087)
REL STD DEV = 0.505 (14.978)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12687, 9um Io = 12992

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6290 (-0.0270)
Sample #2 = 3.5460 (0.0340)
Sample #3 = 3.5370 (0.0800)
Sample #4 = 3.5330 (0.0690)
Avg % Abs = 3.5387 (0.0610)
STD DEV = 0.0067 (0.0240)
REL STD DEV = 0.188 (39.378)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.9420 (-0.0230)
Sample #2 = 6.8380 (0.0810)
Sample #3 = 6.8350 (0.1150)
Sample #4 = 6.8200 (0.1070)
Avg % Abs = 6.8310 (0.1010)
STD DEV = 0.0096 (0.0178)
REL STD DEV = 0.141 (17.600)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12685, 9um Io = 12993

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.2480 (-0.0220)
Sample #2 = 5.1510 (0.1070)
Sample #3 = 5.1310 (0.1110)
Sample #4 = 5.1260 (0.1320)
Avg % Abs = 5.1360 (0.1167)
STD DEV = 0.0132 (0.0134)
REL STD DEV = 0.258 (11.510)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.9990 (-0.0130)
Sample #2 = 9.8150 (0.1840)
Sample #3 = 9.8080 (0.1890)
Sample #4 = 9.7730 (0.2240)
Avg % Abs = 9.7987 (0.1990)
STD DEV = 0.0225 (0.0218)
REL STD DEV = 0.230 (10.952)

***** AUTO CAL DATA *****

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.091
Std Dev = 0.00 Rel Std Dev = 3.30
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.805
Std Dev = 0.02 Rel Std Dev = 2.34
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.836
Std Dev = 0.02 Rel Std Dev = 0.82
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.539
Std Dev = 0.01 Rel Std Dev = 0.19
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.136
Std Dev = 0.01 Rel Std Dev = 0.26
Zero Order Coef = -234.50
First Order Coef = 2636.06
Second Order Coef = 36.80
Standard Deviation = 23.781483

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.132
Std Dev = 0.01 Rel Std Dev = 5.31
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.519
Std Dev = 0.01 Rel Std Dev = 0.98
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.562
Std Dev = 0.02 Rel Std Dev = 0.51
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.831
Std Dev = 0.01 Rel Std Dev = 0.14
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.799
Std Dev = 0.02 Rel Std Dev = 0.23
Zero Order Coef = -148.03
First Order Coef = 1311.65
Second Order Coef = 16.27
Standard Deviation = 34.73555

Opt. Bench
cal. Adj

Post stability checks

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	-0.0002
0.100	0.099	0.0007
0.200	0.201	-0.0006
0.300	0.300	0.0002

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0005
0.040	0.040	0.0005
0.100	0.099	0.0007
0.200	0.201	-0.0010
0.300	0.300	0.0004

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3091.00

Sample #2 = 3093.00

Sample #3 = 3160.00

Sample #4 = 3175.00

Average Result = 3142.6667

STD DEV = 43.6616

REL STD DEV = 1.389

***** CHANNEL 2

Sample #1 = 3297.00

Sample #2 = 3306.00

Sample #3 = 3318.00

Sample #4 = 3343.00

Average Result = 3322.3333

STD DEV = 18.8768

REL STD DEV = 0.568

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1014

3 um H2O Adjust (mg/l*10,000) = 667

9 um H2O Adjust (mg/l*10,000) = 487

**** AUTO CAL PASS

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.050	12:19
Air Blank	0.000	12:20
Control Test	0.049	12:21
Air Blank	0.000	12:21
Control Test	0.049	12:22
Air Blank	0.000	12:22
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.201	12:29
Air Blank	0.000	12:30
Control Test	0.200	12:30
Air Blank	0.000	12:31
Control Test	0.200	12:32
Air Blank	0.000	12:32
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	



Operator's Signature

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:23
Control Test	0.079	12:24
Air Blank	0.000	12:25
Control Test	0.079	12:25
Air Blank	0.000	12:26
Control Test	0.079	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet



Operator's Signature

EGLIN AFB
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001155
07/01/2021
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.079	12:35
Air Blank	0.000	12:36
Control Test	0.079	12:36
Air Blank	0.000	12:36
Control Test	0.080	12:37
Air Blank	0.000	12:37
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	



Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001155, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001155</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>EGLIN AFB</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>07/01/2021</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:07</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Thermometer temperatures are checked with NIST traceable Eutechnics 4400 digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the uses of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Israel Soto
Date: 2021.07.01 15:17:12
-04'00'

Israel Soto

07/01/2021

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1