

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1653	St. Petersburg PD	11/04/2021	TDG ML

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet X																																																																																																
<p>ST. PETERSBURG PD Model: 8000 SN 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> </tbody> </table> <p>Control Test Stats Average 0.0000 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:00	Control Test	0.000	11:01	Air Blank	0.000	11:01	Control Test	0.000	11:02	Air Blank	0.000	11:02	Control Test	0.000	11:03	Air Blank	0.000	11:03	<p>ST. PETERSBURG PD Model: 8000 SN 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> </tbody> </table> <p>Control Test Stats Average 0.0000 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:09	Control Test	0.000	11:10	Air Blank	0.000	11:11	Control Test	0.000	11:11	Air Blank	0.000	11:12	Control Test	0.000	11:12	Air Blank	0.000	11:13	<p>ST. PETERSBURG PD Model: 8000 SN 80-101653 11/04/2021 Software: 8110.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> </tbody> </table> <p>Control Test Stats Average 0.0000 Std Dev 0.0000 Rel. Std Dev(%) 0.0000</p> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:17	Control Test	0.000	11:18	Air Blank	0.000	11:18	Control Test	0.000	11:19	Air Blank	0.000	11:19	Control Test	0.000	11:20	Air Blank	0.000	11:20	<p>ST. PETERSBURG PD Model: 8000 SN 80-101653 11/04/2021 Software: 8110.27</p> <p style="text-align: center;">DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> </tbody> </table> <p>Control Test Stats Average 0.0810 Std Dev 0.0010 Rel. Std Dev(%) 1.2346</p> <p>ML Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:21	Control Test	0.082	11:21	Air Blank	0.000	11:22	Control Test	0.080	11:22	Air Blank	0.000	11:23	Control Test	0.081	11:23	Air Blank	0.000	11:24
Test	g/210L	Time																																																																																																	
Air Blank	0.000	11:00																																																																																																	
Control Test	0.000	11:01																																																																																																	
Air Blank	0.000	11:01																																																																																																	
Control Test	0.000	11:02																																																																																																	
Air Blank	0.000	11:02																																																																																																	
Control Test	0.000	11:03																																																																																																	
Air Blank	0.000	11:03																																																																																																	
Test	g/210L	Time																																																																																																	
Air Blank	0.000	11:09																																																																																																	
Control Test	0.000	11:10																																																																																																	
Air Blank	0.000	11:11																																																																																																	
Control Test	0.000	11:11																																																																																																	
Air Blank	0.000	11:12																																																																																																	
Control Test	0.000	11:12																																																																																																	
Air Blank	0.000	11:13																																																																																																	
Test	g/210L	Time																																																																																																	
Air Blank	0.000	11:17																																																																																																	
Control Test	0.000	11:18																																																																																																	
Air Blank	0.000	11:18																																																																																																	
Control Test	0.000	11:19																																																																																																	
Air Blank	0.000	11:19																																																																																																	
Control Test	0.000	11:20																																																																																																	
Air Blank	0.000	11:20																																																																																																	
Test	g/210L	Time																																																																																																	
Air Blank	0.000	11:21																																																																																																	
Control Test	0.082	11:21																																																																																																	
Air Blank	0.000	11:22																																																																																																	
Control Test	0.080	11:22																																																																																																	
Air Blank	0.000	11:23																																																																																																	
Control Test	0.081	11:23																																																																																																	
Air Blank	0.000	11:24																																																																																																	

Comments: Instrument not pulling samples from sims. No bubbles observed during wet-bath control tests. Sims worked fine on multiple other instruments on bench. Instrument was able to pull wet-bath sample during the 10/28/2021 Agency Inspection. ML 11/04/2021

